



December 22, 2025

Taiwan Surgical Corporation
Ken Chen
Project Director
3f., # 12, Sec. 12, Sheng Yi Rd., Hsinchu County
Zhubei City, TW 30261
Taiwan

Re: K253903

Trade/Device Name: InnoClip Disposable Clip Applier
Regulation Number: 21 CFR 878.4300
Regulation Name: Implantable Clip
Regulatory Class: Class II
Product Code: FZP, GDO
Dated: November 21, 2025
Received: December 5, 2025

Dear Ken Chen:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory->

[assistance/contact-us-division-industry-and-consumer-education-dice](#) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**TEK N.
LAMICHHANE -S**

Tek N. Lamichhane, Ph.D.
Assistant Director
DHT4B: Division of Plastic and
Reconstructive Surgery Devices
OHT4: Office of Surgical and
Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use510(k) Number (*if known*)

K253903

Device Name

InnoClip™ Disposable Clip Applier

Indications for Use (Describe)

InnoClip™ Disposable Clip Applier is indicated for patients undergoing laparoscopic surgical procedures involving occlusion of blood vessel, ducts and other tubular structures, and for radiographic marking.

Type of Use (*Select one or both, as applicable*) Prescription Use (Part 21 CFR 801 Subpart D) Over-The-Counter Use (21 CFR 801 Subpart C)**CONTINUE ON A SEPARATE PAGE IF NEEDED.**

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASstaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary (K253903)

This 510(k) summary of safety and effectiveness information is submitted as part of the Premarket Notification in compliance with requirements of CFR Part 807, Subpart E and Section 807.92

Submitter:

Submitter: TAIWAN SURGICAL CORPORATION
Mailing Address: 3F., No.12, Sec.2, ShengYi Rd., Zhubei City, Hsinchu County 302, Taiwan
Phone Number: +886-3-6588129 ext. 100
Fax Number: +886-3-6588355
Contact Person: Ken Chen
Email: ernest.chen@twsc.com.tw
Date Prepared: November 26, 2025

Device

Trade Name: InnoClip Disposable Clip Applier
Common Name: Implantable clip
Classification Name: Clip, Implantable
Classification Name: 878.4300
Classification Product Code: FZP, GDO
Device Class: II

Predicate Device K202994, InnoClip Disposable Clip Applier, InnoClip Clip Applier (Primary)
K150259, CLIP PLUS Disposable Clip Applier

Device Description The InnoClip™ Disposable Clip Applier consists of a handle piece and 10, 16 or 20 implantable titanium clips. The product line includes two shaft diameters, 10 mm and 5 mm. Titanium clip sizes are 5.5 × 9.0 mm and 5.5 × 11 mm for the 10 mm applier, and 4.6 × 9.1 mm for the 5 mm applier. The applier is designed to be introduced through applicable trocar sleeves, or larger sleeve with a fitted converter.

Indications for Use InnoClip™ Disposable Clip Applier is indicated for patients undergoing laparoscopic surgical procedures involving occlusion of blood vessel, ducts and other tubular structures, and for radiographic marking.

Indication for Use Comparison The InnoClip™ Disposable Clip Applier and the predicate device (K202994 InnoClip™ Disposable Clip Applier) has the same intended use as both devices are intended for patients undergoing laparoscopic surgical procedures involving occlusion of blood vessel, ducts and other tubular structures, and for radiographic marking.

Technological Comparison The InnoClip™ Disposable Clip Applier demonstrates substantial equivalence in terms of intended use, technological characteristics, and performance. The subject device shares identical classification, indications for use, target population, contraindications, mechanism of action, energy source, functional operation, operating environment, patient contact time, sterilization method, and overall design with the predicate device. Minor differences were identified, including modifications to component parts and materials of the components within the InnoClip™ Disposable Clip Applier.

Non-Clinical and/or Clinical Tests Summary & Conclusion **Biocompatibility Testing:**
Biocompatibility testing was conducted in accordance with ISO 10993-1:2018, Biological Evaluation of Medical Devices – Part 1: Evaluation and Testing within a Risk Management Process. The devices are classified as externally communicating medical devices with contact to tissue, bone, or dentin for limited exposure (contact less than 24 hours).

Sterilization Validation:

The sterilization process was validated to achieve a Sterility Assurance Level (SAL) of 10-6 in accordance with ISO11137-1:2015 and ISO11137-2:2015.

Shelf-life

The 3-year shelf-life was established and validated through accelerated aging study conducted in accordance with ASTM F1980 and ISO11607-1/-2.

Bench Performance Testing:

Bench performance testing included evaluation of clipped pull out force, clipped slip force, air tight capability, clipped gap, clip force, and trigger force. The test results showed that the subject device has the similar device performance compared to the predicate device.

Conclusion

All test results demonstrate that the InnoClip™ Disposable Clip Applier meet their acceptance criteria and intended use requirements. Therefore, the subject device (K253903) is considered as safe and effective as the predicate device.