



December 9, 2025

ProVerum Ltd.
Niamh Mahony
Senior Regulatory Director
2nd Floor Astor Hall, 4-8 Eden Quay
Dublin, Dublin D01N5W8
IRELAND

Re: P250011
Trade/Device Name: ProVee System for BPH
Product Code: MER
Filed: April 2, 2025
Amended: September 5, 2025

Dear Niamh Mahony:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your premarket approval application (PMA) for the ProVee System for BPH. This device is indicated for the treatment of obstructive lower urinary tract symptoms secondary to benign prostatic hyperplasia (BPH), in men with prostatic urethral lengths greater than or equal to 3.75 cm and prostatic volumes between 30 cc and 80 cc. Based upon the information submitted, the PMA is approved. You may begin commercial distribution of the device in accordance with the conditions of approval described below. Although this letter refers to your product as a device, please be aware that some approved products may instead be combination products. The Premarket Approval Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPMA/pma.cfm> identifies combination product submissions.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the act). FDA has determined that these restrictions on sale and distribution are necessary to provide reasonable assurance of the safety and effectiveness of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the act, in addition to all other applicable requirements, including those governing the manufacture, distribution, and marketing of devices.

Expiration dating for this device has been established and approved at 12 months. This is to advise you that the protocol you used to establish this expiration dating is considered an approved protocol for the purpose of extending the expiration dating as provided by 21 CFR 814.39(a)(7).

Continued approval of the PMA is contingent upon the submission of periodic reports, required under 21 CFR 814.84, at intervals of one year (unless otherwise specified) from the date of approval of the original PMA. This report, identified as "Annual Report" and bearing the applicable PMA reference number, should

be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and must include the information required by 21 CFR 814.84.

In addition to the above, and in order to provide continued reasonable assurance of the safety and effectiveness of the PMA device, under 21 CFR 814.82(a)(9), the Annual Report must include, separately for each model number (if applicable), the number of devices sold and distributed during the reporting period, including those distributed to distributors. The distribution data will serve as a denominator and provide necessary context for FDA to ascertain the frequency and prevalence of adverse events, as FDA evaluates the continued safety and effectiveness of the device.

In addition to the Annual Report requirements, you must provide the following data in post-approval study (PAS) reports for each PAS listed below.

1. 60-Month Follow-Up for the ProVIDE Pivotal Study (Ver: CIP 002 Rev 05, December 04, 2025)

This study was initiated prior to device approval and is a prospective, multi-center, double-blind, 2:1 randomized, sham controlled clinical study. It was conducted at 17 sites and enrolled 221 subjects. One hundred and fifty (150) subjects were implanted in the treatment arm, and 71 subjects were exposed to a sham control. The 12-month outcomes from this study were used to support PMA approval. The 60-month follow-up data from this study will be used to evaluate the continued safety and effectiveness of the ProVee Expander System.

No greater than 20% the patients enrolled in the study (i.e., patients enrolled in the treatment arm and patients crossed over from the sham to treatment arm) will be lost to follow up through 60-months.

You must collect and report the following clinical outcomes annually through 60-months:

Effectiveness

- Percent of subjects who experienced at least a 30% improvement in International Prostate Symptom Score (IPSS) from baseline
- Percent change in International Prostate Symptom Score (IPSS) in the treatment arm compared to baseline.
- Mean improvement from baseline in quality of life (QoL) questionnaire.
- Mean improvement from baseline uroflowmetry measures of peak flow rate (Qmax).
- Mean improvement from baseline in post void residual (PVR) volume.
- Post-procedure incidence of secondary reintervention using alternate surgical lower urinary tract symptoms (LUTS) therapy.
- Post-procedure incidence of secondary reintervention using standard pharmacological agents for LUTS therapy.

- Post-procedure device removal rate and summary of the reasons for removal (including the number and rate for each reason).

Safety

- Rate of device or procedure related adverse events.
- Change in sexual health characterized by change in International Index of Erectile Function, specifically the Erectile Function domain (IIEF-5) and Male Sexual Health Questionnaire – Ejaculatory Domain (MSHQ-EjD).
- Proportion of subjects with adverse events classified as Clavien-Dindo Grade IIIb or higher or any event resulting in persistent disability.
- Assessment of retrieval procedure related adverse events related to implant removal.

Data must be summarized descriptively, without statistical testing. Patient and physician labeling must be updated annually via a PMA supplement to include the effectiveness and safety outcomes listed above after the patients enrolled in this study complete each annual year of follow up.

2. 60-Month Follow-Up of the ProVIDE II Bridging Study (Ver: CIP 003 Rev 01, December 04, 2025)

This study was initiated prior to device approval and is a prospective, multi-center, non-randomized bridging study. It was conducted at 8 sites and enrolled 40 subjects. Forty (40) subjects were implanted in the study. The 3-month outcomes from the study were used to support PMA approval. The 60-month follow-up data from this study will be used to evaluate the continued safety and effectiveness of the ProVee System for BPH.

No greater than 20% of the patients enrolled in the study will be lost to follow up through 60-months.

You must collect and report the following clinical outcomes annually through 60-months:

Effectiveness

- Percent of subjects who experienced at least a 30% improvement in International Prostate Symptom Score (IPSS) from baseline.
- Percent change in IPSS compared to baseline.
- Mean improvement from baseline in quality of life (QoL) questionnaire.
- Mean improvement from baseline uroflowmetry measures of peak flow rate (Qmax).
- Mean improvement from baseline in post void residual (PVR) volume.

- Post-procedure incidence of secondary reintervention using an alternative surgical lower urinary tract symptoms (LUTS) therapy.
- Post-procedure incidence of secondary reintervention using standard pharmacological agents for LUTS therapy.
- Post procedure device removal rate and summary of the reasons for removal (including the number and rate for each reason)

Safety

- Rate of device or procedure related adverse events.
- Change in sexual health characterized by change in International Index of Erectile Function, specifically the Erectile Function domain (IIEF-5) and Male Sexual Health Questionnaire – Ejaculatory Domain (MSHQ-EjD).
- Proportion of subjects with adverse events classified as Clavien-Dindo Grade IIIb or higher or any event resulting in persistent disability.
- Assessment of retrieval procedure related adverse events related to implant removal.

Data must be summarized descriptively, without statistical testing. You must update the patient and physician labeling annually via a PMA supplement to include the effectiveness and safety outcomes listed above after the patients enrolled in this study complete each annual year of follow up.

The PAS Progress Reports must be submitted every six (6) months for the first year and annually thereafter, from the date of the PMA approval letter, unless otherwise specified by FDA. The Final PAS Report should be submitted no later than three (3) months after study completion (i.e., last subject's last follow-up date).

Each PAS report should be submitted to the address below identified as a "PMA Post-Approval Study Report" in accordance with how the study is identified above and bearing the applicable PMA reference number.

Be advised that failure to comply with any post-approval requirement, including initiation, enrollment, and completion requirements outlined above, constitutes grounds for FDA withdrawal of approval of the PMA in accordance with 21 CFR 814.82(c) and 814.46(a)(2).

Be advised that the failure to conduct any such study in compliance with the good clinical laboratory practices in 21 CFR part 58 (if a non-clinical study subject to part 58) or the institutional review board regulations in 21 CFR part 56 and the informed consent regulations in 21 CFR part 50 (if a clinical study involving human subjects) may be grounds for FDA withdrawal of approval of the PMA in accordance with 21 CFR 814.46(a)(3)-(4).

Be advised that protocol information, interim and final results will be published on the Post-Approval Studies Program Database Webpage, available at https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPMA/pma_pas.cfm.

In addition, the results from any post approval study should be included in the labeling as these data become available. Under 21 CFR 814.39, any updated labeling must be submitted to FDA in the form of a PMA Supplement. For more information on post-approval studies, see the FDA guidance document entitled, "Procedures for Handling Post-Approval Studies Imposed by Premarket Approval Application Order" (<https://www.fda.gov/media/71327/download>).

This is a reminder that as of September 24, 2014, class III devices are subject to certain provisions of the final Unique Device Identification (UDI) rule. These provisions include the requirement to provide a UDI on the device label and packages (21 CFR 801.20), format dates on the device label in accordance with 21 CFR 801.18, and submit data to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). Additionally, 21 CFR 814.84 (b)(4) requires PMA annual reports submitted after September 24, 2014, to identify each device identifier currently in use for the subject device, and the device identifiers for devices that have been discontinued since the previous periodic report. It is not necessary to identify any device identifier discontinued prior to December 23, 2013. Combination Products may also be subject to UDI requirements (see 21 CFR 801.30). For more information on these requirements, please see the UDI website available at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-udi-system>.

Before making any change affecting the safety or effectiveness of the PMA device, you must submit a PMA supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.39. All PMA supplements and alternate submissions (30-day notice) must comply with the applicable requirements in 21 CFR 814.39. Additional information about changes that may require a PMA supplement are provided in the FDA guidance document entitled, "Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process" <https://www.fda.gov/media/81431/download>.

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production and process controls (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52 for devices or post-marketing safety reporting (21 CFR Part 4, Subpart B) for combination products, you are required to report adverse events for this device. Manufacturers of medical devices, including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or

2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems> and on combination product post-marketing safety reporting is available at <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>.

In accordance with the recall requirements specified in 21 CFR 806.10 for devices or the post-marketing safety reporting requirements (21 CFR Part 4, Subpart B) for combination products, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at <https://www.fda.gov/safety/recalls-market-withdrawals-safety-alerts/industry-guidance-recalls>.

The CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading. CDRH will notify the public of its decision to approve your PMA by making available, among other information, a summary of the safety and effectiveness data upon which the approval is based. The information can be found at <https://www.fda.gov/medical-devices/device-approvals-denials-and-clearances/pma-approvals>. Written requests for this information can also be made to the Food and Drug Administration, Dockets Management Branch, (HFA-305), 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. The written request should include the PMA number or docket number. Within 30 days from the date that this information is placed on the Internet, any interested person may seek review of this decision by submitting a petition for review under section 515(g) of the act and requesting either a hearing or review by an independent advisory committee. FDA may, for good cause, extend this 30-day filing period.

Failure to comply with any post-approval requirement constitutes a ground for withdrawal of approval of a PMA. The introduction or delivery for introduction into interstate commerce of a device that is not in compliance with its conditions of approval is a violation of law.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this PMA submission with a copy of all final labeling. Final labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final labeling is identical to the labeling approved in draft form. If the final labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted to the CDRH Portal and should reference the above PMA number to facilitate processing. For more information on the CDRH Portal, please visit <https://www.fda.gov/medical-devices/industry-medical-devices/send-and-track-medical-device-premarket-submissions-online-cdrh-portal>.

If you have any questions concerning this approval order, please contact Feba Abraham at (301) 796-5772 or Feba.Abraham@fda.hhs.gov.

Sincerely,

Sharon M. Andrews
Director
DHT3B: Division of Reproductive,
Gynecology, and Urology Devices
OHT3: Office of Gastrorenal, ObGyn,
General Hospital, and Urology Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health