



March 26, 2026

bioMérieux, Inc.
Neal Bertels
Senior Regulatory Affairs Specialist
595 Anglum Rd.
Hazelwood, Missouri 63042

Re: K260281

Trade/Device Name: VITEK 2 AST-*Streptococcus* Cefuroxime (≤ 0.125 - ≥ 8 $\mu\text{g/mL}$)
Regulation Number: 21 CFR 866.1645
Regulation Name: Fully automated short-term incubation cycle antimicrobial susceptibility system
Regulatory Class: Class II
Product Code: LON, LTT, LTW
Dated: January 29, 2026
Received: January 29, 2026

Dear Neal Bertels:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

FDA's substantial equivalence determination also included the review and clearance of your Predetermined Change Control Plan (PCCP). Under section 515C(b)(1) of the Act, a new premarket notification is not

required for a change to a device cleared under section 510(k) of the Act, if such change is consistent with an established PCCP granted pursuant to section 515C(b)(2) of the Act. Under 21 CFR 807.81(a)(3), a new premarket notification is required if there is a major change or modification in the intended use of a device, or if there is a change or modification in a device that could significantly affect the safety or effectiveness of the device, e.g., a significant change or modification in design, material, chemical composition, energy source, or manufacturing process. Accordingly, if deviations from the established PCCP result in a major change or modification in the intended use of the device, or result in a change or modification in the device that could significantly affect the safety or effectiveness of the device, then a new premarket notification would be required consistent with section 515C(b)(1) of the Act and 21 CFR 807.81(a)(3). Failure to submit such a premarket submission would constitute adulteration and misbranding under sections 501(f)(1)(B) and 502(o) of the Act, respectively.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality Management System Regulation (QMSR) (21 CFR Part 820), which includes, but is not limited to, ISO 13485 clause 7.3 (Design controls), ISO 13484 clause 8.3 (Nonconforming product), and ISO 13485 clause 8.5 (Corrective and preventative action). Please note that regardless of whether a change requires premarket review, the QMSR requires device manufacturers to review and approve changes to device design and production (ISO 13485 clause 7.3 and 21 CFR 820.70) and document changes and approvals in the Medical Device File (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the Quality Management System Regulation (QMSR) (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Ribhi Shawar -S

Ribhi Shawar, Ph.D. (ABMM)
Branch Chief, General Bacteriology and Antimicrobial
Susceptibility Branch
Division of Microbiology Devices
OHT7: Office of In Vitro Diagnostics
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K260281

Device Name
VITEK 2 AST-Streptococcus Cefuroxime ($\leq 0.125 - \geq 8$ $\mu\text{g/mL}$)

Indications for Use (Describe)

VITEK 2 AST-Streptococcus Cefuroxime ($\leq 0.125 - \geq 8$) is designed for antimicrobial susceptibility testing of Streptococcus species and is intended for use with the VITEK 2 Systems as a laboratory aid in the determination of in vitro susceptibility to antimicrobial agents.

VITEK 2 AST-Streptococcus Cefuroxime ($\leq 0.125 - \geq 8$) is a quantitative test. Testing is indicated for Streptococcus pneumoniae as recognized by the FDA Susceptibility Test Interpretive Criteria (STIC).

VITEK 2 AST-Streptococcus Cefuroxime ($\leq 0.125 - \geq 8$) has demonstrated acceptable performance with the following organism:
Streptococcus pneumoniae

The VITEK 2 Antimicrobial Susceptibility Test (AST) is intended to be used with the VITEK 2 Systems in clinical laboratories as an in vitro test to determine the susceptibility of Streptococcus pneumoniae, beta-hemolytic Streptococcus and Viridans Streptococcus to antimicrobial agents when used as instructed.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



510(k) SUMMARY

VITEK[®] 2 AST-*Streptococcus* Cefuroxime

A. 510(k) Submission Information:

| | |
|----------------------|--|
| Submitter's Name: | bioMérieux, Inc. |
| Address: | 595 Anglum Road Hazelwood, MO 63042 |
| Contact Person: | Neal Bertels Senior Regulatory Affairs Specialist |
| Phone Number: | 314 -898-2195 |
| Fax Number: | 314-731-8689 |
| Date of Preparation: | March 23, 2025 |

B. Device Name:

| | |
|----------------------|---|
| Formal/Trade Name: | VITEK [®] 2 AST- <i>Streptococcus</i> Cefuroxime ($\leq 0.125 - \geq 8$ $\mu\text{g/mL}$) |
| Classification Name: | 21 CFR 866.1645 Fully Automated Short-Term Incubation Cycle Antimicrobial Susceptibility System Product Codes: LON, LTT, LTW |
| Common Name: | VITEK [®] 2 AST-ST Cefuroxime |

C. Predicate Device: VITEK[®] 2 AST-*Streptococcus* Penicillin (K232201)

D. Device Description:

The principle of the VITEK[®] 2 AST cards is based on the microdilution minimum inhibitory concentration (MIC) technique reported by MacLowry and Marsh⁽¹⁾ and Gerlach⁽²⁾. The VITEK[®] 2 AST card is essentially a miniaturized, abbreviated and automated version of the doubling dilution technique⁽³⁾.

Each VITEK[®] 2 AST card contains 64 wells. A control well which only contains microbiological culture media is resident on all cards. The remaining wells contain premeasured portions of a specific antibiotic combined with culture media. The bacterial or yeast isolate to be tested is diluted to a standardized concentration with 0.45 – 0.5% saline before being used to rehydrate the antimicrobial medium within the card. The VITEK[®] 2 System automatically fills, seals and places the card into the

incubator/reader. The VITEK[®] 2 Compact has a manual filling, sealing and loading operation. The VITEK[®] 2 Systems monitor the growth of each well in the card over a defined period of time. At the completion of the incubation cycle, a report is generated that contains the MIC value along with the interpretive category result for each antibiotic contained on the card.

VITEK[®] 2 AST-*Streptococcus* Cefuroxime has the following concentrations in the card: 0.25, 0.5, 1, and 2 (equivalent standard method concentration by efficacy in µg/mL).

E. Substantial Equivalence Information

The similarities and differences of the VITEK 2 AST-*Streptococcus* Cefuroxime when compared to the predicate device, VITEK 2 AST-*Streptococcus* Penicillin (K232201), are described in the following table. The only differences between both devices are the Indications for Use and the breakpoints used to analyze the data performance. The below table provides the similarities and differences:

| Item | Device: VITEK [®] 2 AST- <i>Streptococcus</i> Cefuroxime | VITEK [®] 2 AST- <i>Streptococcus</i> Penicillin (K232201) |
|---------------------|---|--|
| Similarities | | |
| Intended Use | <p>VITEK[®] 2 AST-<i>Streptococcus</i> Cefuroxime (≤0.125 - ≥8) is designed for antimicrobial susceptibility testing of <i>Streptococcus</i> species and is intended for use with the VITEK[®] 2 Systems as a laboratory aid in the determination of <i>in vitro</i> susceptibility to antimicrobial agents. VITEK[®] 2 AST-<i>Streptococcus</i> Cefuroxime (≤0.125 - ≥8) is a quantitative test. Testing is indicated for <i>Streptococcus pneumoniae</i>, as recognized by the FDA Susceptibility Testing Criteria (STIC).</p> <p>The VITEK 2 <i>Streptococcus</i> Susceptibility Card is intended for use with the VITEK[®] 2 Systems in clinical laboratories as an <i>in vitro</i> test to determine the susceptibility of <i>Streptococcus pneumoniae</i>, beta-hemolytic <i>Streptococcus</i>, and Viridans <i>Streptococcus</i> to antimicrobial agents when used as instructed.</p> <p>Testing is indicated for <i>Streptococcus pneumoniae</i> as recognized by the FDA Susceptibility Test Interpretive Criteria (STIC). VITEK[®] 2 AST- <i>Streptococcus</i> Cefuroxime has demonstrated acceptable performance with the following organism: <i>Streptococcus pneumoniae</i></p> | <p>VITEK[®] 2 AST-<i>Streptococcus</i> Penicillin is designed for antimicrobial susceptibility testing of <i>Streptococcus</i> species and is intended for use with the VITEK[®] 2 and VITEK[®] 2 Compact Systems as a laboratory aid in the determination of <i>in vitro</i> susceptibility to antimicrobial agents. VITEK[®] 2 <i>Streptococcus</i> Penicillin is a quantitative test. Penicillin has been shown to be active against most strains of the microorganisms listed below, according to the FDA label for this antimicrobial.</p> <p>The VITEK[®] 2 <i>Streptococcus</i> Susceptibility Card is intended for use with the VITEK[®] 2 Systems in clinical laboratories as an <i>in vitro</i> test to determine the susceptibility of clinically significant <i>Streptococcus</i> to antimicrobial agents when used as instructed.</p> |

| | | |
|-------------------------------------|--|--|
| Test Methodology | Automated quantitative antimicrobial susceptibility test for use with the VITEK [®] 2 Systems to determine the <i>in vitro</i> susceptibility of microorganisms | Automated quantitative antimicrobial susceptibility test for use with the VITEK [®] 2 and VITEK [®] 2 Compact Systems to determine the <i>in vitro</i> susceptibility of microorganisms |
| Inoculum | Saline suspension of organism | Saline suspension of organism |
| Test Card | VITEK [®] 2 <i>Streptococcus</i> Susceptibility Test Card | VITEK [®] 2 <i>Streptococcus</i> Susceptibility Test Card |
| Differences | | |
| Antimicrobial Agent | Cefuroxime | Penicillin |
| Instrument | VITEK [®] 2 Systems (VITEK [®] 2, VITEK [®] 2 Compact, and VITEK [®] COMPACT PRO) | VITEK [®] 2 Systems (VITEK [®] 2 and VITEK [®] 2 Compact) |
| Antimicrobial Concentrations | 0.25, 0.5, 1, 2 | 0.06, 0.12, 0.5, 2 |
| Analysis Algorithms | Growth Pattern Analysis | Discriminant Analysis |
| Base Broth | Modified ST1 | ST4 |
| Tested Species | <i>Streptococcus pneumoniae</i> | Beta hemolytic streptococci groups C and G <i>Streptococcus pyogenes</i> <i>Streptococcus agalactiae</i> <i>Streptococcus viridans</i> group <i>Streptococcus pneumoniae</i> |
| Breakpoints (S/I/R) | <i>S. pneumoniae</i> (oral) [≤0.5/ 1 / ≥2] <i>S. pneumoniae</i> (parenteral) [≤1/ 2 / ≥4] | <i>S. pneumoniae</i> (non-meningitis) [≤2 / 4 / ≥8] <i>S. pneumoniae</i> (meningitis) [≤0.06/ - / ≥0.12] Viridans Streptococci [≤0.12/ 0.25-2 / ≥4] Beta-hemolytic <i>Streptococcus</i> [≤0.12/ - / -] <i>S. pneumoniae</i> (oral/non-meningitis) [≤0.06/ 0.12-1 / ≥2] |

F. Intended Use/Indications for Use:

VITEK[®] 2 AST-*Streptococcus* Cefuroxime (≤ 0.125 - ≥ 8) is designed for antimicrobial susceptibility testing of *Streptococcus* species and is intended for use with the VITEK[®] 2 Systems as a laboratory aid in the determination of in vitro susceptibility to antimicrobial agents.

VITEK[®] 2 AST- *Streptococcus* Cefuroxime (≤ 0.125 - ≥ 8) is a quantitative test. Testing is indicated for *Streptococcus pneumoniae* as recognized by the FDA Susceptibility Test Interpretive Criteria (STIC).

VITEK[®] 2 AST- *Streptococcus* Cefuroxime (≤ 0.125 - ≥ 8) has demonstrated acceptable performance with the following organism:

Streptococcus pneumoniae

The VITEK[®] 2 Antimicrobial Susceptibility Test (AST) is intended to be used with the VITEK[®] 2 Systems in clinical laboratories as an in vitro test to determine the susceptibility of *Streptococcus pneumoniae*, beta-hemolytic *Streptococcus* and Viridans *Streptococcus* to antimicrobial agents when used as instructed.

G. Performance Overview:

VITEK[®] 2 AST-*Streptococcus* Cefuroxime demonstrated substantially equivalent performance when compared with the CLSI broth microdilution reference method, as defined in the FDA Class II Special Controls Guidance Document: Antimicrobial Susceptibility Test (AST) Systems; Guidance for Industry and FDA (Issued August 28, 2009).

The Premarket Notification (Traditional 510[k]) presents data in support of VITEK[®] 2 AST-*Streptococcus* Cefuroxime. An external evaluation was conducted with fresh and stock clinical isolates, as well as a set of challenge strains. The external evaluations were designed to confirm the acceptability of VITEK[®] 2 AST-*Streptococcus* Cefuroxime by comparing its performance with the CLSI broth microdilution reference method incubated for 24 hours. The data is representative of performance on the VITEK[®] 2 Systems.

VITEK[®] 2 AST-*Streptococcus* Cefuroxime (parenteral) demonstrated acceptable performance of 95.5% overall Essential Agreement and 97.7% overall Category Agreement with the reference method. VITEK[®] 2 AST-*Streptococcus* Cefuroxime (oral) demonstrated acceptable performance of 95.5% overall Essential Agreement and 95.5% overall Category Agreement with the reference method.

Limitation:

Due to the occurrence of very major errors, perform an alternative method of testing prior to reporting of results for the following antibiotic/organism combination:

Cefuroxime (cxm01n): *S. pneumoniae* when the VITEK[®] MIC is 0.5 $\mu\text{g/mL}$ (based on parenteral breakpoints)

| Antimicrobial | Antimicrobial Code ¹ | Antibiotic Version | Bp ² | Comment | Essential Agreement | | | | Category Agreement | | | | % Reproducibility |
|---------------|---------------------------------|--------------------|---|--|---------------------|-----|-----|-----|--------------------|----------------|----------------|-----------------|-------------------|
| | | | | | % Error | | | | % Error | | | | |
| | | | | | % EA | VME | ME | mE | % CA | VME | ME | mE | |
| Cefuroxime | CXM | cxm01n | CLSI (FDA) | #, <i>E. Streptococcus pneumoniae</i> (Parenteral) | (337/353) 95.5 | N/A | N/A | N/A | (345/353) 97.7 | (5/103) 4.9 | (0/247) 0.0 | (3/353) 0.8 | 98.1 |
| | | | CLSI (FDA) | #, <i>E. Streptococcus pneumoniae</i> (oral) | (337/353) 95.5 | N/A | N/A | N/A | (337/353) 95.5 | (1/95) 1.1 | (0/250) 0.0 | (15/353) 4.2 | |
| | | | VITEK [®] 2 Cefuroxime MIC values tended to be in exact agreement or at least one doubling dilution lower when testing <i>Streptococcus pneumoniae</i> by VITEK [®] 2 auto-dilution method compared to the CLSI reference broth microdilution method. | | | | | | | | | | |

¹The VITEK[®] 2 AST antimicrobial code may not match the BCI host code. Please contact your local bioMérieux representative for correct BCI code.

²Abbreviations — Bp = breakpoint committee; EA = essential agreement; CA = category agreement; VME = Very Major Error (susceptible result with resistant reference result); ME = Major Error (resistant result with susceptible reference result); mE = minor Error (susceptible or resistant result with an intermediate reference result, or an intermediate result with a susceptible or resistant reference result).

= US Food and Drug Administration
510(k) cleared
CLSI = Clinical and Laboratory Standards Institute
E = External performance data
N/A = Not applicable

Reproducibility and Quality Control demonstrated acceptable results.

H. Conclusion:

The performance data presented in this submission support a substantial equivalence decision. VITEK[®] 2 AST-*Streptococcus* Cefuroxime ($\leq 0.125 - \geq 8$ µg/mL) is substantially equivalent to VITEK[®] 2 AST-*Streptococcus* Penicillin (K232201).

References:

1. MacLowry, J.D. and Marsh, H.H., Semi-automatic Microtechnique for Serial Dilution Antibiotic Sensitivity Testing in the Clinical laboratory, *Journal of Laboratory Clinical Medicine*, 72:685-687, 1968.
2. Gerlach, E.H., Microdilution 1: A Comparative Study, p. 63-76. *Current Techniques for Antibiotic Susceptibility Testing*. A. Balows (ed.), Charles C. Thomas, Springfield, IL, 1974.
3. Barry, A.L., *The Antimicrobial Susceptibility Test, Principles and Practices*, Lea and Febiger, Philadelphia, PA, 1976.