



July 31, 2026

Olympus Medical Systems Corporation
% Deborah Kipp
Program Manager, Regulatory Affairs
Olympus Corporation of the Americas
800 W. Park Dr.
Westborough, Massachusetts 01581

Re: K261183

Trade/Device Name: EVIS EUS Ultrasound Gastrointestinal Videoscope (Olympus GF-UE190)
Regulation Number: 21 CFR 876.1500
Regulation Name: Endoscope And Accessories
Regulatory Class: Class II
Product Code: ODG, ITX
Dated: June 30, 2026
Received: June 30, 2026

Dear Deborah Kipp:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality Management System Regulation (QMSR) (21 CFR Part 820), which includes, but is not limited to, ISO 13485 clause 7.3 (Design controls), ISO 13485 clause 8.3 (Nonconforming product), ISO 13485 clause 8.5.2 (Corrective action), and ISO 13485 clause 8.5.3 (Preventative action). Please note that regardless of whether a change requires premarket review, the QMSR requires device manufacturers to review and approve changes to device design and production (ISO 13485 clause 7.3 and ISO 13485 clause 7.5) and document changes and approvals in the Medical Device File (ISO 13485 clause 4.2.3).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the Quality Management System Regulation (QMSR) (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory->

[assistance/contact-us-division-industry-and-consumer-education-dice](#)) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

SIVAKAMI VENKATACHALAM -S

for

Shanil P. Haugen, Ph.D.

Assistant Director

DHT3A: Division of Renal, Gastrointestinal,
Obesity and Transplant Devices

OHT3: Office of Gastrorenal, ObGyn,

General Hospital, and Urology Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Please type in the marketing application/submission number, if it is known. This textbox will be left blank for original applications/submissions.

K261183

?

Please provide the device trade name(s).

?

EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE (OLYMPUS GF-UE190)

Please provide your Indications for Use below.

?

The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 has been designed to be used with diagnostic ultrasound system, video system center, light source, documentation equipment, monitor, endoscopic therapy accessories and other ancillary equipment.

The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 is designed for endoscopic real-time ultrasound imaging, and other endoscopic procedures within the upper gastrointestinal tract and surrounding organs.

Operator qualifications

Appropriately trained healthcare professionals.

Device use settings

Healthcare facility

Please select the types of uses (select one or both, as applicable).

- ☒ Prescription Use (Part 21 CFR 801 Subpart D)
☐ Over-The-Counter Use (21 CFR 801 Subpart C)

?

Date Prepared: 04/10/2026

510(k) Summary**1. GENERAL INFORMATION**

- 510(k) Submitter: OLYMPUS MEDICAL SYSTEMS CORP.
2951 Ishikawa-cho, Hachioji-shi, Tokyo 192-8507, Japan
- Contact Person: Deborah Kipp
Olympus Corporation of the Americas
3500 Corporate Parkway PO Box 610
Center Valley, PA 18034-0610, USA
Phone: 402-278-4109
Fax: 484-896-7128
Email: deborah.kipp@olympus.com
- Manufacturing site: Shirakawa Olympus Co., Ltd.,
3-1 Oaza-Odakura-Aza-Okamiyama, Nishigo-mura, Nishishirakawa-gun,
Fukushima 961-8061, Japan

2. DEVICE IDENTIFICATION

Trade/Device Name	EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS
Model Name	GF-UE190
Common Name	Gastroscope and accessories, flexible/rigid
Regulation Number	21 CFR 876.1500
Regulation Name	Endoscope and accessories
Regulatory Class	II
Product Code	ODG (Endoscopic Ultrasound System, Gastroenterology-urology) ITX (Transducer, ultrasonic, diagnostic)
Classification Panel	Gastroenterology-Urology

3. PREDICATE DEVICE / REFERENCE DEVICE

Predicate Device Name	510(k) Submitter	510(k) No.
EVIS EUS ULTRASOUND GASTROINTENSTINAL VIDEOSCOPE OLYMPUS GF-UE190	OLYMPUS MEDICAL SYSTEMS CORP.	K251859

4. DEVICE DESCRIPTION

The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 has been designed to be used with diagnostic ultrasound system, video system center, light source, documentation equipment, monitor, endoscopic therapy accessories and other ancillary equipment.

The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 is designed for endoscopic real-time ultrasound imaging, and other endoscopic procedures within the upper gastrointestinal tract and surrounding organs.

5. PRINCIPLE OF OPERATION

The GF-UE190 has been designed to be used with diagnostic ultrasound system, video system center, light source, documentation equipment, monitor and endoscopic therapy accessories and other ancillary equipment. This instrument is designed for endoscopic real-time ultrasound imaging, and other endoscopic procedures within the upper gastrointestinal tract, anus, rectum, and surrounding organs.

The Control Section

The RIGHT/LEFT and UP/DOWN angulation control knob on the control section is connected to the tip of the tip of the bending section by a series of wires. By operating the RIGHT/LEFT and UP/DOWN angulation control knob, the bending section at the distal end bends vertically to guide the distal end for insertion and observation within the gastrointestinal tract.

The control section contains two cylinders: the suction cylinder and the air/water cylinder. The suction valve is attached to the suction cylinder and the air/water valve is attached to the air/water cylinder. When suction valve is depressed to the first stage to activate suction, the suction valve is operated to remove any fluids, debris, flatus, or air from the patient. The suction valve is depressed completely to activate suction of the water from the balloon through the balloon channel. The hole in the air/water valve is covered to insufflate air and the air/water valve is depressed to the first stage to feed water for lens washing. It also can be used to feed air to remove any fluid or debris adhering to the objective lens. The air/water valve is depressed completely to fill the balloon with sterile water through the balloon channel.

The Insertion Section

The insertion section has main parts including the image sensor, light guides that bring light from the light source through the endoscope, instrument channel where therapeutic tools can be pushed in and out (also the suction channel), and the ultrasound transducer that emits and receives ultrasound waves. The light from the light source

travels through the light guide to the light guide lens at the distal end. The light source can offer both the white light for normal observation and the narrow band imaging (NBI). The image sensor transduces the incident light from the objective lens to electrical signal. The video processor transduces electrical signal to video signal.

The Endoscope Connector Section

The connector section has the endoscope connector that connects the endoscope with the light source (CLV-190 or CV-1500). The connector section also has the ultrasound cable connector. This ultrasound cable connector connects the endoscope with the Olympus ultrasound center through the ultrasound cable. The Ultrasound (US) Connector cap attached to the ultrasound cable connector protects the endoscope from water penetration during reprocessing.

The subject device has been designed to be used with the following systems:

- ALOKA ARIETTA 850 : K183456
- EU-ME3: K243502
- Aplio i800: K233195
- EVIS EXERA III 190 System: K112680, K121959
- EVIS X1 1500 System: K222584

6. INDICATIONS FOR USE

The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 has been designed to be used with diagnostic ultrasound system, video system center, light source, documentation equipment, monitor, endoscopic therapy accessories and other ancillary equipment.

The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 is designed for endoscopic real-time ultrasound imaging, and other endoscopic procedures within the upper gastrointestinal tract and surrounding organs.

Operator qualifications

Appropriately trained healthcare professionals.

Device use settings

Healthcare facility

INDICATIONS FOR USE COMPARISON

The indications for use of the subject device are identical to the EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 cleared under K251859.

7. COMPARISON OF TECHNOLOGY CHARACTERISTICS WITH THE PREDICATE DEVICE

The subject and predicate devices have the same design, materials, principles of operation, and energy source. The subject device has the same technological characteristics as the device cleared under K251859.

A side-by-side comparison of the subject device and the predicate device is provided below.

Item	Subject Device (SD)	Predicate Device (PD)
General Information		
Model	EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190	EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190
Company	OLYMPUS MEDICAL SYSTEMS CORP.	OLYMPUS MEDICAL SYSTEMS CORP.
510(k) Number	-	K251859
Indications for use	The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 has been designed to be used with diagnostic ultrasound system, video system center, light source, documentation equipment, monitor, endoscopic therapy accessories and other ancillary equipment. The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 is designed for endoscopic real-time ultrasound imaging, and other endoscopic procedures within the upper gastrointestinal tract and surrounding organs.	The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 has been designed to be used with diagnostic ultrasound system, video system center, light source, documentation equipment, monitor, endoscopic therapy accessories and other ancillary equipment. The EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 is designed for endoscopic real-time ultrasound imaging, and other endoscopic procedures within the upper gastrointestinal tract and surrounding organs.

Item	Subject Device (SD)	Predicate Device (PD)
	<u>Operator qualifications</u> Appropriately trained healthcare professionals. <u>Device use settings</u> Healthcare facility	<u>Operator qualifications</u> Appropriately trained healthcare professionals. <u>Device use settings</u> Healthcare facility
Regulation name	Endoscope and accessories	Endoscope and accessories
Regulation number	876.1500	876.1500
Product Code	ODG, ITX	ODG, ITX
Ultrasound Information		
Transducer Frequency	7-10 MHz (Nominal ultrasound frequencies of the transducer assembly)	7-10 MHz (Nominal ultrasound frequencies of the transducer assembly)
Mode	Used with ALOKA ARIETTA 850 B, M, PWD, Color Doppler, Combined, Other.	Used with EU-ME3 B, PWD, Color Doppler, Combined, Other Used with Aplio i800 B, M, PWD, Color Doppler, Combined, Other
Acoustic Output	I _{SPTA} :720 mW/cm ² or less MI :1.9 or less	I _{SPTA} :720 mW/cm ² or less MI :1.9 or less

Item	Subject Device (SD)	Predicate Device (PD)
Scanning Direction	Perpendicular to the insertion direction	Perpendicular to the insertion direction
Scanning Field of View [deg]	360	360
Axial resolution	1mm or less	1mm or less
Lateral resolution	3mm or less	3mm or less
Depth of penetration	Used with ARIETTA 850: 85mm or more	Used with EU-ME3: 90mm or more Used with Aplio i800: 85mm or more
Scanning Method	Electronic radial array	Electronic radial array
Contact Method	Balloon method, sterile de-aerated water immersion method	Balloon method, sterile de-aerated water immersion method
Endoscope Specifications		
Field of View	100 °	100 °
Depth of Field	3 – 100 mm	3 – 100 mm
Direction of View	50° (forward-oblique)	50° (forward-oblique)
Optimum Working Distance	14 mm	14 mm
Type of CCD Chip	Color	Color
Total Number of Pixels	182,484	182,484
Pixels per square mm	226,757	226,757
Size of Pixel	2.1µm(H) x 2.1µm(V)	2.1µm(H) x 2.1µm(V)
Active Area of CCD Chip	865µm(H) x 857µm(V)	865µm(H) x 857µm(V)
Outer Diameter of Distal End	ø 13.4 mm	ø 13.4 mm

Item	Subject Device (SD)	Predicate Device (PD)
Maximum distal end outer diameter	ø 14.7 mm	ø 14.7 mm
Outer Diameter of Insertion Tube	ø 10.9 mm	ø 10.9 mm
Angulation UP/DOWN /LEFT/RIGHT [deg]	130/90/90/90	130/90/90/90
Working Length	1250 mm	1250 mm
Inner Diameter of Instrument Channel	Φ2.2mm	Φ2.2mm
Combination use with Electrosurgical instruments	Not applicable	Not applicable
Distal end	Stainless	Stainless
A rubber adhesive	ES-8SC	ES-8SC
Ultrasound Cable	The ultrasound cable (MAJ-2056 or UITS-AI800A) is separated from ultrasound connector.	The ultrasound cable (MAJ-2056 or UITS-AI800A) is separated from ultrasound connector.
Scope connector	Electrical connector is integrated to scope connector. Ultrasound connector is separated. When combined with the water-resistant cap, ultrasound connector is water resistant.	Electrical connector is integrated to scope connector. Ultrasound connector is separated. When combined with the water-resistant cap, ultrasound connector is water resistant.
Protect water penetration during reprocessing.	The US connector cap (MAJ-2295) is attached to the ultrasound cable connector on the endoscope to protect the ultrasound cable connector and the endoscope from water penetration during reprocessing.	The US connector cap (MAJ-2295) is attached to the ultrasound cable connector on the endoscope to protect the ultrasound cable connector and the endoscope from water penetration during reprocessing.
Sterilization methods for reprocessing endoscope	Ethylene oxide gas	Ethylene oxide gas
High-frequency accessories	Not available	Not available

Item	Subject Device (SD)	Predicate Device (PD)
RFID tag for communication with endoscope reprocessor	Available	Available
Individual scope information (Scope ID)	Available	Available

8. PERFORMANCE DATA

The following performance data were provided in support of the substantial equivalence determination for GF-UE190:

1) Performance testing - Bench

Bench testing as listed below was conducted to ensure that the subject device performs as intended and meet design specifications.

- Thermal Safety
- Acoustic Output
- Clinical Accuracy
- Elastography Function
- Shear Wave Speed Quantification Measurement
- Transducer Element Check

2) Performance testing - Animal

Animal testing was not performed.

3) Risk management

Risk management was performed in accordance with ISO 14971:2019. The design verification tests and their acceptance criteria were performed and identified as a result of this risk management.

9. CONCLUSIONS

Based on the indications for use, technological characteristics, performance testing and technological comparison to the predicate device, EVIS EUS ULTRASOUND GASTROINTESTINAL VIDEOSCOPE OLYMPUS GF-UE190 raises no new issue of safety and effectiveness and is substantially equivalent to the predicate device in terms of safety, effectiveness, and performance.