

SUMMARY OF SAFETY AND EFFECTIVENESS

FEB 05 2003

NAME OF FIRM: DePuy Orthopaedics, Inc.
P.O. Box 988
700 Orthopaedic Drive
Warsaw, IN 46581-0988

510(k) CONTACT: Cheryl Hastings
Director, Regulatory Affairs

TRADE NAME: DePuy Summit Basic Press-Fit Hip Prosthesis

COMMON NAME: Total Hip Joint Replacement Prosthesis

CLASSIFICATION: 888.3353: Hip joint metal/ceramic/polymer semi-constrained cemented or non-porous uncemented prosthesis; Class II

DEVICE PRODUCT CODE: 87 LZ0

SUBSTANTIALLY EQUIVALENT DEVICE: DePuy Response 2000 Hip Stems, K000392

DEVICE DESCRIPTION:

The Summit Basic Press-Fit Hip Stem is a collared, tapered Titanium femoral stem with a grit blasted finish. The Summit Basic Press-Fit Hip Stem is offered in 7 sizes with a constant neck offset. The stem is intended for cementless, press-fit fixation and is designed specifically to treat femoral head and neck fractures but can be used for any of the indications listed below.

INTENDED USE AND INDICATIONS:

Total hip arthroplasty is intended to provide increased patient mobility and reduce pain by replacing the damaged hip joint articulation in patients where there is evidence of sufficient sound bone to seat and support the components. Total hip replacement is indicated in the following conditions:

- A severely painful and/or disabled joint from osteoarthritis, traumatic arthritis, rheumatoid arthritis, or congenital hip dysplasia;
- Avascular necrosis of the femoral head;
- Acute traumatic fracture of the femoral head or neck;
- Failed previous hip surgery including joint reconstruction, internal fixation, arthrodesis, hemiarthroplasty, surface replacement arthroplasty, or total hip replacement;
- Certain cases of ankylosis.

BASIS OF SUBSTANTIAL EQUIVALENCE:

The Summit Basic Press-Fit Hip Stem has the same basic design and has the same intended use as the Response 2000 Hip Stem. The Summit Basic Press-Fit Hip Stem is manufactured from Ti-4V-6Al alloy while the Response 2000 Hip Stem is manufactured from Co-Cr-Mo alloy. The designs are similar – both stems are available in a single offset, but the Summit Basic Press-Fit Hip Stem is offered in only one stature while the Response 2000 Hip Stem is offered in large and small statures.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

FEB 05 2003

Ms. Cheryl Hastings
Director, Regulatory Affairs
DePuy Orthopaedics, Inc.
P.O. Box 988
700 Orthopaedic Drive
Warsaw, Indiana 46581-0988

Re: K030122

Trade/Device Name: DePuy Summit Basic Press-Fit Hip Prosthesis
Regulation Number: 21 CFR 888.3350
Regulation Name: Hip Joint Metal/Polymer Semi-constrained Cemented Prosthesis
Regulatory Class: Class II
Product Code: LWJ
Dated: January 13, 2003
Received: January 14, 2003

Dear Ms. Hastings:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21

Page 2 – Ms. Cheryl Hastings


CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address

<http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

A handwritten signature in black ink, appearing to read "Celia M. Witten". The signature is written in a cursive style with a large initial "C" and a long horizontal flourish at the end.

Celia M. Witten, Ph.D., M.D.

Director

Division of General, Restorative and
Neurological Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

510(k) Number (if known) K030122

Device Name DePuy Summit Basic Press-Fit Hip Stem


Intended Use and Indications:

Total hip arthroplasty is intended to provide increased patient mobility and reduce pain by replacing the damaged hip joint articulation in patients where there is evidence of sufficient sound bone to seat and support the components.

Total hip replacement is indicated in the following conditions:

- A severely painful and/or disabled joint from osteoarthritis, traumatic arthritis, rheumatoid arthritis, or congenital hip dysplasia;
- Avascular necrosis of the femoral head;
- Acute traumatic fracture of the femoral head or neck;
- Failed previous hip surgery including joint reconstruction, internal fixation, arthrodesis, hemiarthroplasty, surface replacement arthroplasty, or total hip replacement;
- Certain cases of ankylosis.

Concurrence of CDRH, Office of Device Evaluation


(Division Sign-Off)
Division of General Restorative
and Neurological Devices

510(k) Number K030122

Prescription Use Yes
(Per 21 CFR 801.109)

OR

Over-The Counter Use NO

0000004