

MAY - 8 2003

---

**I. 510(K) SUMMARY OF SAFETY AND EFFECTIVENESS**

---

510(k) Summary Of Safety and Effectiveness**I. General Information**

This Summary of Safety and Effectiveness information is being submitted in accordance with the requirements of the SMDA of 1990 and 21 § 807.92

Establishment:

- Address: BD Vacutainer Systems, Preanalytical Solutions  
1 Becton Drive  
Franklin Lakes, NJ 07417-1885
- Registration Number: 2243072
- Contact Person: M. Wendy Ballesteros  
Regulatory Affairs Specialist  
Telephone no.: 201-847-6280  
Fax No. 201-847-4858
- Date of Summary: February 21, 2003

Device

- Trade Name: BD Vacutainer™ Push Button Blood Collection Set
- Classification Name: Tubes, Vials, Systems, Serum Separators, Blood Collection
- Classification: Class II
- Performance Standards: None Established under 514 of the Food, Drug and Cosmetic Act

## II. Safety and Effectiveness Information Supporting Substantial Equivalence

- Device Description

The BD Vacutainer™ Push Button Blood Collection Set is for venous blood collection and IV administration. It contains a needle that will retract into the body of the device when a button is depressed, helping to prevent accidental needle sticks.

- Intended Use

The BD Vacutainer™ Push Button Blood Collection Set is a sterile, multiple-sample, single-use winged blood collection set intended for venipuncture to obtain blood specimens from patients.

The BD Vacutainer™ Push Button Blood Collection Set is also indicated for the intravenous administration of fluids as indicated in 21 CFR §880.5440. It may be used for any patient population with consideration given to patient size and appropriateness for the solution being infused and duration of therapy.

The recommended use of the device is to activate the needle prior to removal from the venipuncture site. The retraction of the intravenous (IV) end of the needle aids in the prevention of accidental needlestick injury.

- Synopsis of Performance Study Results

Based upon previously demonstrated performance and successful completion of biocompatibility testing, Push Button Blood Collection Set will perform as intended.

## III. Predicate Device Summary Table

- Substantial Equivalence

Based on comparison of the device features, materials, intended use and performance, the BD Vacutainer™ Push Button Blood Collection Set is shown to be substantially equivalent to the commercially available predicate devices indicated in the table below. The predicate devices, K number, and clearance date are also identified in the table below.

Manufacturer	Predicate Device	K-Number	Clearance Date
BD Vacutainer Systems, Preanalytical Solutions	BD VACUTAINER™ Brand Safety-Lok™ Blood Collection Set	K980414	March 3, 1998
BD Vacutainer Systems, Preanalytical Solutions	BD Vacutainer™ Push Button Blood Collection Set	K022875	September 11, 2002

M. Wendy Ballesteros

M. Wendy Ballesteros  
 Regulatory Affairs Specialist  
 BD Vacutainer Systems, Preanalytical Solutions  
 Becton Dickinson and Company

2/21/03

Date



**MAY - 8 2003**

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

BD Vacutainer Systems, Preanalytical Solutions  
Ms. M. Wendy Ballesteros  
Regulatory Affairs Specialist  
1 Becton Drive  
Franklin Lakes, New Jersey 07417-1885

Re: K030573

Trade/Device Name: BD Vacutainer™ Push Button Blood Collection Set  
Regulation Number: 21 CFR 862.1675, 21 CFR 880.5440  
Regulation Name: Blood Specimen Collection Device, Intravascular  
Administration Set  
Regulatory Class: II  
Product Code: 75 JKA, 80 FPA  
Dated: February 21, 2003  
Received: February 24, 2003

Dear Ms. Ballesteros:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4618. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Susan Runner, DDS, MA  
Interim Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

---

**B. INDICATIONS FOR USE**

---

510(k) Number (if known): K030573

Device Name: BD Vacutainer™ Push Button Blood Collection Set

**Indications for Use:**

The BD Vacutainer™ Push Button Blood Collection Set is a sterile, multiple-sample, single-use winged blood collection set intended for venipuncture to obtain blood specimens from patients.

The BD Vacutainer™ Push Button Blood Collection Set is also indicated for intravenous administration of fluids as indicated in 21 CFR §880.5440. It may be used for any patient population with consideration given to patient size and appropriateness for the solution being infused and duration of therapy.

The recommended use of the device is to activate the needle prior to removal from the venipuncture site. The retraction of the intravenous (IV) end of the needle aids in the prevention of accidental needlestick injury.

(Please do not write below this line-continue on another page if needed)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use  Or Over-the-Counter Use

(Per 21 CFR § 801.109)

*Patricia Cuervo*  
(Division Sign-Off)  
Division of Anesthesiology, General Hospital,  
Infection Control, Dental Devices

(Optional format 1-2-96)

510(k) Number: K030573