



Food and Drug Administration  
2098 Gaither Road  
Rockville MD 20850

**JUL 11 2003**

David M. Lyerly, Ph.D.  
Vice-President of Research and Development  
TECHLAB, Inc.  
VPI Research Park  
1861 Pratt Drive, Suite 1030  
Blacksburg, VA 24060-6364

Re: k030991  
Trade/Device Name: C. Diff Check™ - 30  
Regulation Number: 21 CFR 866.2660  
Regulation Name: Microorganism Differentiation and Identification Device  
Regulatory Class: Class I  
Product Code: LLH  
Dated: June 20, 2003  
Received: June 23, 2003

Dear Dr. Lyerly:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820).

Page 2 –

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (301) 594-3084. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>.

Sincerely yours,



Steven I. Gutman, M.D., M.B.A.  
Director  
Office of *In Vitro* Diagnostic Device  
Evaluation and Safety  
Center for Devices and  
Radiological Health

Enclosure

2. STATEMENT OF INTENDED USE

510(k) Number (if known): K030991 ~~Not known~~

Device Name: C. DIFF CHEK™ - 30


Indications for Use

The C. DIFF CHEK™ - 30 test is an enzyme immunoassay for use as a screening test to detect *Clostridium difficile* antigen, glutamate dehydrogenase, in fecal specimens from persons suspected of having *C. difficile* disease. The test does not distinguish toxigenic from non toxigenic strains of *C. difficile*. With the use of additional tests that detect *C. difficile* toxins, the test is to be used as an aid in the diagnosis of *C. difficile* disease. As with other *C. difficile* tests, results should be considered in conjunction with the patient history. FOR IN VITRO DIAGNOSTIC USE.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

\_\_\_\_\_ Concurrence of CDRH, Office of Device Evaluation (ODE) \_\_\_\_\_

Prescription Use  \_\_\_\_\_ OR Over-The Counter Use \_\_\_\_\_  
(Per 21 CFR 801.109)

 (Optional format 1-2-96)  
Division Sign-Off

Office of In Vitro Diagnostic Device  
Evaluation and Safety

510(k) K030991 \_\_\_\_\_