

**JUL 30 2003**

**ENCLOSURE L**

K 031911

Premarket Notification (510(k) - Mercury Medical Mini StatCO<sub>2</sub><sup>TM</sup> End Tidal CO<sub>2</sub> Detector

**Summary of Safety and Effectiveness**

**Non-Confidential Summary of Safety and Effectiveness**

Page 1 of 2  
April 28, 2003

**Mercury Enterprises, Inc./Mercury Medical  
11300 49th St. N.  
Clearwater, FL 33762  
Tel: (800) 237-6418  
Fax: (727) 572-4501**

**Official Contact: Wayne Glover  
Manager, QA/RA**

**Proprietary or Trade Name: Mini StatCO<sub>2</sub><sup>TM</sup> Detector**

**Common/Usual Name: End Tidal CO<sub>2</sub>**

**Classification: Class II, CCK, 21 CFR 868.1400**

**Classification Name: Analyzer, Gas, Carbon Dioxide, Gaseous Phase**

**Device: Mini StatCO<sub>2</sub><sup>TM</sup> Detector**

**Predicate Devices: Nellcor Pedi-CAP<sup>TM</sup> (K944400)**

**Device Description:**

**The Mercury Medical Mini StatCO<sub>2</sub><sup>TM</sup> uses the same technology as the Nellcor Pedi-CAP<sup>TM</sup> (K944400). The Mini StatCO<sub>2</sub><sup>TM</sup> End Tidal CO<sub>2</sub> Detector is a "Single Patient Use" only device. The device assists in verification of tube placement during endotracheal or nasotracheal intubation. It may be used on intubated patients to detect approximate ranges of end tidal CO<sub>2</sub> when clinically significant.**

## ENCLOSURE L

Premarket Notification (510(k) - Mercury Medical Mini StatCO<sub>2</sub><sup>™</sup> End Tidal CO<sub>2</sub> Detector

### Non-Confidential Summary of Safety and Effectiveness

Page 2 of 2  
April 28, 2003

**The device consists of a housing unit made with material of polycarbonate, filter media consisting of 3M Filtrete GSB30 and Bregas C-it Indicator, 510(k) K000520 mounted to the housing.**

#### **Intended Use:**

**Indicated Use:** The Mercury Medical Mini StatCO<sub>2</sub><sup>™</sup> End Tidal CO<sub>2</sub> Detector is to provide a semi-quantitative visualization of the CO<sub>2</sub> in the patient airway. It is an adjunct in patient assessment, to be used in conjunction with other methods to determine clinical signs and symptoms by or on the order of a physician.

**Technical Characteristics:** The device has the same technical characteristics as the predicate device marketed by Nellcor, Pedi-CAP<sup>™</sup>.

**Non-Clinical Data:** Performance and specifications of the modified device are consistent with all requirements for this device type specified by ISO 5356-1: 1987 – Anesthetic and Respiratory Equipment-Conical connectors-Part 1: Cones and Sockets. ASTM F1054 – Standard Specification for Conical Fittings of 15mm.

**Environment of Use:** Hospital/Transport

**Conclusions:** The comparison to the predicate devices demonstrates that the proposed device is safe and effective and is substantially equivalent to the predicate devices.



JUL 30 2003

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Mr. Wayne Glover  
Manager, Quality Assurance and Regulatory Affairs  
Mercury Medical  
11300 49<sup>th</sup> Street North  
Clearwater, FL 33762

Re: K031411  
Trade/Device Name: Mini StatCO<sub>2</sub><sup>TM</sup> CO<sub>2</sub> Detector  
Regulation Number: 21 CFR 868.1400  
Regulation Name: Carbon dioxide gas analyzer  
Regulatory Class: II  
Product Code: CCK  
Dated: April 28, 2003  
Received: May 5, 2003

Dear Mr. Glover:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2 – Mr. Wayne Glover

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4646. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>.

Sincerely yours,



Susan Runner, DDS, MA  
Interim Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

ENCLOSURE B

Indications for Use

Page 1 of 1


510(k) Number:     K031411     (To be assigned)

Device Name: **Mini StatCO<sub>2</sub><sup>TM</sup>**

Intended Use: The Mercury Medical Mini StatCO<sub>2</sub><sup>TM</sup> End Tidal CO<sub>2</sub> Detector is to provide a semi-quantitative visualization of the CO<sub>2</sub> in the patient airway. It is an adjunct in patient assessment, to be used in conjunction with other methods to determine clinical signs and symptoms by or on the order of a physician.

---

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)  
Division of Anesthesiology, General Hospital,  
Infection Control, Dental Devices

510(k) Number:     K031411    

Prescription Use   X    
(Per CFR 801.109)

or

Over-the-counter use