

OCT 24 2003

3 510(K) SUMMARY

K032433

510(k) SUMMARY—Nasal Jacks Mask

**Submitter Name:** ResMed Corp.

**Submitter Address:** 14040 Danielson Street, Poway CA 92064-6857  
USA

**Contact Person:** David D'Cruz, VP US Clinical & Regulatory Affairs

**Phone Number:** (858) 746 2238

**Fax Number:** (858) 746 2900

**Date Prepared:** August 1, 2003

**Device Trade Name:** Not available at time of submission

**Device Common Name/  
Classification Name:** Nasal Jacks Mask

**Predicate Devices:** Primary predicate: ADAM™ Interface System. K900164  
Secondary predicate: Mirage Nasal Mask (cleared in  
SULLIVAN® AUTOSET® Nasal CPAP System K980721)

**Device Description:**

The Nasal Jacks Mask is designed for adult patients for the delivery of non-invasive ventilatory support using continuous positive airway pressure or bi-level therapy. It is intended for single-patient, multi-use and is minimally obtrusive to the user providing a high level of comfort, ease-of-use and seal.

**Intended Use:**

The Nasal Jacks Mask is an accessory to a non-continuous ventilator (respirator), intended for single-patient multi-use for adult patients prescribed continuous positive airway pressure (CPAP) and bi-level therapy in hospital, clinical, and home environments.

## **Device Technological Characteristics and Comparison to Predicate Device(s):**

### **Overview:**

The Nasal Jacks Mask is supported by headgear to allow a seal with the patients nostrils via the nasal cushions, then connected via tubing to a CPAP or bi-level flow generator. Positive pressure ventilation is thus applied to the lungs in a non-invasive manner.

The Nasal Jacks Mask comes in one frame size and has three nasal cushion sizes.

The Nasal Jacks Mask design is substantially equivalent to predicate device(s) (ADAM™ Interface System and Mirage Nasal Mask). The Nasal Jacks Mask design has the same intended use and has the same fundamental scientific technology as its predicates.

### **Performance Data:**

Performance testing has been carried out to verify and validate the safety and effectiveness of the Nasal Jacks Mask. The Nasal Jacks Mask was tested to determine the pressure-flow characteristic, functional dead space (CO<sub>2</sub> re-breathing), physical dead space and flow impedance. The results of the performance data show that the mask is substantially equivalent to the predicate mask (refer section 5.2.1)

### **Materials Biocompatibility**

Materials have been carefully selected to ensure patient safety and efficacy of the product. The materials used to create components of the Nasal Jacks Mask, which contact the skin, mucosal membrane and/or the air-path, are either predicate materials (i.e., cleared previously for the same intended use), or have been tested to the ISO 10993 standards by an independent certified laboratory. The details are referenced in section 5.4.



OCT 24 2003

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

ResMed Limited  
C/O Mr. David D'Cruz  
Vice President US Clinical & Regulatory Affairs  
ResMed Corporation  
14040 Danielson Street  
Poway, California 92064-6857

Re: K032433  
Trade/Device Name: Nasal Jacks Mask  
Regulation Number: 868.5905  
Regulation Name: Noncontinuous Ventilator (IPPB)  
Regulatory Class: II  
Product Code: BZD  
Dated: August 1, 2003  
Received: August 11, 2003

Dear Mr. D'Cruz:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA’s issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act’s requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4646. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Sponsor:  
ResMed Ltd

Nasal Jacks Mask  
Traditional 510(k) Premarket Notification


#### 4 INDICATIONS FOR USE

510(k) Number (if known): K032433

Device Name: **Nasal Jacks Mask**

#### Indications for Use:

The Nasal Jacks Mask is an accessory to a non-continuous ventilator (respirator), intended for single-patient multiple use for adult patients prescribed continuous positive airway pressure (CPAP) and bi-level therapy in hospital, clinical, and home environments.

  
\_\_\_\_\_  
(Division Sign-Off)  
Division of Anesthesiology, General Hospital,  
Infection Control, Dental Devices

510(k) Number: K032433

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

\_\_\_\_\_  
Concurrence of CDRH; Office of Device Evaluation (ODE)

Prescription  
Use   
(Per 21 CFR 801.109)

OR

Over-The-Counter  
Use \_\_\_\_\_

(Optional Format 1-2-96)