

K032979

FEB 20 2004

510(k) Summary
(As required by 21 C.F.R. §807.92)

Submitted by: Egon Pfeil
Philips Medizin Systeme Boeblingen GmbH
Cardiac and Monitoring Systems
Hewlett-Packard Str.2
71034 Boeblingen
Germany

Date of Summary: September 19, 2003

Device Name: The Philips Reusable SpO₂ Sensors M1191T, M1192T, and M1193T.

Common Name: SpO₂ Sensor

Classification Name: Classification Name: Oximeter
Regulation Number: 21 C.F.R §870.2700

Predicate Devices: Philips M1191A, M1192A, M1193A, M1901B (Nelcor/Tyco Oxisensor II™ N-25), M1903B (Nelcor/Tyco Oxisensor II™ D-20) and M1904B (Nelcor/Tyco Oxisensor II™ D-25) disposable SpO₂ sensors cleared pursuant to K882609, 1/19/89; K990972, 4/19/99, and K000822, 4/6/00.

Modifications: The modification involves a change to the connector from the proprietary Philips version to a Nelcor compatible connector with corresponding change in sensor wave length coding from the Philips Rtype to the Nelcor Rcal.

Device Description: The Philips SpO₂ devices measure, non-invasively, the arterial oxygen saturation of blood. The measurement method is based on the red and infrared light absorption of hemoglobin and oxyhemoglobin. Light of a red and infrared light source is emitted through human tissue and received by a photodiode.

The measurement is based on the absorption of light, which is emitted through human tissue (for example through the index finger). The light comes from two sources (red LED and infrared LED) with different wavelengths and is received by a photodiode. Out of the different absorption behavior of the red and infrared light a so-called Ratio can be calculated. The saturation value is defined by the percentage ratio of the oxygenated hemoglobin [HbO₂] to the total amount of hemoglobin [Hb].

$$SpO_2 = [HbO_2]/([Hb]+[HbO_2])$$

Out of calibration curves, which are based on controlled hypoxia studies with healthy non-smoking adult volunteers over a specified saturation range (SaO₂ from 100%-70%), the Ratio can be related to a SpO₂ value.

The devices contain a red and infrared light source and a photodiode receiving the non-absorbed red and infrared light. The received

signals are forwarded to a measurement device that amplifies the acquired signal and an algorithm that calculates the ratio and converts via a validated calibration table the ratio to a saturation value.

Intended Use

The Philips Reusable SpO₂ Sensors are intended for acquiring non-invasively the arterial oxygen saturation to support the measurement of oxygen saturation.

M1191T is indicated for adult patients, M1192T is indicated for pediatric patients, and M1193T is indicated for neonatal patients.

Technological characteristics

The Philips Reusable SpO₂ Sensors have the same technological characteristics as the legally marketed predicate devices.

Testing

Verification and validation testing activities were conducted to establish the performance and reliability characteristics of the modified device.

Testing involved environmental and clinical evaluations for accuracy. Hardware verification testing and cable interface verification testing were also conducted. Test results showed substantial equivalence.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

FEB 20 2004

Mr. Egon Pfeil
Regulatory Affairs Engineer
Philips Medizin Systeme Bielefeld GmbH
Cardiac and Monitoring Systems, Incorporated
Hewlett-Packard Str.2
71034 Bielefeld
GERMANY

Re: K032979/S2

Trade/Device Name: The Philips Reusable SpO₂ Sensors M1191T,
M1192T and M1193

Regulation Number: 21 CFR 870.2700

Regulation Name: Oximeter

Regulatory Class: II

Product Code: DQA

Dated: January 27, 2004

Received: January 30, 2004

Dear Mr. Pfeil:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4646. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Chiu Lin, Ph., D.

Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure


Indications for Use Statement

510(k) Number
(if known)

Device Name: The Philips Reusable SpO₂ Sensors M1191T, M1192T, and M1193T

Indications for Use: The Philips Reusable SpO₂ Sensors are intended for acquiring non-invasively the arterial oxygen saturation to support the measurement of oxygen saturation.

M1191T is indicated for adult patients, M1192T is indicated for pediatric patients, and M1193T is indicated for neonatal patients.



(Division Sign-Off)
Division of Anesthesiology, General Hospital,
Infection Control, Dental Devices
510(k) Number: K032979

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE
IF NEEDED

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use

OR

Over-The-Counter Use _____

Page __ of __