

OCT 16 2003

510(k) Summary - COBAS Integra Albumin Gen.2

Introduction According to the requirements of 21 CFR 807.92, the following information provides sufficient detail to understand the basis for a determination of substantial equivalence

Submitter name, address, contact Roche Diagnostics Corporation
9115 Hague Rd
Indianapolis IN 46250
(317) 521-3831

Contact person: Sherri L. Coenen

Date prepared: September 24, 2003

Device Name Proprietary name: Roche Diagnostics COBAS Integra Albumin Gen.2

Common name: Albumin Assay

Classification name: Albumin test system

Device description The COBAS Integra Albumin Gen.2 is a colorimetric assay for the determination of albumin concentration in serum or plasma. At a pH value of 4.1, albumin displays a sufficiently cationic character to be able to bind with bromocresol green (BCG) to form a blue-green complex. The color intensity is directly proportional to the albumin concentration and can be determined photometrically.

Intended use The cassette COBAS Integra Albumin Gen.2 (ALB2) contains an in vitro diagnostic reagent system intended for use on COBAS Integra systems for the quantitative determination of the albumin concentration in human serum and plasma.

Predicate Device We claim substantial equivalence to the currently marketed COBAS Integra Albumin Assay. (K951595).

510(k) Summary - COBAS Integra Creatinine plus ver.2,
continued

**Reagent
Summary**

The following table describes the similarities and differences between the COBAS Integra Albumin Gen.2 and the predicate device.

Topic	COBAS Integra Albumin (K951595)	COBAS Integra Albumin Gen.2 (Modified Device)
Intended Use	The cassette COBAS Integra Albumin (ALB) contains an in vitro diagnostic reagent system intended for use on COBAS Integra systems for the quantitative determination of the albumin concentration in serum and plasma.	The cassette COBAS Integra Albumin Gen.2 (ALB2) contains an in vitro diagnostic reagent system intended for use on COBAS Integra systems for the quantitative determination of the albumin concentration in serum and plasma.
Method	BCG colorimetric assay with endpoint method	Same
Sample type	Serum Heparin or EDTA plasma	Same
Measuring range	0.08 - 60 g/L	2 - 60 g/L
Expected values	Adults (18-60 yrs): 3.5 - 5.0 g/dL Adults (> 60 yrs): 3.4 - 4.8 g/dL Newborn: 2.8 - 4.4 g/dL Children (4 days - 14 yrs): 3.8 - 5.4 g/dL Children (14 - 18 years): 3.2 - 4.5 g/dL	Adults: 3.4 - 4.8 g/dL Newborn: 2.8 - 4.4 g/dL Children (4 days - 14 yrs): 3.8 - 5.4 g/dL Children (14 - 18 years): 3.2 - 4.5 g/dL



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
2098 Gaither Road
Rockville MD 20850

OCT 16 2003

Ms. Sherri L. Coenen
Regulatory Affairs Consultant
Regulatory Submissions, Centralized Diagnostics
Roche Diagnostics Corporation
9115 Hague Road
P.O. Box 50457
Indianapolis, IN 46250-0457

Re: k033009
Trade/Device Name: COBAS Integra Albumin Gen.2
Regulation Number: 21 CFR 862.1035
Regulation Name: Albumin test system
Regulatory Class: Class II
Product Code: CIX
Dated: September 24, 2003
Received: September 26, 2003

Dear Ms. Coenen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820).

Page 2 –

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (301) 594-3084. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>.

Sincerely yours,



Steven I. Gutman, M.D., M.B.A.
Director
Office of *In Vitro* Diagnostic Device
Evaluation and Safety
Center for Devices and
Radiological Health

Enclosure

Indications for Use Statement

510(k) Number (if known): N/A K033009

Device Name: COBAS Integra Albumin Gen.2

Indications For Use:

The cassette COBAS Integra Albumin Gen.2 (ALB2) contains an in vitro diagnostic reagent system intended for use on COBAS Integra systems for the quantitative determination of the albumin concentration in human serum and plasma. Albumin measurements are used in the diagnosis and treatment of numerous diseases involving primarily the liver or kidneys.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)
Prescription Use ~~_____~~ OR Over-The-Counter Use _____
(Per 21 CFR 801.109)

(Optional Format 1-2-96)

Carol C Benson for Jean Cooper, DVM
Division Sign-Off

Office of In Vitro Diagnostic Device
Evaluation and Safety

510(k) K 033 009