

K 0 3 3 8 8 3

P. 1/4

FEB 13 2004

**Exactech Optetrak[®] Total Knee System
Line Extension-Optetrak[®] HI-FLEX Knee Components
Special 510(k)**

Summary of Safety and Effectiveness

.....

Sponsor: Exactech[®] Inc.
2320 N.W. 66th Court
Gainesville, Florida 32653
Phone: (352) 377-1140
Fax: (352) 378-2617

FDA Establishment Number: 1038671

Contact: Dr. Gary Miller
Vice President of Research and Development

Date: _____

Exactech Optetrak[®] Total Knee System
Line Extension-Optetrak[®] HI-FLEX Knee Components
Special 510(k)

Summary of Safety and Effectiveness

.....

Classifications / Proprietary Names:

Classification Name:	Prosthesis, Knee, Patellofemorotibial, Semi-Constrained, cemented, Polymer/Metal/Polymer
Trade / Proprietary Model Name:	Optetrak [®] Total Knee System <ul style="list-style-type: none"> • <i>HI-FLEX Asymmetric Posterior-Stabilized Cemented Femoral Components</i> • <i>HI-FLEX Posterior-Stabilized Tibial Insert Components</i>
Product Code:	JWH
C.F.R. Section:	888.3560
Device Class:	II
Classification Panel:	Orthopedic

Exactech Legally Marketed Devices for Substantial Equivalence Comparison:

<u>Model</u>	<u>510(k) Number</u>
Optetrak [®] Total Knee System Tibial Component	K933610
Optetrak [®] Size 0 and 1 Delta Line Extension	K011976
Optetrak [®] Total Knee Asymmetric Femoral Components	K032606
Optetrak [®] B-Series Total Knee System	K010434

Device Information:

INTENDED USE

The Optetrak[®] HI-FLEX Asymmetric Posterior-Stabilized Cemented Femoral Components (herein referred to as the Optetrak[®] HI-FLEX PS Femoral Components) and Optetrak[®] HI-FLEX Posterior-Stabilized Tibial Inserts (Herein referred to as the HI-FLEX PS Tibial Inserts) are intended to be used with an Optetrak[®] Tibial Tray to replace the patient's distal femur and proximal tibia during primary or revision total knee arthroplasty. The Optetrak[®] HI-FLEX PS Femoral components are intended for use when needed to closely

**Exactech Optetrak[®] Total Knee System
Line Extension-Optetrak[®] HI-FLEX Knee Components
Special 510(k)**

Summary of Safety and Effectiveness

.....

match the geometry of the patient's resected distal femur and allows for more flexion than the standard Optetrak[®] components.

The Optetrak[®] HI-FLEX PS Tibial Insert is intended for use with the Optetrak[®] HI-FLEX PS Femoral components to afford the patient a higher degree of flexion than the standard Optetrak[®] femoral/tibial insert combination.

The proposed components are intended for use in total knee arthroplasty procedures in which the Posterior Cruciate Ligament (PCL) must be sacrificed.

All proposed femoral components are intended for cemented use only.

INDICATIONS

The OPTETRAK[®] Total Knee Systems are indicated for use in skeletally mature individuals undergoing primary surgery for total knee replacement due to osteoarthritis, osteonecrosis, rheumatoid arthritis and/or post-traumatic degenerative problems. They are also indicated for revision of failed previous reconstructions where sufficient bone stock and soft tissue integrity are present.

CONTRAINDICATIONS

The OPTETRAK[®] Total Knee Systems are contraindicated in patients with active infection, patients without sufficient bone stock to allow appropriate insertion and fixation of the prosthesis, patients without sufficient soft tissue integrity to provide adequate stability, and in patients with either mental or neuromuscular disorders that do not allow control of the knee joint, and in patients whose weight, age, or activity level might cause extreme loads and early failure of the system.

Optetrak[®] HI-FLEX Tibial Inserts must be used with Optetrak[®] HI-FLEX Femoral Components.

CAUTION: In the USA, for cemented use only.

Device Modifications

The device modifications to the femoral components presented in this Special 510(k) represent changes to the Optetrak[®] Asymmetric (AK), Posterior-Stabilized, Cemented

**Exactech Optetrak[®] Total Knee System
Line Extension-Optetrak[®] HI-FLEX Knee Components
Special 510(k)**

Summary of Safety and Effectiveness

.....

Femoral Components (K032606) and the Optetrak[®] B-Series Posterior-Stabilized Femoral Component (K010434). These changes represent a combination of:

1. The condyle, anterior cam, asymmetric flange and PCL stabilizing box geometries of the Optetrak[®] AK Femoral components, and
2. The Posterior condyle radius and cam geometries of the Optetrak[®] B-Series Femoral Component

The device modifications to the tibial insert components presented in this Special 510(k) represent changes to the Optetrak[®] Posterior-Stabilized Tibial Inserts (K011976 and K933610) and changes to the Optetrak[®] B-series Tibial Inserts (K010434). These changes represent a combination of:

1. The articulating surfaces, mating geometry and anterior spine geometry of the Optetrak[®] Posterior-Stabilized Tibial Inserts, and
2. The posterior portion of the spine and the posterior scallops of the Optetrak[®] B-Series Tibial Insert Component

No changes were made to the patellar or tibial tray components of the Optetrak[®] Total Knee System.

PERFORMANCE DATA SUMMARY

Verification and Validation analyses were conducted to verify that the implant performance would be adequate for anticipated *in vivo* loading.


We conclude that the Optetrak[®] HI-FLEX PS Femoral Components and the Optetrak[®] HI-FLEX PS Tibial Inserts are substantially equivalent to other devices legally marketed in the United States, most notably Exactech's predicate Optetrak[®] AK Femoral and Optetrak[®] Posterior-Stabilized tibial insert.

Page 2 - Gary J. Miller, Ph.D.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4659. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

A handwritten signature in black ink, appearing to read "Celia M. Witten". The signature is written in a cursive style and is positioned to the right of the typed name.

Celia M. Witten, Ph.D., M.D.
Director

Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

**Exactech Optetrak® Total Knee System
Line Extension-Optetrak® HI-FLEX Knee Components
Special 510(k)**

Indications for Use
.....

510(k) Number: #K033883

Device Name: Optetrak® Total Knce System

- *HI-FLEX Asymmetric Posterior-Stabilized Cemented Femoral Components*
- *HI-FLEX Posterior-Stabilized Tibial Insert Components*

INDICATIONS

The OPTETRAK® Total Knee Systems are indicated for use in skeletally mature individuals undergoing primary surgery for total knee replacement due to osteoarthritis, osteonecrosis, rheumatoid arthritis and/or post-traumatic degenerative problems. They are also indicated for revision of failed previous reconstructions where sufficient bone stock and soft tissue integrity are present.

CONTRAINDICATIONS

The OPTETRAK® Total Knee Systems are contraindicated in patients with active infection, patients without sufficient bone stock to allow appropriate insertion and fixation of the prosthesis, patients without sufficient soft tissue integrity to provide adequate stability, and in patients with either mental or neuromuscular disorders that do not allow control of the knee joint, and in patients whose weight, age, or activity level might cause extreme loads and early failure of the system.

OPTETRAK® HI-FLEX Tibial Inserts must be used with Optetrak® HI-FLEX Femoral Components.

CAUTION: In the USA, for cemented use only.

Prescription Use X or Over the Counter Use
(Part 21 CFR 801 D)

[Handwritten Signature]

Please do not write below this line - use another page if needed.

(Division Sign-Off)

**Division of General, Restorative,
and Neurological Devices** Concurrence of CDRH, Office of Device Evaluation (ODE)

510(k) Number K033883