

JAN - 7 2004

510(k) Summary

K033937

Introduction According to the requirements of 21 CFR 807.92, the following information provides sufficient detail to understand the basis for a determination of substantial equivalence.

Submitter name, address, contact Roche Diagnostics Corporation
9115 Hague Road
Indianapolis, IN 46250
317-521-3723

Contact Person: Theresa M. Ambrose

Date Prepared: December 18, 2003

Device Name Proprietary name: Elecsys® PreciControl MultiAnalyte

Common name: PreciControl MultiAnalyte

Classification name: Multi-analyte Controls (assayed and unassayed)

Predicate device The Elecsys® PreciControl MultiAnalyte is substantially equivalent to the currently marketed Elecsys® PreciControl Cardiac (K032089).

Device Description The Elecsys® PreciControl MultiAnalyte is a lyophilized product consisting of synthetic human C-Peptide and recombinant human insulin in a buffered equine serum matrix. During manufacture, the analytes are spiked into the matrix at the desired concentration levels.

Intended use The Elecsys® PreciControl MultiAnalyte is used for quality control of the Elecsys C-Peptide and Elecsys Insulin immunoassays on the Elecsys immunoassay systems.

Continued on next page

510(k) Summary, Continued

Comparison to predicate device

The Elecsys® PreciControl MultiAnalyte is substantially equivalent to the currently marketed Elecsys® PreciControl Cardiac (K032089). The below tables compare Elecsys® PreciControl MultiAnalyte with the predicate device, Elecsys® PreciControl Cardiac (K032089).

Similarities

Characteristic	Elecsys® PreciControl MultiAnalyte	Predicate device Elecsys® PreciControl Cardiac
Intended Use	PreciControl MultiAnalyte is used for quality control of the Elecsys C-Peptide and Elecsys Insulin immunoassays on the Elecsys immunoassay systems.	PreciControl Cardiac is used for quality control of the Elecsys CK-MB, Digoxin, Myoglobin, and NT-proBNP immunoassays on the Elecsys immunoassay systems.
Levels	Two	Same
Format	Lyophilized	Same
Handling	Reconstitute with exactly 2.0 mL distilled water and allow to stand closed for 15 minutes.	Same

Continued on next page

510(k) Summary, Continued

Differences

Characteristic	Elecsys® PreciControl MultiAnalyte	Predicate device Elecsys® PreciControl Cardiac
Matrix	Equine serum with added C-Peptide and insulin	Human serum with added CK-MB, digoxin, myoglobin, and NT-proBNP
Stability	<u>Unopened:</u> • Store at 2-8°C until expiration date <u>Reconstituted:</u> • At -20°C : 1 months • On the analyzers : 3 hrs	<u>Unopened:</u> • Store at 2-8°C until expiration date <u>Reconstituted:</u> • At 2-8°C : 3 days • At -20°C : 3 months • On the analyzers : 3 hrs

Performance Characteristics

The Elecsys® PreciControl MultiAnalyte was evaluated for value assignment and stability.



JAN - 7 2004

Food and Drug Administration
2098 Gaither Road
Rockville MD 20850

Theresa M. Ambrose, Ph.D.
Regulatory Principal
Roche Diagnostics Corporation
9115 Hague Road
PO Box 50457
Indianapolis, IN 46250-0457

Re: k033937
Trade/Device Name: Elecsys[®] PreciControl MultiAnalyte
Regulation Number: 21 CFR 862.1660
Regulation Name: Quality control material (assayed and unassayed)
Regulatory Class: Class I
Product Code: JJY
Dated: December 18, 2003
Received: December 19, 2003

Dear Dr. Ambrose:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820).

Page 2 –

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (301) 594-3084. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>.

Sincerely yours,



Steven I. Gutman, M.D., M.B.A.
Director
Office of *In Vitro* Diagnostic Device
Evaluation and Safety
Center for Devices and
Radiological Health

Enclosure

Indications for Use Statement

510(k) Number (if known): N/A K 033937

Device Name: Elecsys® PreciControl MultiAnalyte

Indications For Use:

The Elecsys® PreciControl MultiAnalyte is used for quality control of the Elecsys C-Peptide and Elecsys Insulin immunoassays on the Elecsys immunoassay systems.

Prescription Use
(Per 21 CFR 801.109)

OR

Over-The-Counter Use

(Optional Format 1-2-96)

Carol Benson

Division Sign-Off

Office of In Vitro Diagnostic Device
Evaluation and Safety

510(k) K 033937