



February 5, 2024

Boston Scientific Corp.
Balaka Das
Senior Fellow, Regulatory Affairs
Boston Scientific Neuromodulation Corporation
25155 Rye Canyon Loop
Valencia, California 91355

Re: P030017/S363

Trade/Device Name: Spectra WaveWriter SCS System, WaveWriter Alpha SCS System, WaveWriter Alpha Prime SCS System

Product Code: LGW, QRB

Filed: May 30, 2023

Amended: November 7, 2023

Dear Balaka Das:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your premarket approval application (PMA) supplement for the Spectra WaveWriter SCS System, WaveWriter Alpha SCS System, WaveWriter Alpha Prime SCS System for expanding the indications to include patients with chronic intractable low back and leg pain without prior back surgery. The Boston Scientific Spinal Cord Stimulator (SCS) Systems* are indicated as an aid in the management of chronic intractable pain of the trunk and/or limbs including unilateral or bilateral pain associated with the following:

- Failed Back Surgery Syndrome
- Complex Regional Pain Syndrome (CRPS) Types I and II
- Intractable low back pain and leg pain with or without prior back surgery
- Radicular pain syndrome
- Radiculopathies resulting in pain secondary to failed back syndrome or herniated disc,
- Epidural fibrosis
- Degenerative disc disease (herniated disc pain refractory to conservative and surgical interventions)
- Arachnoiditis
- Multiple back surgeries.

The Boston Scientific Spectra WaveWriter™, WaveWriter Alpha™ and WaveWriter Alpha™ Prime SCS Systems are also indicated as an aid in the management of chronic intractable unilateral or bilateral low back and leg pain without prior back surgery.

*The Boston Scientific Spinal Cord Stimulator (SCS) Systems include the following:

- Precision™ System

- Precision Spectra™ System
- Precision Novi™ System
- Precision Montage™ MRI System
- Spectra WaveWriter™ System
- WaveWriter Alpha™ System
- WaveWriter Alpha™ Prime System

Based upon the information submitted, the PMA supplement is approved. You may begin commercial distribution of the device in accordance with the conditions of approval described below. Although this letter refers to your product as a device, please be aware that some approved products may instead be combination products. The Premarket Approval Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPMA/pma.cfm> identifies combination product submissions.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the act). FDA has determined that these restrictions on sale and distribution are necessary to provide reasonable assurance of the safety and effectiveness of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the act, in addition to all other applicable requirements, including those governing the manufacture, distribution, and marketing of devices.

Expiration dating for this device has been established and approved at 2 years in prior supplements for all SCS models. This is to advise you that the protocol you used to establish this expiration dating is considered an approved protocol for the purpose of extending the expiration dating as provided by 21 CFR 814.39(a)(7).

Continued approval of the PMA is contingent upon the submission of periodic reports, required under 21 CFR 814.84, at intervals of one year (unless otherwise specified) from the date of approval of the original PMA. This report, identified as "Annual Report" and bearing the applicable PMA reference number, should be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and must include the information required by 21 CFR 814.84.

In addition to the above, and in order to provide continued reasonable assurance of the safety and effectiveness of the PMA device, under 21 CFR 814.82(a)(9), the Annual Report must include, separately for each model number (if applicable), the number of devices sold and distributed during the reporting period, including those distributed to distributors. The distribution data will serve as a denominator and provide necessary context for FDA to ascertain the frequency and prevalence of adverse events, as FDA evaluates the continued safety and effectiveness of the device.

This is a reminder that as of September 24, 2014, class III devices are subject to certain provisions of the final Unique Device Identification (UDI) rule. These provisions include the requirement to provide a UDI on the device label and packages (21 CFR 801.20), format dates on the device label in accordance with 21 CFR 801.18, and submit data to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). Additionally, 21 CFR 814.84 (b)(4) requires PMA annual reports submitted after September 24, 2014, to identify each device identifier currently in use for the subject device, and the device identifiers for devices that have been discontinued since the previous periodic report. It is not necessary to identify any device identifier discontinued prior to December 23, 2013. Combination Products may also be subject to

UDI requirements (see 21 CFR 801.30). For more information on these requirements, please see the UDI website available at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-udi-system>.

Before making any change affecting the safety or effectiveness of the PMA device, you must submit a PMA supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.39. All PMA supplements and alternate submissions (30-day notice) must comply with the applicable requirements in 21 CFR 814.39. Additional information about changes that may require a PMA supplement are provided in the FDA guidance document entitled, "Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process" <https://www.fda.gov/media/81431/download>.

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production and process controls (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52 for devices or post-marketing safety reporting (21 CFR Part 4, Subpart B) for combination products, you are required to report adverse events for this device. Manufacturers of medical devices, including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or
2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems> and on combination product post-marketing safety reporting is available at <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>.

In accordance with the recall requirements specified in 21 CFR 806.10 for devices or the post-marketing safety reporting requirements (21 CFR Part 4, Subpart B) for combination products, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at <https://www.fda.gov/safety/recalls-market-withdrawals-safety-alerts/industry-guidance-recalls>.

CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading. CDRH will notify the public of its decision to approve your PMA by making available, among other information, a summary of the safety and effectiveness data upon which the approval is based. The information can be found at <https://www.fda.gov/medical-devices/device-approvals-denials-and-clearances/pma-approvals>. Written requests for this information can also be made to the Food and Drug Administration, Dockets Management Branch, (HFA-305), 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. The written request should include the PMA number or docket number. Within 30 days from the date that this information is placed on the Internet, any interested person may seek review of this decision by submitting a petition for review under section 515(g) of the act and requesting either a hearing or review by an independent advisory committee. FDA may, for good cause, extend this 30-day filing period.

Failure to comply with any post-approval requirement constitutes a ground for withdrawal of approval of a PMA. The introduction or delivery for introduction into interstate commerce of a device that is not in compliance with its conditions of approval is a violation of law.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this PMA submission with a copy of all final labeling. Final labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final labeling is identical to the labeling approved in draft form. If the final labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted, unless otherwise specified, to the address below and should reference the above PMA number to facilitate processing.

U.S. Food and Drug Administration
Center for Devices and Radiological Health
Document Control Center - WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

If you have any questions concerning this approval order, please contact Kaitlin Olsen at Kaitlin.Olsen@fda.hhs.gov.

Sincerely,

Vivek J. Pinto -S

Vivek Pinto, PhD

Director

DHT5B: Division of Neuromodulation
and Physical Medicine Devices

OHT5: Office of Neurological
and Physical Medicine Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health