

KD41813

Special 510(k) Premarket Notification
GE LOGIQ 7 Ultrasound BT04
July 2, 2004

JUL 23 2004

Attachment B:
Summary of Safety and Effectiveness
Prepared in accordance with 21 CFR Part 807.92(c).



GE Healthcare

General Electric Company
P.O. Box 414, Milwaukee, WI 53201

Section a):

1. Submitter: GE Medical Systems, Ultrasound and Primary Care Diagnostics, LLC
PO Box 414
Milwaukee, WI 53201
Contact Person: Allen Schuh,
Manager, Safety and Regulatory Engineering
Telephone: 414-647-4385; Fax: 414-647-4090
Date Prepared: July 2, 2004
2. Device Name: GE LOGIQ 7 Diagnostic Ultrasound System, BT04
Ultrasonic Pulsed Echo Imaging System, 21 CFR 892.1560, 90-IYO
Ultrasonic Pulsed Doppler Imaging System, 21 CFR 892.1550, 90-IYN
3. Marketed Device: GE LOGIQ 7 Diagnostic Ultrasound System K010329/K032182 (IYO/IYN)
A device currently in commercial distribution.
4. Device Description: The GE LOGIQ 7 is a full featured general purpose diagnostic ultrasound system. It consists of a mobile console approximately 60 cm wide, 100 cm deep and 140-160 cm (adjustable) high that provides digital acquisition, processing and display capability. The user interface includes a computer keyboard, specialized controls and a color video CRT and LCD touch panel. This modification will provide users with improved ergonomics, image enhancement and productivity.
5. Indications for Use: The device is intended for use by a qualified physician for ultrasound evaluation of Fetal; Abdominal; Pediatric; Small Organ (breast, testes, thyroid); Neonatal Cephalic; Adult Cephalic; Cardiac (adult and pediatric); Peripheral Vascular; Musculo-skeletal Conventional and Superficial; Urology (including prostate); Transrectal; Transvaginal; Transesophageal and Intraoperative (abdominal, thoracic, vascular and neurosurgical).
6. Comparison with Predicate Device: The GE LOGIQ 7 BT04 is of a comparable type and substantially equivalent to the current GE LOGIQ 7. It has the same technological characteristics, key safety and effectiveness features, physical design, construction, and materials, and has the same intended uses and basic operating modes as the predicate device.

Section b):

1. Non-clinical Tests: The device has been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness as well as thermal, electrical and mechanical safety, and has been found to conform with applicable medical device safety standards.
2. Clinical Tests: None required.
3. Conclusion: Intended uses and other key features are consistent with traditional clinical practice, FDA guidelines, and established methods of patient examination. The design and development process of the manufacturer conforms with 21 CFR 820, ISO 9001:2000 and ISO 13485 quality systems. The device conforms to applicable medical device safety standards and compliance is verified through independent evaluation with ongoing factory surveillance. Diagnostic ultrasound has accumulated a long history of safe and effective performance. Therefore, it is the opinion of GE Medical Systems that the GE LOGIQ 7 BT04 Diagnostic Ultrasound is substantially equivalent with respect to safety and effectiveness to devices currently cleared for market.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUL 23 2004

Mr. Allen Schuh
Manager, GE Ultrasound Safety
and Regulatory Engineering
GE Healthcare
General Electric Company
GE Medical Systems, Ultrasound and Primary Care Diagnostics, LLC
4855 West Electric Avenue
WEST MILWAUKEE WI 53219

Re: K041813

Trade Name: GE LOGIQ 7 Diagnostic Ultrasound System (with BT04 Modification)

Regulation Number: 21 CFR 892.1550

Regulation Name: Ultrasonic pulsed doppler imaging system

Regulation Number: 21 CFR 892.1560

Regulation Name: Ultrasonic pulsed echo imaging system

Regulation Number: 21 CFR 892.1570

Regulation Name: Diagnostic ultrasonic transducer

Regulatory Class: II

Product Code: 90 IYN, IYO, and ITX

Dated: July 2, 2004

Received: July 6, 2004

Dear Mr. Schuh:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the GE LOGIQ 7 Diagnostic Ultrasound System (with BT04 Modification), as described in your premarket notification:

Transducer Model Number

<u>3.5C</u>	<u>7L</u>	<u>7S</u>
<u>3.5Cs</u>	<u>10L</u>	<u>10S</u>
<u>3C</u>	<u>12L</u>	<u>6T</u>
<u>5C</u>	<u>M12L</u>	<u>P2D</u>
<u>M7C</u>	<u>3S</u>	<u>P6D</u>
<u>8C</u>	<u>M3S</u>	<u>i8L</u>
<u>E8C</u>	<u>4S</u>	<u>i12L</u>
<u>BE9C</u>	<u>5S</u>	

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This determination of substantial equivalence is granted on the condition that prior to shipping the first device, you submit a postclearance special report. This report should contain complete information, including acoustic output measurements based on production line devices, requested in Appendix G, (enclosed) of the Center's September 30, 1997 "Information for Manufacturers Seeking Marketing Clearance of Diagnostic Ultrasound Systems and Transducers." If the special report is incomplete or contains unacceptable values (e.g., acoustic output greater than approved levels), then the 510(k) clearance may not apply to the production units which as a result may be considered adulterated or misbranded.

The special report should reference the manufacturer's 510(k) number. It should be clearly and prominently marked "ADD-TO-FILE" and should be submitted in duplicate to:


Food and Drug Administration
Center for Devices and Radiological Health
Document Mail Center (HFZ-401)
9200 Corporate Boulevard
Rockville, Maryland 20850

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801, please contact the Office of Compliance at (301) 594-4591. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or at (301) 443-6597 or at its Internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

If you have any questions regarding the content of this letter, please contact Rodrigo C. Perez at (301) 594-1212.

Sincerely yours,


for Nancy C. Brogdon

Director, Division of Reproductive,
Abdominal and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure(s)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 Ultrasound System

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	Other
Ophthalmic											
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P	P	
Abdominal ^[1]	P	P	P	P	P	P	P	P	P	P	
Pediatric	P	P	P	P	P	P	P	P	P	P	
Small Organ ^[2]	P	P	P		P	P	P	P	P		
Neonatal Cephalic	P	P	P	P	P	P	P	P	P	P	
Adult Cephalic	P	P	P	P	P	P	P	P	P	P	
Cardiac ^[3]	P	P	P	P	P	P	P	P	P	P	
Peripheral Vascular	P	P	P	P	P	P	P	P	P	P	
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	
Other ^[4]	P	P	P	P	P	P	P	P	P	P	
<i>Exam Type, Means of Access</i>											
Transesophageal	P	P	P	P	P	P	P	P	P		
Transrectal	P	P	P		P	P	P	P	P		
Transvaginal	P	P	P		P	P	P	P	P		
Transurethral											
Intraoperative ^[5]	P	P	P	P	P	P	P	P	P	P	
Intraoperative Neurological	P	P	P		P	P	P	P	P	P	
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

- Notes: [1] Abdominal includes renal, GYN/Pelvic
 [2] Small organ includes breast, testes, thyroid.
 [3] Cardiac is Adult and Pediatric.
 [4] Other use includes Urology/Prostate
 [5] Intraoperative includes abdominal, thoracic (cardiac), and vascular (PV).
 [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. [Signature]

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 3.5C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	Other
<i>Anatomy/Region of Interest</i>											
Ophthalmic											
Fetal / Obstetrics	P	P	P		P	P	P	P	P	P	
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	
Pediatric											
Small Organ (specify)	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P	P	P	P	P	P	
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN;

[4] Other use includes Urology;

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David R. [Signature]

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number

KD41813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 3.5Cs Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	Other
Ophthalmic											
Fetal / Obstetrics	P	P	P		P	P	P	P	P	P	
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	
Pediatric											
Small Organ (specify)	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P	P	P	P	P	P	
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E


Notes: [1] Abdominal includes GYN;

[4] Other use includes Urology;

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 3C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	Other
Ophthalmic											
Fetal / Obstetrics	P	P	P		P	P	P	P	P	P	
Abdominal ^[1]	P	P	P		P	P	P	P	P	P	
Pediatric	P	P	P		P	P	P	P	P	P	
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P	P	P	P	P	P	
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E


Notes: [1] Abdominal includes GYN;

[4] Other use includes Urology;

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 5C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	Other
Ophthalmic											
Fetal / Obstetrics	P	P	P		P	P	P	P	P		
Abdominal ^[1]	P	P	P		P	P	P	P	P		
Pediatric	P	P	P		P	P	P	P	P		
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P	P	P	P	P		
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P		P	P	P	P	P		
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN/Pelvic, Renal and Aorta-iliac artery;

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Legrom

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number

2041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with M7C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	P	P	P		P	P	P	P	P	P	
Abdominal	P	P	P		P	P	P	P	P	P	
Pediatric	P	P	P		P	P	P	P	P	P	
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify)											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											


N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [2] Small organ includes breast, testes, thyroid.

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number 2041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 8C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/ Region of Interest</i>	Mode of Operation											
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse		
Ophthalmic												
Fetal / Obstetrics												
Abdominal ^[1]	P	P	P		P	P	P	P	P	P		
Pediatric	P	P	P		P	P	P	P	P	P		
Small Organ (specify)	P	P	P		P	P	P	P	P	P		
Neonatal Cephalic	P	P	P		P	P	P	P	P	P		
Adult Cephalic												
Cardiac												
Peripheral Vascular												
Musculo-skeletal Conventional												
Musculo-skeletal Superficial												
Other ⁽⁴⁾												
<i>Exam Type, Means of Access</i>												
Transesophageal												
Transrectal												
Transvaginal												
Transurethral												
Intraoperative (specify)												
Intraoperative Neurological												
Intravascular												
Laparoscopic												


N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN/Pelvic;

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K044813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with E8C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/ Region of Interest	Mode of Operation											
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse		
Ophthalmic												
Fetal / Obstetrics	P	P	P		P	P	P	P	P			
Abdominal ^[1]	P	P	P		P	P	P	P	P			
Pediatric												
Small Organ (specify)												
Neonatal Cephalic												
Adult Cephalic												
Cardiac												
Peripheral Vascular												
Musculo-skeletal Conventional												
Musculo-skeletal Superficial												
Other ^[4]	P	P	P		P	P	P	P	P			
Exam Type, Means of Access												
Transesophageal												
Transrectal	P	P	P		P	P	P	P	P			
Transvaginal	P	P	P		P	P	P	P	P			
Transurethral												
Intraoperative (specify)												
Intraoperative Neurological												
Intravascular												
Laparoscopic												

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN/Pelvic;

[4] Other use includes Urology/Prostate;

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Seymour

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K04813

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with BE9C Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/ Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics											
Abdominal ^[1]											
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	N	N	N		N	N	N	N	N		
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal	N	N	N		N	N	N	N	N		
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E


Notes: [1] Abdominal includes GYN/Pelvic;

[4] Other use includes Urology/Prostate;

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K041813

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 7L Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	P	P	P		P	P	P	P	P		
Abdominal	P	P	P		P	P	P	P	P		
Pediatric											
Small Organ ^[2]	P	P	P		P	P	P	P	P		
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P	P	P	P	P		
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P		
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P		
Other ^[4]											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [2] Small organ includes breast, testes, thyroid.

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Lynn

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number

2041813

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 10L Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation									
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse
Ophthalmic										
Fetal / Obstetrics	P	P	P		P	P	P	P	P	P
Abdominal	P	P	P		P	P	P	P	P	P
Pediatric	P	P	P		P	P	P	P	P	P
Small Organ ^[2]	P	P	P		P	P	P	P	P	P
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Peripheral Vascular	P	P	P		P	P	P	P	P	P
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P
Other (specify)										
Exam Type, Means of Access										
Transesophageal										
Transrectal										
Transvaginal										
Transurethral										
Intraoperative ^[5]	P	P	P		P	P	P	P	P	P
Intraoperative Neurological	P	P	P		P	P	P	P	P	P
Intravascular										
Laparoscopic										

N = new indication; P = previously cleared by FDA; E = added under Appendix E


Notes: [2] Small organ includes breast, testes, thyroid.

[5] Intraoperative includes abdominal, thoracic, and vascular. Neurosurgical added via K970901.

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number 14041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 12L Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics											
Abdominal											
Pediatric	P	P	P		P	P	P	P	P		
Small Organ ^[2]	P	P	P		P	P	P	P	P		
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P	P	P	P	P		
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P		
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P		
Other (specify)											
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative ^[5] (specify)	P	P	P		P	P	P	P	P		
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [2] Small organ includes breast, testes, thyroid.

[5] Intraoperative includes abdominal, thoracic, and vascular.

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David G. Seaman

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number

KD41813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with M12L Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics											
Abdominal											
Pediatric	P	P	P		P	P	P	P	P	P	
Small Organ ^[2]	P	P	P		P	P	P	P	P	P	
Neonatal Cephalic											
Adult Cephalic											
Cardiac											
Peripheral Vascular	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Conventional	P	P	P		P	P	P	P	P	P	
Musculo-skeletal Superficial	P	P	P		P	P	P	P	P	P	
Other (specify)											
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative ^[5] (specify)	P	P	P		P	P	P	P	P	P	
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [2] Small organ includes breast, testes, thyroid.

[5] Intraoperative includes abdominal, thoracic, and vascular.

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Seymour

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K041813

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 3S Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation											
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse		
Ophthalmic												
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P	P	P	
Abdominal ^[1]	P	P	P	P	P	P	P	P	P	P	P	
Pediatric	P	P	P	P	P	P	P	P	P	P	P	
Small Organ (specify)												
Neonatal Cephalic	P	P	P	P	P	P	P	P	P	P	P	
Adult Cephalic	P	P	P	P	P	P	P	P	P	P	P	
Cardiac ^[3]	P	P	P	P	P	P	P	P	P	P	P	
Peripheral Vascular												
Musculo-skeletal Conventional												
Musculo-skeletal Superficial												
Other ^[4]	P	P	P	P	P	P	P	P	P	P	P	
<i>Exam Type, Means of Access</i>												
Transesophageal												
Transrectal												
Transvaginal												
Transurethral												
Intraoperative (specify)												
Intraoperative Neurological												
Intravascular												
Laparoscopic												

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN;

[3] Cardiac is Adult and Pediatric;

[4] Other use includes Urology;

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Seymour

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K024181B

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with M3S Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P		
Abdominal ^[1]	P	P	P	P	P	P	P	P	P		
Pediatric	P	P	P	P	P	P	P	P	P		
Small Organ (specify)											
Neonatal Cephalic	P	P	P	P	P	P	P	P	P		
Adult Cephalic	P	P	P	P	P	P	P	P	P		
Cardiac ^[3]	P	P	P	P	P	P	P	P	P		
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify)	P	P	P	P	P	P	P	P	P		
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes Renal and GYN;

[3] Cardiac is Adult and Pediatric.

[4] Other use includes Urology;

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Johnson

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K041813

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 4S Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/Region of Interest	Mode of Operation									
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes*	Harmonic Imaging	Coded Pulse
Ophthalmic										
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P	P
Abdominal ^[1]	P	P	P	P	P	P	P	P	P	P
Pediatric	P	P	P	P	P	P	P	P	P	P
Small Organ (specify)										
Neonatal Cephalic	P	P	P	P	P	P	P	P	P	P
Adult Cephalic	P	P	P	P	P	P	P	P	P	P
Cardiac ^[3]	P	P	P	P	P	P	P	P	P	P
Peripheral Vascular										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)	P	P	P	P	P	P	P	P	P	P
Exam Type, Means of Access										
Transesophageal										
Transrectal										
Transvaginal										
Transurethral										
Intraoperative (specify)										
Intraoperative Neurological										
Intravascular										
Laparoscopic										

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes Renal and GYN;

[3] Cardiac is Adult and Pediatric.

[4] Other use includes Urology;

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Seymour

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number KD41813

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 5S Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:


Clinical Application Anatomy/Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P		
Abdominal ^[1]	P	P	P	P	P	P	P	P	P		
Pediatric	P	P	P	P	P	P	P	P	P		
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic	P	P	P	P	P	P	P	P	P		
Cardiac ^[3]	P	P	P	P	P	P	P	P	P		
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other ^[4]	P	P	P	P	P	P	P	P	P		
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

- Notes: [1] Abdominal includes GYN;
 [3] Cardiac is Adult and Pediatric;
 [4] Other use includes Urology;
 [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K041813

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 7S Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics	P	P	P	P	P	P	P	P	P		
Abdominal ^[1]	P	P	P	P	P	P	P	P	P		
Pediatric	P	P	P	P	P	P	P	P	P		
Small Organ (specify)											
Neonatal Cephalic	P	P	P	P	P	P	P	P	P		
Adult Cephalic	P	P	P	P	P	P	P	P	P		
Cardiac ^[3]	P	P	P	P	P	P	P	P	P		
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify)											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal includes GYN;

[3] Cardiac is Adult and Pediatric.

[4] Other use includes Urology and GYN.

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Leggett

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 10S Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/Region of Interest</i>	Mode of Operation											
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse		
Ophthalmic												
Fetal / Obstetrics												
Abdominal	P	P	P	P	P	P	P	P	P	P		
Pediatric	P	P	P	P	P	P	P	P	P	P		
Small Organ (specify)												
Neonatal Cephalic	P	P	P	P	P	P	P	P	P	P		
Adult Cephalic	P	P	P	P	P	P	P	P	P	P		
Cardiac ^[3]	P	P	P	P	P	P	P	P	P	P		
Peripheral Vascular												
Musculo-skeletal Conventional												
Musculo-skeletal Superficial												
Other (specify)												
<i>Exam Type, Means of Access</i>												
Transesophageal												
Transrectal												
Transvaginal												
Transurethral												
Intraoperative (specify)												
Intraoperative Neurological												
Intravascular												
Laparoscopic												

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [3] Cardiac is Adult and Pediatric.

[*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Seymour

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number

K041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with 6T Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:


Clinical Application <i>Anatomy/ Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics											
Abdominal											
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac	P	P	P	P	P	P	P	P	P		
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify)											
<i>Exam Type, Means of Access</i>											
Transesophageal	P	P	P	P	P	P	P	P	P		
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [*] Combined modes are B/M, B/Color M, B/PWD or CWD, B/Color/PWD or CWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 510(k) Number K041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with P2D Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/ Region of Interest</i>	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics											
Abdominal											
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac ^[3]				P							
Peripheral Vascular				P							
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify)											
<i>Exam Type, Means of Access</i>											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative (specify)											
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [3] Cardiac is Adult and Pediatric.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David A. Segerson

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 ID Number

12041813

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with P6D Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:


Clinical Application <i>Anatomy/ Region of Interest</i>	Mode of Operation											
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse		
Ophthalmic												
Fetal / Obstetrics												
Abdominal												
Pediatric												
Small Organ (specify)												
Neonatal Cephalic												
Adult Cephalic												
Cardiac ^[3]					P							
Peripheral Vascular					P							
Musculo-skeletal Conventional												
Musculo-skeletal Superficial												
Other (specify)												
<i>Exam Type, Means of Access</i>												
Transesophageal												
Transrectal												
Transvaginal												
Transurethral												
Intraoperative (specify)												
Intraoperative Neurological												
Intravascular												
Laparoscopic												

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [3] Cardiac is Adult and Pediatric.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 File Number K041813

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with i8L Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application <i>Anatomy/ Region of Interest</i>	Mode of Operation											
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse		
Ophthalmic												
Fetal / Obstetrics												
Abdominal ^[1]	P	P	P		P	P	P	P	P			
Pediatric												
Small Organ (specify)												
Neonatal Cephalic												
Adult Cephalic												
Cardiac ^[3]	P	P	P		P	P	P	P	P			
Peripheral Vascular												
Musculo-skeletal Conventional												
Musculo-skeletal Superficial												
Other (specify)												
<i>Exam Type, Means of Access</i>												
Transesophageal												
Transrectal												
Transvaginal												
Transurethral												
Intraoperative ^[5]	P	P	P		P	P	P	P	P			
Intraoperative Neurological												
Intravascular												
Laparoscopic												

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal is via Intraoperative;

[3] Cardiac is Adult and Pediatric via Intraoperative;

[5] Intraoperative includes abdominal, thoracic, and vascular.

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David G. Leggett

(Division Sign-Off)
 Division of Reproductive, Abdominal,
 and Radiological Devices
 File Number

K041813

Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form
GE LOGIQ 7 with i12L Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application Anatomy/ Region of Interest	Mode of Operation										
	B	M	PW Doppler	CW Doppler	Color Doppler	Color M Doppler	Power Doppler	Combined Modes	Harmonic Imaging	Coded Pulse	
Ophthalmic											
Fetal / Obstetrics											
Abdominal ^[1]	P	P	P		P	P	P	P			
Pediatric											
Small Organ (specify)											
Neonatal Cephalic											
Adult Cephalic											
Cardiac ^[3]	P	P	P		P	P	P	P			
Peripheral Vascular											
Musculo-skeletal Conventional											
Musculo-skeletal Superficial											
Other (specify)											
Exam Type, Means of Access											
Transesophageal											
Transrectal											
Transvaginal											
Transurethral											
Intraoperative ^[5]	P	P	P		P	P	P	P			
Intraoperative Neurological											
Intravascular											
Laparoscopic											

N = new indication; P = previously cleared by FDA; E = added under Appendix E

Notes: [1] Abdominal is via Intraoperative;

[3] Cardiac is Adult and Pediatric via Intraoperative;

[5] Intraoperative includes abdominal, thoracic, and vascular.

[*] Combined modes are B/M, B/Color M, B/PWD, B/Color/PWD, B/Power/PWD.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David E. Larson

(Division Sign-Off)
 Division of Reproductive, Abdominal
 and Radiological Devices
 FDA Number:

 K041813

Prescription Use (Per 21 CFR 801.109)