

K042515

Smith & Nephew, Inc.
Summary of Safety and Effectiveness
High Performance Knee

MAR 14 2005

Contact Person and Address

Janet Johnson Akil
Director, Regulatory Affairs
Smith & Nephew, Inc.
Orthopaedic Division
1450 Brooks Road
Memphis, TN 38116
(901) 399-5153

Date of Summary: September 15, 2004

Name of Device: High Performance Knee

Common Name: Knee prosthesis

Device Description

The High Performance Knee consists of femoral, tibial insert and patellar components to be used with existing Genesis II tibial trays. The High Performance Knee is a total knee system which provides the ability for greater flexion to those patients who have the anatomical capability to allow a greater flexion range.

Device Classification

21 CFR 888.3560 Knee joint patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis – Class II

Mechanical and Clinical Data

A review of the mechanical test data indicated that the High Performance Knee is equivalent to devices currently used clinically and is capable of withstanding expected *in vivo* loading without failure.

Indications for Use

Total knee components are indicated for rheumatoid arthritis; post-traumatic arthritis, osteoarthritis, or degenerative arthritis in older patients whose age, weight, and activity level are compatible with an adequate long-term result; failed osteotomies, unicompartmental replacement, or total knee replacement. Posterior stabilized knee systems are designed for use in patients in primary and revision surgery, where the anterior and posterior cruciate ligaments are absent or incompetent and the collateral ligaments remain intact. Smith & Nephew, Inc. High Performance Knee components are indicated for use only with cement and are single use devices.

Substantial Equivalence Information

The substantial equivalence of the High Performance Knee is based on its similarities in indications for use, design features, operational principles, and material composition to the following predicate devices – Smith & Nephew Genesis II Knee System and Genesis II & Profix Zirconium Femoral Knee Components (K951987 and K962557, respectively) and the Zimmer NexGen Complete Knee Solution LPS-Flex Fixed Bearing Femoral and Articular Surface Components (K991581).



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

MAR 1 4 2005

Ms. Janet Johnson Akil
Director, Regulatory Affairs
Smith & Nephew, Inc.
Orthopaedic Division
1450 E. Brooks Road
Memphis, Tennessee 38116

Re: K042515

Trade/Device Name: High Performance Knee

Regulation Numbers: 21 CFR 888.3560

Regulation Names: Knee joint, patellofemorotibial, polymer/metal/polymer semi-
constrained cemented prosthesis

Regulatory Class: II

Product Codes: JWH

Dated: March 1, 2005

Received: March 2, 2005

Dear Ms. Akil:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21

CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/edrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Miriam Provost", with a long horizontal flourish extending to the right.

Miriam Provost, Ph.D.
Acting Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K042515

Device Name: High Performance Knee


Indications for Use:

Total knee components are indicated for rheumatoid arthritis; post-traumatic arthritis, osteoarthritis, or degenerative arthritis in older patients whose age, weight, and activity level are compatible with an adequate long-term result; failed osteotomies, unicompartmental replacement, or total knee replacement. Posterior stabilized knee systems are designed for use in patients in primary and revision surgery, where the anterior and posterior cruciate ligaments are absent or incompetent and the collateral ligaments remain intact. Smith & Nephew, Inc. High Performance Knee components are indicated for use only with cement and are single use devices.

Prescription Use X AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

for 
(Division Sign-Off)

**Division of General, Restorative,
and Neurological Devices**

Page 1 of _____

510(k) Number K042515