

APR 8 2005

K050244

**II. 510(k) SUMMARY**

Submitted By: **NRS Global Partners Sdn Bhd**  
No. 30, Jalan 3/146, Bandar Tasik Selatan  
5700 Kuala Lumpur, Malaysia

Contact Persons: **Eli J. Carter, Consultant**  
1219 Little Creek Road  
Durham, NC 27713

**Mr. Anthony Tan Jian Loong**  
Chief Operating Officer  
NRS Global Partners Shn Bhd

Date Prepared: January 31, 2005

Proprietary Name: None

Common Name: Male Latex Condom

Classification Name: Male Latex Condom

Predicate Device: Innolates Male Latex Condom; **K002393**

Description of Device: This condom is made of a natural latex sheath, which completely covers the erect penis with a closely fitted membrane. This condom is straight sided, contoured or flared in shape, either textured or non-textured, with a reservoir tip; nominal length 180-mm, nominal width 53-mm, and nominal thickness 0.07mm. It is lubricated with silicone or Astroglide™, and USP grade cornstarch is used as a dressing material. This condom is designed to conform to established national and international voluntary standards including ASTM D3492 and ISO 4074.

**This condom is identical to the predicate device with respect to design, physical properties, and formulation.**

Intended Use of the Device:

This latex condom has the same intended use as the predicate condom. The condom is used for contraception and for prophylactic purposes to help prevent pregnancy and the transmission of sexually transmitted diseases, including HIV.

Technological Characteristics:

This condom has the same technological characteristics as the predicate condom identified above. The design is in conformance with ASTM Latex Condom Standard D3492 and that the condom is made of natural rubber latex.



APR 8 2005

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

NRS Global Partners Sdn Bhd  
% Mr. Eli Carter  
Consultant  
1219 Little Creek Rd.  
DURHAM NC 27713

Re: K050244  
Trade/Device Name: Male Natural Latex Condom  
Regulation Number: 21 CFR 884.5300  
Regulation Name: Condom  
Regulatory Class: II  
Product Code: HIS  
Dated: January 31, 2005  
Received: February 4, 2005

Dear Mr. Carter:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at one of the following numbers, based on the regulation number at the top of this letter:

21 CFR 876.xxxx	(Gastroenterology/Renal/Urology)	240-276-0115
21 CFR 884.xxxx	(Obstetrics/Gynecology)	240-276-0115
21 CFR 892.xxxx	(Radiology)	240-276-0120
Other		240-276-0100

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Nancy C. Brogdon  
Director, Division of Reproductive,  
Abdominal, and Radiological Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure

K050244

**INDICATIONS FOR USE**

510(k) Number: ~~Not Known~~ K050244  
Device Name: Male Natural Rubber Latex Condom  
Indications for Use: The NRS Global Partners male latex condom is used for contraceptive and for prophylactic purposes (to help prevent pregnancy and the transmission of sexually transmitted diseases)

Prescription Use \_\_\_\_\_  
(Per 21 CFR 801 Subpart D)

OR

Over-the-Counter Use  \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

Nancy C. Brogdon  
(Division Sign-Off)  
Division of Reproductive, Abdominal,  
And Radiological Devices

510(k) Number K050244