



K050646

1/1

APR 26 2005

3.0 Summary of Safety and Effectiveness Information

Page 1 of 1

SPONSOR: Synthes (USA)  
1302 Wrights Lane East  
West Chester, PA ~~19380~~ 19380  
(610) 719-5000

DEVICE NAME: Synthes (USA) 3.5 / 4.5mm LCP® Medial Proximal Tibia Plates

CLASSIFICATION: Class II, §888.3030 – Plate, Fixation, Bone

PREDICATE DEVICE: Synthes (USA) 3.5 / 4.5mm LCP® Medial Proximal Tibia Plates

DEVICE DESCRIPTION: The Synthes (USA) 3.5 / 4.5 mm LCP® Medial Proximal Tibia Plates are contoured to match the anatomy of the medial proximal tibia with a limited contact low profile design. The plates are designed for either the right or left medial proximal tibia in a variety of shaft lengths. These plates will be available in both 3.5 mm and 4.5 mm versions. The plate head exhibits 2.0 mm holes for preliminary fixation with k-wires, or meniscal repair with sutures.

INTENDED USE: The Synthes (USA) 3.5 / 4.5mm LCP® Medial Proximal Tibia Plates are intended to buttress metaphyseal fractures of the medial tibia plateau, split-type fractures of the medial tibia plateau, medial split fractures with associated depressions and split or depression fractures of the medial tibia plateau. The plates may also be used for fixation of the proximal quarter (lateral and medial) of the tibia as well as segmental fractures of the proximal tibia. The 4.5mm version may also be used for fixation of nonunions and malunions of the medial proximal tibia and tibia shaft, as well as opening and closing wedge tibial osteotomies.

SUBSTANTIAL EQUIVALENCE Comparative information presented supports substantial equivalence.

MATERIAL: Titanium and Stainless Steel



APR 26 2005

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Ms. Kathy Anderson  
Regulatory Affairs Manager  
Synthes (USA)  
1302 Wrights Lane East  
West Chester, Pennsylvania 19380

Re: K050646

Trade/Device Name: Synthes (USA) 3.5/ 4.5mm LCP® Medial Proximal Tibia Plates  
Regulation Number: 21 CFR 888.3030  
Regulation Name: Single/ multiple component metallic bone fixation appliances and accessories

Regulatory Class: II  
Product Code: HRS  
Dated: March 9, 2005  
Received: March 14, 2005

Dear Ms. Anderson:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Ms. Kathy Anderson

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Miriam C. Provost", is written over the typed name. The signature is fluid and cursive.

Miriam C. Provost, Ph.D.  
Acting Director  
Division of General, Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure



2.0 Indications for Use Statement

Page 1 of 1

510(k) Number (if known): \_\_\_\_\_

Device Name: Synthes (USA) 3.5 / 4.5mm LCP® Medial Proximal Tibia Plates

Indications/Contraindications:

The Synthes (USA) 3.5 / 4.5mm LCP® Medial Proximal Tibia Plates are intended to buttress metaphyseal fractures of the medial tibia plateau, split-type fractures of the medial tibia plateau, medial split fractures with associated depressions and split or depression fractures of the medial tibia plateau. The plates may also be used for fixation of the proximal quarter (lateral and medial) of the tibia as well as segmental fractures of the proximal tibia. The 4.5mm version may also be used for fixation of nonunions and malunions of the medial proximal tibia and tibia shaft, as well as opening and closing wedge tibial osteotomies.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use X OR Over-The-Counter Use \_\_\_\_\_  
(Per 21 CFR 801.109)

(Division Sign-Off)

Division of General, Restorative,  
and Neurological Devices

510(k) Number K050646