

JUN 21 2005

K050860

Date: April 1, 2005  
Subject: 510(k) Summary of Safety and Effectiveness Information  
for the GE Datex-Ohmeda Centiva/5 Ventilator

Proprietary: GE Datex-Ohmeda Centiva/5 Ventilator

Common: Ventilator, Continuous

Classification: Anesthesiology, 73 CBK, 21 CFR 868.5895

The 510(k) summary of safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 1992.

The GE Datex-Ohmeda Centiva/5 is substantially equivalent to the following currently marketed device:

Drager Evita 4- Class II - 21CFR868.5895, which has been the subject of a cleared 510(k) with FDA log number K992608.

Drager Evita 2- Class II - 21CFR868.5895, which has been the subject of a cleared 510(k) with FDA log number K970165.

GE Datex-Ohmeda Engstrom Carestation –Class II - 21CFR868.5895, which has been the subject of several cleared 510(k)s, most recently with FDA log number K041775.

The Centiva/5 is a flexible, intuitive and simple to operate critical care ventilator. It offers a configurable and compact unit, which places significant ventilation power to the hands of the clinician yet maintains simple operation. The Centiva/5 is equipped with a RS-232 serial data interface to enable communication with Clinical Information Management Systems.

The GE Datex-Ohmeda Centiva/5 is designed to provide mechanical ventilation for adults and pediatrics weighing 5kg and above having degrees of pulmonary impairment varying from minor to severe. The modes of ventilation are available include:

- Volume Controlled (VCV)
- Synchronized Intermittent Mandatory Ventilation, Volume Controlled (SIMV-VC)
- Bi-level Airway Pressure Ventilation
- Constant Positive Airway Pressure/Pressure Support Ventilation (CPAP/PSV)
- Apnea backup (active in CPAP/PSV)

The GE Datex-Ohmeda Centiva/5 is a microprocessor based, electronically controlled, pneumatically driven ventilator that includes integrated FiO<sub>2</sub>, airway pressure, flow and volume monitoring. The Centiva/5 is equipped with a pneumatic outlet capable of driving standard pneumatic drug nebulizers.

The system is designed for facility use, including within-facility transport, and should only be used under the orders of a clinician.

The ventilator consists of two main components: a display or control panel and a ventilator unit. The display allows the user to interface with the system and control settings. The ventilator unit controls electrical power, nebulization, and pneumatic gas flow to and from the patient.

Optional accessories include a trolley/cart, support arm, optional humidifier and an external battery unit.

The GE Datex-Ohmeda Centiva/5 Ventilator was designed to comply with the applicable portions of the following voluntary standards;

1. UL 2601 – General requirements for Medical Electrical Equipment
2. ASTM F1100 – Particular Requirements for Critical Care Ventilators
3. EN/IEC 60601-1: General requirements for Medical Electrical Equipment
4. EN/IEC 60601-1-2: 1998 - Medical Electrical Equipment - Electromagnetic Compatibility
5. EN 475 – Electrically Generated Alarm Signals
6. CGA V-1 ad ISO 5145 Medical Gas Cylinders – Threaded Cylinders
7. EN 980 Graphical Symbols
8. EN/IEC 60601-2-12, Medical Electrical Equipment – Critical Care Ventilators

The GE Datex-Ohmeda Centiva/5 Ventilator and the currently marketed device are substantially equivalent in design concepts, technologies and materials. The GE Datex-Ohmeda Centiva/5 Ventilator has been validated through rigorous testing that, in part, supports the compliance of GE Datex-Ohmeda Centiva/5 Ventilator to the standards listed above.

Contact: Dan Kosednar, RAC  
Manager, Regulatory Planning and Submissions



JUN 21 2005

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Mr. Dan Kosednar  
Manager, Regulatory Planning and Submissions  
GE Healthcare  
Datex-Ohmeda, Incorporated  
P.O. Box 7550  
Madison, Wisconsin 53707

Re: K050860  
Trade Name: GE Datex-Ohmeda Centiva/5 Ventilator  
Regulation Number: 21 CFR 868.5895  
Regulation Name: Continuous Ventilator  
Regulatory Class: Class II  
Product Code: CBK  
Dated: June 6, 2005  
Received: June 8, 2005

Dear Mr. Kosednar:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must

Page 2 – Mr. Dan Kosednar

comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known): K050860

Device Name: GE Datex-Ohmeda Centiva/5 Critical Care Ventilator

Indications For Use:

The GE Datex-Ohmeda Centiva/5 Critical Care Ventilator is designed as a critical care ventilator for adult to pediatric patients. It provides volume control, pressure limited volume, bi-level and pressure support ventilation (PSV), Continuous Positive Airway Pressure (CPAP) and PSV with apnea back-up ventilation modes. It incorporates integrated airway pressure, patient flow, tidal volume, minute volume, peak pressure, PEEP, rate and FIO2 monitoring.

The system is designed for facility use, including within-facility transport, and should only be used under the orders of a clinician.

Prescription Use XXX  
(Part 21 CFR 801 Subpart D)


AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page 1 of 1



(Division Sign-Off)  
Division of Anesthesiology, General Hospital,  
Infection Control, Dental Devices

510(k) Number: K050860