

Section 2 - Summary of Safety and Effectiveness

(1) Company Information

MicroVention, Inc.
75 Columbia
Aliso Viejo, CA 92656
Telephone: (949) 461-3314
Fax: (949) 461-3329
www.microvention.com

(2) Contact Information

Vincent Cutarelli
Telephone: (949) 461-3314 ext. 105
Fax: (949) 461-3329
E-mail: vinc@microvention.com

(3) Device Name

Trade/Proprietary Name:	MicroPlex™ Coil System (MCS) HydroCoil® Embolic System (HES)
Classification Name:	Device, Artificial Embolization Device, Embolization, Arterial
Common/Usual Name:	Endovascular Embolization Coil

(4) Device Description

The MicroPlex® Coil System (MCS) and HydroCoil® Embolic System (HES) consist of an implantable coil attached to a delivery system called a Delivery Pusher. The Delivery Pusher is powered by a hand-held, battery-powered Detachment Controller designed specifically for the MCS and HES. The Detachment Controller is provided separately.

The coil is delivered to treatment site on the Delivery Pusher through standard neuro-interventional micro-catheters. A removable introducer sheath on the outside of the Delivery Pusher assists in the placement of the MCS and HES into the micro-catheter. During deployment, the proximal end of the Delivery Pusher is connected to the Detachment Controller. When the controller is turned on, the coil detaches.

(5) **Indications for Use**

The MicroPlex[®] Coil System (MCS) and HydroCoil[®] Embolic System (HES) are intended for the endovascular embolization of intracranial aneurysms and other neurovascular abnormalities such as arteriovenous malformations and arteriovenous fistulae.

The MCS and HES are also intended for vascular occlusion of blood vessels within the neurovascular system to permanently obstruct blood flow to an aneurysm or other vascular malformation and for arterial and venous embolizations in the peripheral vasculature.

(6) **Name of Predicate or Legally Marketed Device**

The MicroVention MicroPlex[®] Coil System (MCS) and HydroCoil[®] Embolic System (HES) with a modified detachment system are substantially equivalent to the MicroPlex[®] Coil System (MCS) and HydroCoil[®] Embolic System (HES) with expanded indications for use that were determined to be substantially equivalent on October 22, 2003 (reference K032590), the HydroCoil[®] Embolic System (HES) with HES-HC-S (10) Coils that was determined to be substantially equivalent on December 30, 2003 (reference K033836), the HydroCoil[®] Embolic System (HES) that was determined to be substantially equivalent on July 7, 2004 (reference K041551) and the Micrus MicroCoil System that was determined to be substantially equivalent on August 1, 2003 (reference K031578).

(7) **Technological Characteristics and Substantial Equivalence**

The MicroPlex[®] Coil System (MCS) and HydroCoil[®] Embolic System (HES) with a modified detachment system are substantially equivalent in design, materials, operating principle, method of application, indications for use, packaging and sterilization to the predicate devices.

(8) **Performance Data Summary**

Performance testing has demonstrated that the MicroPlex[®] Coil System (MCS) and HydroCoil[®] Embolic System (HES) with a modified detachment system are equivalent in performance to the predicate devices.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUN 28 2005

Mr. Vincent Cutarelli
Vice President, Regulatory Affairs and Quality Assurance
MicroVention Incorporated
75 Columbia, Suite A
Aliso Viejo, California 92656

Re: K050954

Trade/Device Name: *MicroPlex*® Coil System (MCS) and *HydroCoil*® Embolic System (HES)

Regulation Number: 21 CFR 882.5950

Regulation Name: Artificial embolization device

Regulatory Class: II

Product Code: HCG

Dated: April 14, 2005

Received: April 15, 2005

Dear Mr. Cutarelli:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Mr. Vincent Cutarelli

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115 . Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,


for Miriam C. Provost, Ph.D.
Acting Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K050954

Device Name: MicroPlex[®] Coil System (MCS) and HydroCoil[®] Embolic System (HES)

Indications for Use: The MicroPlex[®] Coil System (MCS) and HydroCoil[®] Embolic System (HES) are intended for the endovascular embolization of intracranial aneurysms and other neurovascular abnormalities such as arteriovenous malformations and arteriovenous fistulae.

The MCS and HES are also intended for vascular occlusion of blood vessels within the neurovascular system to permanently obstruct blood flow to an aneurysm or other vascular malformation and for arterial and venous embolizations in the peripheral vasculature.


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of General, Restorative
and Neurological Devices

510(k) Number K050954