



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

JUL 25 2005

Sarstedt, Incorporated  
c/o Mr. Peter Rumswinkel  
VP/General Manager  
1025 St. James Church Rd.  
Newton, North Carolina 28658

Re: K051019

Trade/Device Name: Safety Needle

Regulation Number: 21 CFR 862.1675

Regulation Name: Blood Collection Needle, Accessory To Tubes, Vials, System

Regulatory Class: II

Product Code: JKA

Dated: July 06, 2005

Received: July 7, 2005

Dear Mr. Rumswinkel:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

# SARSTEDT



K051019

Sarstedt, Inc. • P.O. Box 468 • Newton, NC 28658-0468

Instruments  
and Disposables  
for Medicine  
and Science

## Indications for Use

510(k) Number: K051019

Device Name: Safety Needle

### Indications For Use:

The Safety-Needle is a single lumen needle for blood collection via venipuncture in conjunction with the S-Monovette<sup>®</sup> system. The Safety-Needle needle protector, when activated after removal of the needle from the patient, covers the used needle until it can be properly disposed; therefore, aiding in the prevention of needle stick injuries.

The Safety-Needle is individually packaged, sterile, single-use and non-pyrogenic.

The Safety-Needle is intended to be used in all hospitals and/or practices where S-Monovettes are used.

Prescription Use    
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use    
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)  
Division of Anesthesiology, General Hospital,  
Infection Control, Dental Devices

510(k) Number: K051019

Page 1 or 1