DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
2098 Gaither Road
Rockville MD 20850

APR - 6 2006

Michael J. Vincent
V.P. Operations
Immunalysis Corp.
829 Towne Center Drive
Pomona, CA 91767

Re: k051579
Trade/Device Name: Immunalysis Amphetamine ELISA for Oral Fluids
Regulation Number: 21 CFR 862.3100
Regulation Name: Amphetamine test system
Regulatory Class: Class II
Product Code: DKZ
Dated: August 10, 2005
Received: August 22, 2005

Dear Mr. Vincent:

This letter corrects our substantially equivalent letter of October 21, 2005.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820).

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (240) 276-0484. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address http://www.fda.gov/cdrh/industry/support/index.html.

Sincerely yours,

Alberto Gutierrez, Ph.D.
Director
Division of Chemistry and Toxicology
Office of In Vitro Diagnostic Device
Evaluation and Safety
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K051579

Device Name: Immunalysis Amphetamine ELISA for Oral Fluids

Indications For Use:

The Immunalysis Amphetamine ELISA test system utilizes an Enzyme Linked Immunoassay (ELISA) for the qualitative detection of Amphetamine in oral fluid samples collected with the Quantisal™ oral fluid collection device using a cutoff of 50 ng/mL of d-Amphetamine. This in-vitro diagnostic device is intended for clinical laboratory use only.

The Immunalysis Amphetamine ELISA Kit for Oral Fluids provides only a preliminary analytical test result. A more specific alternate chemical method must be used in order to obtain a confirmed analytical result. Gas chromatography/ mass spectrometry (GS-MS) is the preferred confirmatory method . Clinical and Professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are used.

Prescription Use X AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Tracy Phillips
Division Sign-Off

Office of In Vitro Diagnostic Device
Evaluation and Safety

510(k) K051579