



February 10, 2023

Vascular Solutions, Inc.
Linda Busklein
Senior Regulatory Affairs Associate
6464 Sycamore Court
Minneapolis, Minnesota 55369

Re: K053054
Trade/Device Name: Thrombix 3x3 hemostatic pad
Regulatory Class: Unclassified
Product Code: QSX

Dear Linda Busklein:

The Food and Drug Administration (FDA) is sending this letter to notify you of an administrative change related to your previous substantial equivalence (SE) determination letter dated November 23, 2005. Specifically, FDA is updating this SE Letter because FDA has better categorized your device technology under product code QSX.

Please note that the 510(k) submission was not re-reviewed. For questions regarding this letter please contact Julie Morabito, OHT4: Office of Surgical and Infection Control Devices, 240-402-3839, Julie.Morabito@fda.hhs.gov.

Sincerely,

Julie A. Morabito -S

Julie Morabito, Ph.D.
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

NOV 23 2005

Linda Busklein
Senior Regulatory Affairs Associate
Vascular Solutions, Inc.
6464 Sycamore Court
Minneapolis, Minnesota 55369

Re: K053054
Trade/Device Name: Thrombix 3x3 hemostatic pad
Regulatory Class: Unclassified
Product Code: FRO
Dated: October 28, 2005
Received: November 4, 2005

Dear Ms. Busklein:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Mark N. Melkerson". The signature is written in a cursive style with a large initial "M".

Mark N. Melkerson
Acting Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use Statement

510(k) Number: K053054

Device Name: **Thrombix 3x3 hemostatic pad**

Indications for Use:

The Thrombix 3X3 pad is applied topically and is indicated as a trauma dressing for temporary control of moderate to severely bleeding wounds and for the control of surface bleeding from vascular access sites and percutaneous catheters or tubes. ✓

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


Barbara Buchman
(Division Sign-Off)

**Division of General, Restorative,
and Neurological Devices**

Page 1 of _____

510(k) Number K053054

K053054

page 1/1

NOV 23 2005

510(K) SUMMARY

Common/Usual Name: Topical Hemostatic

Product Trade Name: Thrombix 3x3 hemostatic pad

Classification Name: Unclassified, Product Code FRO

Manufacturer: Vascular Solutions, Inc.
6464 Sycamore Court
Minneapolis, Minnesota 55369

Establishment Registration: 2134812

Contact: Linda Busklein
Sr. Regulatory Affairs Associate
(763) 656-4349 phone
(763) 656-4250 fax

Performance Standards: No performance standards have been developed under section 514 for this device.

Device Description: The Thrombix 3x3 hemostatic pad consists of a lyophilized pad containing thrombin, sodium carboxymethylcellulose, and calcium chloride in a non-woven gauze.

Intended Use: The Thrombix 3X3 pad is applied topically and is indicated as a trauma dressing for temporary control of moderate to severely bleeding wounds and for the control of surface bleeding from vascular access sites and percutaneous catheters or tubes.

Summary of Non-Clinical Testing: Tests conducted included assessment of the potlife of the pad after addition of up to 10 mls of saline.

Predicate Devices: D-Stat Dry 3x3 hemostatic pad (K040510)
ThrombiGel thrombin/gelatin foam hemostat (K050511)

Conclusions: The Thrombix 3x3 hemostatic pad is substantially equivalent to the currently marketed D-Stat Dry 3x3 hemostatic pad and the ThrombiGel thrombin/gelatin foam hemostat, based on a comparison of the indications for use and the technological characteristics of the device.