

K060524

JUN 15 2006

16. 510(k) SUMMARY OF SAFETY AND EFFECTIVENESS

**Dolphin 2000™ Datex Compatible Oximeter sensors
2/16/06**

Submitter (Consultant name and Address)

Bill Curnan
9433 S. Morning Glory Lane
Highlands Ranch, CO 80130

Phone: 720-939-6482
Fax: 786-551-8221

Sponsor Company Name and Address and Contact Person

Dolphin Medical Inc.
12525 Chadron Avenue
Hawthorne, CA 90250

Tammy Conway, QA Manager
phone: (310) 978-0516
fax: (310) 978-1816

Manufacturing Facility Names and Addresses

Opto Sensors (M) Sdn. Bhd.
No. 6 Jalan Angkasa Mas 1
Tabrau Industrial Estate II
81100 Johor Bahru, Malaysia

OSI Systems Pvt. Ltd.
B1, Vikram Puri
Secunderbad, Andhra Pradesh,
5000076, India

Common, Classification & Proprietary Names

Common Name: oximetry sensor
Classification Name: oximeter
Proprietary Name: Dolphin™ 2000 Oximetry Sensors

Predicate Devices

Model #	Predicate Model #	Predicate 510(k) #	Predicate Model #	Predicate 510(k) #
2030 Sensor	Model 2010	K030952		
2331 Sensor	2351	K012989	251-1	K970769
2332 Sensor	2352	K012989	251-2	K970769
2333 Sensor	2353	K012989	251-3	K970769
2334 Sensor	2354	K012989	251-4	K970769
2413 & 2423 Extension Cables	2421	K030952		

Device Description

The Dolphin 2000™ Datex Compatible Oximetry Sensors are fully compatible disposable and re-usable replacement sensors for use with major brands of pulse oximeter monitors. They represent a design change to the existing Dolphin Aristo Datex Compatible Sensors and the Dolphin 2000 Nellcor Compatible Sensors.

The disposable Dolphin 2000 Oximetry Sensors are constructed in a similar manner to predicate devices. The emitter and detector diodes are embedded in a laminate of tapes that is connected to the cable assembly. The sensors have an adhesive bandage backing that allows the sensor to be applied to the patient by wrapping it around a finger or toe (measurement site). Four sizes of disposable Dolphin 2000 Oximetry Sensors are available, which are indicated for use for adult, pediatric, infant and neonatal application sites. The Dolphin 2000 disposable sensors are provided non-sterile for single patient use. Disposable sensors are connected to the host oximeter through the use of an adapter cable.

The re-usable Dolphin 2000 Finger Clip Oximetry Sensor is an adult-sized clothespin-style clip that is placed on the end of a finger. The finger clip sensor consists of the emitter and detector components mounted in opposing clip halves, maintained in mild compression by a spring hinge. The molded outer components house the optoelectric components within contoured pads that maintain contact with the patient's finger. Clear windows within these pads permit the optical energy to pass through the finger for the measurements.

Intended Use

The Dolphin 2000™ Oximetry Sensors are indicated for use in continuous monitoring of arterial oxygen saturation and pulse rate.

Technological Characteristics Comparison

The Dolphin 2000™ Oximetry Sensors are substantially equivalent in intended use, design, principles of operation, materials, and performance to commercially available oximetry sensors.

All of the Dolphin 2000™ oximetry sensors and the predicate devices operate on the identical principles of non-invasive optical assessment of tissue oxygenation using emitters (LEDs) and detectors (photodiode).

The Dolphin 2000™ oximetry sensors are designed, configured, and manufactured for full compatibility for use with the labeled, commercially-available oximetry monitors. They are constructed of similar materials and components of equivalent specifications as used in the predicate devices.

The Dolphin 2000™ oximetry sensors, like the predicate devices are available in both disposable and re-usable styles, labeled for use in adult, pediatric, infant and neonatal populations.

The labeled accuracy of the Dolphin 2000 sensors is equivalent to those of the predicate devices.

Performance Testing

▪ **Biocompatibility**

The sensors in this application utilize only previously approved skin contact materials. Biocompatibility tests, appropriate for skin-contacting devices for prolonged exposure, were performed on each of the device components used in the assembly of the Dolphin 2000™ pulse oximetry sensors. Test results demonstrated the materials to be non-toxic, non-irritant, and non-sensitizing.

▪ **Electrical Safety**

The Dolphin 2000 Oximetry Sensors have been tested and found to comply with the applicable clauses of the following standards:

- EN 60601-1 (1990) Medical electrical equipment - part 1: General requirements for safety
- ASTM F1415-92 Standard Specification for Pulse Oximeters

Clinical Testing

The sensors were validated in breathe-down protocols at the VA Hospital of Wisconsin – Milwaukee, (Dr. Phillip Clifford, MD.). Scientific accuracy was demonstrated by statistically comparing Dolphin 2000 SpO₂ values to functional SaO₂ values. Volunteers participated in the breathe-down protocol at rest (i.e. no motion) while fully conscious at SaO₂ values ranging from 70-100%. Data was analyzed to determine the ARMS for each probe.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUN 15 2006

Dolphin Medical Incorporated
C/O Mr. Bill Curnan
Regulatory Specialist
9433 S. Morning Glory Lane
Highlands Ranch, Colorado 80130

Re: K060524

Trade/Device Name: Dolphin 2000 Oximetry Sensors
Regulation Number: 21 CFR 870.2700
Regulation Name: Oximeter
Regulatory Class: II
Product Code: DQA
Dated: May 6, 2006
Received: May 22, 2006

Dear Mr. Curnan

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number (if known):

Device Name: Dolphin 2000 Oximetry Sensors

Indications For Use:

The Dolphin 2000 Oximetry Sensors are indicated for use in continuous monitoring of arterial oxygen saturation and pulse rate.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

[Handwritten Signature]

(Division Sign- C)
Division of Physiology, General Hospital,
Infection Control, Dental Devices

510(k) Number: K 060524