

K061312



JUN - 6 2006

510(k) Summary

Applicant/Sponsor: Biomet Manufacturing Corp.

Contact Person: Patricia Sandborn Beres
Senior Regulatory Specialist

Proprietary Name: 36mm BioloX® *delta* Ceramic Heads

Common or Usual Name: Ceramic Modular Head

Classification Name: Hip joint/ceramic/polymer semiconstrained cemented or non-cemented prosthesis

Legally Marketed Devices To Which Substantial Equivalence Is Claimed:
BioloX® *delta* Ceramic Heads – K042091 & K051411

Device Description: BioloX® *delta* Ceramic Heads are designed to be the bearing surface of a total hip joint replacement. The highly polished spherical surface articulates with a polyethylene liner acetabular component. The modular head attaches to a metallic femoral stem with a Biomet Type I taper. This submission covers 36mm diameter heads with neck lengths of -3, 0, +3 and +6.

Indications For Use: BioloX® *delta* Ceramic Heads are indicated for use in total hip replacement with cemented or non-cemented femoral and acetabular components in cases of:

- 1) Non-inflammatory degenerative joint disease including osteoarthritis and avascular necrosis
- 2) Rheumatoid arthritis
- 3) Correction of functional deformity
- 4) Treatment of non-union, femoral neck fracture, and trochanteric fractures of the proximal femur with head involvement, unmanageable by other techniques.
- 5) Revision procedures where other treatment or devices have failed.

Specific indications for compatible components that can be used with the above modular heads include:

Salvage/Oncology Hip and Total Femur System components are also indicated for cases of ligament deficiency, tumor resection, trauma and revision of unsuccessful osteotomy or arthrosis.

MAILING ADDRESS
P.O. Box 587
Warsaw, IN 46581-0587

SHIPPING ADDRESS
56 E. Bell Drive
Warsaw, IN 46582

OFFICE
574.267.6639

FAX
574.267.8137

E-MAIL
biomet@biomet.com

Interlocking hip stems are indicated for non-cemented application in cases of revision, trauma, fracture, oncology or other situations where severe proximal bone loss may compromise the fixation and stability of a standard type hip replacement prosthesis.

Summary of Technologies: The 36mm Biolox® *delta* Ceramic Heads are technologically similar to the predicate devices.

Non-Clinical Testing: All parameters of the "Guidance Document for the Preparation of PreMarket Notifications for Ceramic Ball Hip Systems" were met for the devices contained in this 510(k).

Clinical Testing: None provided



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUN - 6 2006

Biomet Manufacturing Corp.
c/o Ms. Patricia Sandborn Beres,
Senior Regulatory Specialist
P.O. Box 587
Warsaw, Indiana 46581-0587

Re: K061312

Trade/Device Name: 36mm Biolox[®] *delta* Ceramic Heads

Regulation Number: 21 CFR 888.3353

Regulation Name: Hip joint metal/ceramic/polymer semi-constrained cemented or nonporous uncemented prosthesis

Regulatory Class: Class II

Product Code: LZO

Dated: May 9, 2006

Received: May 10, 2006

Dear Ms. Beres:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21

CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



for Mark N. Melkerson

Director
Division of General, Restorative and
Neurological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K061312

Device Name: 36mm BioloX® delta Ceramic Heads

Indications For Use: BioloX® delta Ceramic Heads are indicated for use in total hip replacement with cemented or non-cemented femoral and acetabular components in cases of:

- 1) Non-inflammatory degenerative joint disease including osteoarthritis and avascular necrosis
- 2) Rheumatoid arthritis
- 3) Correction of functional deformity
- 4) Treatment of non-union, femoral neck fracture, and trochanteric fractures of the proximal femur with head involvement, unmanageable by other techniques.
- 5) Revision procedures where other treatment or devices have failed.

Specific indications for compatible components that can be used with the above modular heads include:

Salvage/Oncology Hip and Total Femur System components are also indicated for cases of ligament deficiency, tumor resection, trauma and revision of unsuccessful osteotomy or arthrosis. (K974558, K002757, K021380, K033871)

Interlocking hip stems are indicated for non-cemented application in cases of revision, trauma, fracture, oncology or other situations where severe proximal bone loss may compromise the fixation and stability of a standard type hip replacement prosthesis. (K990830, K042774)


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use No
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

**Division of General, Restorative,
and Neurological Devices**

Page 1 of 1

510(k) Number K061312