

K 061768

**SUMMARY OF SAFETY AND EFFECTIVENESS  
MP-1 MICRO PERIMETER**

SEP 28 2006

**SUBMITTER INFORMATION**

- A. Company Name: Nidek Technologies Srl
- B. Company Address: Via Dell'Artigianato, 6/A  
Albignasego (Padova), Italy 35020
- C. Company Phone: +39 049 86 29 200  
Company Fax: +39 049 86 26 824
- D. Contact Person: Mr. Aldo Cocchiglia  
Managing Director  
Nidek Technologies Srl
- E. Date Summary Prepared: August 08, 2006

**DEVICE IDENTIFICATION**

- A. Generic Device Name: Ophthalmic camera, AC-powered  
Automated Micro Perimeter, AC-powered
- B. Trade/Proprietary Name: MP-1 MICRO PERIMETER
- C. Classification: Class II
- D. Product Code: HKI , HPT

K06 1768

## DEVICE DESCRIPTION

The Nidek Technologies Srl MP-1 MICRO PERIMETER is an instrument for the diagnosis of retinal diseases. It is able to capture infrared live sequences and color images of the patient retina through CCD cameras and at the same time to project light stimuli on the retina in those retinal positions chosen by the operator.

The MP-1 MICRO PERIMETER is supplied with Xenon light and Halogen light sources suitable for performing the above described examinations.

The system includes also a suitable Insulation transformer, a LCD S-VGA Monitor and an IBM Personal Computer, which incorporates the dedicated NAVIS software (with Windows 2000 operative system).

## INTENDED USE

The MP-1 MICRO PERIMETER is intended for use as:

1) *Color retinography*

Color image of the fundus obtained in non-mydratic conditions, using an IR sensible camera as a viewfinder and a visible flash for illuminating the retina at picture taking.

2) *Fixation exam*

This examination consists in:

- locating, in the patient's fundus, his/her fixation site (anatomical information);
- recording the positions, during a certain interval, of the point at which the patient is fixating, as a measure of fixation stability (functional information).

3) *Fundus-related microperimetry*

Consists in:

- projecting light stimuli on patient retina at given positions;
- recording the patient's subjective answer to each stimulus (seen/ not seen);
- associating precisely the subjective answers (functional information) with the retinal location of the stimuli (anatomical information), compensating properly for eye fundus movements during the examination;
- generating a sensitivity map of the inspected retinal region.

V. 061768

**4) Visual rehabilitation**

The biofeedback tool consists of a software module which implements a new exam mode, called the "feedback exam".

The feedback exam represents a visual rehabilitation tool, which helps the doctor training patients with unstable fixation to develop a new preferred retinal locus (PLR) being led by an audible bio-feedback signal

**SUBSTANTIAL EQUIVALENCE**

The Nidek Technologies Srl MP-1 MICRO PERIMETER device is of comparable type and is substantially equivalent to the following predicate device:

Predicate Device	510(K) Holder	510(k) No.	Date Cleared
MP-1 MICRO PERIMETER	NIDEK TECHNOLOGIES SRL	K023719	12 / 23 / 2002
AR-1 PHYSIOLOGICAL PRODESSOR	ATAL RESEARCHES LTD.	K810880	05 / 27 / 1981

In further support of a substantial equivalence determination, Section 5 provides a comparison chart of the modified MP-1 MICRO PERIMETER and the predicate devices.

**TECHNOLOGICAL CHARACTERISTICS**

A comparison of the technological characteristics of the modified MP-1 MICRO PERIMETER and the predicate devices has been performed. The results of this comparison demonstrate that the modified MP-1 MICRO PERIMETER device is equivalent to the marketed predicate devices.

**PERFORMANCE DATA**

The performance data indicate that the modified MP-1 MICRO PERIMETER device meets all specified requirements, and is substantially equivalent to the predicate device.



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

SEP 28 2006

Nidek Technologies SRL  
c/o Mr. Aldo Cocchiglia  
Via Dell' Artigianato, 6/A  
Albignasego (Padova), Italy 35020

Re: K061768

Trade/Device Name: MP-1 Micro Perimeter  
Regulation Number: 21 CFR 886.1120  
Regulation Name: Ophthalmic camera  
Regulatory Class: II  
Product Code: HKI, HPT  
Dated: August 11, 2006  
Received: August 23, 2006

Dear Mr. Cocchiglia:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Malvina B. Eydelman, M.D.

Director

Division of Ophthalmic and Ear, Nose  
and Throat Devices

Office of Device Evaluation

Center for Devices and  
Radiological Health

Enclosure

---

## INDICATIONS FOR USE

510(k) Number (if known): K061768

Device Name: MP-1 MICRO PERIMETER

Indications for Use: The MP-1 MICRO PERIMETER is indicated for use as :

- *Color retinography*
- *Fixation examiner*
- *Fundus-related microperimetry*
- *Visual rehabilitation.*

Prescription Use   
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

*Denis L. McCarthy*

(Division Sign-Off)

Division of Ophthalmic Ear,  
Nose and Throat Devices

510(k) Number K061768