

K062046

5. 510(k) Summary



JAN 19 2007

General Information

Trade Name	Modified Merci [®] Retriever
Common Name	Endovascular Retriever
Classification	Class II, Catheter, Thrombus Retriever per 21 CFR § 870.1250
Submitter/Owner	Concentric Medical, Inc. 1380 Shorebird Way Mountain View, CA 94043 Phone: 650-938-2100 Fax: 650-938-2700
Establishment Registration Number	2954917
Contact	Kirsten Valley Senior Vice President, Regulatory Affairs and Operations
Prepared	July 18, 2006

Intended Use

The Modified Merci Retriever is intended to restore blood flow in the neurovasculature by removing thrombus in patients experiencing an ischemic stroke. Patients who are ineligible for treatment with intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment. The Modified Merci Retriever is also indicated for use in the retrieval of foreign bodies misplaced during interventional radiological procedures in the neuro, peripheral and coronary vascular systems.

Predicate Device

Merci[®] Retriever X6

K062046

Device Description and Technology Comparison to Predicate Device

Like the predicate device, the Modified Merci Retriever consists of a flexible, Nitinol core wire with shaped distal loops covered by a platinum coil for radiopacity. It is coated with a hydrophilic lubricious coating.

Both the Modified Merci Retriever and the predicate device are placed distal to the thrombus or foreign body using a microcatheter which is retracted to deploy the loops of the Retriever. The Retriever and microcatheter are pulled back into the thrombus or foreign body, engaging it in the loops of the Retriever. The Retriever, the thrombus or foreign body, and the microcatheter are then removed from the body.

Materials

All materials used in the manufacture of the Modified Merci Retriever are suitable for the intended use of the device and have been used in numerous previously cleared products.

Testing Summary

The Modified Merci Retriever was tested in the same manner as the predicate device. All devices met the required specifications for the completed tests. The Modified Merci Retriever was designed under the Concentric Quality System that is in compliance with 21CFR§820.30.

Summary of Substantial Equivalence

The Modified Merci Retriever is substantially equivalent to the predicate device, the Merci Retriever X6.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Concentric Medical, Inc.
% Ms. Kirsten Valley
Senior VP, Operations and
Regulatory Affairs
1380 Shorebird Way
Mountain View, California 94043

JAN 19 2007

Re: K062046
Trade/Device Name: Modified Merci[®] Retriever
Regulation Number: 21 CFR 870.1250
Regulation Name: Percutaneous catheter
Regulatory Class: II
Product Code: NRY
Dated: November 17, 2006
Received: November 20, 2006

Dear Ms. Valley:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Ms. Kirsten Valley

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read 'F-A Melkerson', with a long horizontal flourish extending to the right.

Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

4. Indications for Use Statement

510(k) Number (if known): (intentionally left blank) K062046

Device Name: Modified Merci® Retriever

Indications for Use: The Modified Merci Retriever is intended to restore blood flow in the neurovasculature by removing thrombus in patients experiencing ischemic stroke. Patients who are ineligible for treatment with intravenous tissue plasminogen activator (IV t-PA) or who fail IV t-PA therapy are candidates for treatment. The Modified Merci Retriever is also indicated for use in the retrieval of foreign bodies misplaced during interventional radiological procedures in the neuro, peripheral and coronary vasculature.

Prescription Use X
(Per 21 CFR 801.Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
**Division of General, Restorative,
and Neurological Devices**
510(k) Number K062046