

Attachment 1
Summary of Safety and Effectiveness

K062451

Page 1 of 2

This 510(k) summary of Safety and Effectiveness information is submitted in accordance with the requirements of 21 CFR Part 807.87(h)

General Information:

Device Name:	NeuViz Dual Multi-slice CT Scanner System
Classification Name:	21 CFR Part 892.1750 Computed Tomography X-ray System
Model Name	NeuViz Dual (H); NeuViz Dual (L)
Classification:	Class II
Performance Standard:	21 CFR Subchapter J, Federal Diagnostic X-ray Equipment Standard
Manufacture:	PHILIPS AND NEUSOFT MEDICAL SYSTEMS CO., LTD. Neusoft Park, Hun Nan Industrial Area, Shenyang 110179, P.R.China
Distributor:	NEUSOFT MEDICAL SYSTEMS CO., LTD. No.3-11, Wenhua Road, Heping District, Shenyang, P.R.China Post Code : 110004
Submitter:	Contact : Tianyanfang Title : Manager of Q&R Department Tel : 86-24-83660649 Fax : 86-24-83780480 E-Mail : Tianyanfang@neusoft.com

Summary prepared : September 10th, 2006

K062451

Page 2 of 2

Safety and Effectiveness information**Intended Uses:**

The NeuViz Dual systems are intended to produce cross-section images of head and whole body by computer reconstruction of X-ray transmission data taken at different angles.

Device Description:

The NeuViz Dual systems are whole body X-ray computed tomography scanner which features a continuously rotating tube-detector system and multi-slice capability of 2 slices simultaneously. There are two variants with different tube capacity: NeuViz Dual (H) and NeuViz Dual (L). The system software is an interactive program used for X-ray scan control, image reconstruction, and image archive/evaluation.

Predicated Device:

CT-C3000DUAL and CT-C2800DUAL Family of Dual-slice CT Scanner Systems (K041542)

Statement of Substantial Equivalence:

The NeuViz Dual systems are of comparable type and substantially equivalent to the CT-C3000DUAL and CT-C2800DUAL Family of Dual-slice CT Scanner Systems (K041542) that comply with the same or equivalent standards and has the same intended uses. Both of these system use on-board high frequency High-Voltage generator to generate X-radiation from X-ray tube. The X-ray transmission data is detected by the solid-state detector and is reconstructed by the computer which has an interactive user interface. Both of these devices produce two dimensional image and 3D image that can be filmed or electronically stored for future review.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
9200 Corporate Blvd.
Rockville MD 20850

SEP 28 2006

Neusoft Medical Systems Co., Ltd.
c/o Mr. Tamas Borsai
TÜV Rheinland of North America
12 Commerce Road
NEWTOWN CT 06470

Re: K062451
Trade/Device Name: NeuViz Dual Multi-slice CT Scanner System
Regulation Number: 21 CFR §892.1750
Regulation Name: Computed tomography x-ray system
Regulatory Class: II
Product Code: JAK
Dated: September 12, 2006
Received: September 13, 2006

Dear Mr. Borsai:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.



Protecting and Promoting Public Health

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at one of the following numbers, based on the regulation number at the top of this letter:

21 CFR 876.xxx	(Gastroenterology/Renal/Urology	240-276-0115
21 CFR 884.xxx	(Obstetrics/Gynecology)	240-276-0115
21 CFR 894.xxx	(Radiology)	240-276-0120
Other		240-276-0100

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Nancy C. Brogdon
Director, Division of Reproductive,
Abdominal, and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

**Attachment 2
Indications for Use**Page 1 of 1510(k) Number (if known): K062451

Device Name: NeuViz Dual Multi-slice CT Scanner System

Indications for Use:

The NeuViz Dual Multi-slice CT Scanner System -**NeuViz Dual (H)** and **NeuViz Dual (L)** - are intended to produce cross-section images of head and body by computer reconstruction of X-ray transmission data taken at different angles.

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ✓Nancy C Brogdon
(Division Sign-Off)Division of Reproductive, Abdominal,
and Radiological Devices510(k) Number K062451

Indications for Use