

510(k) Summary

K 062589

Pursuant to CFR 807.92, the following 510(k) Summary is provided:

1. (a) **Submitter's Address:** George J. Hattub
MedicSense, USA
291 Hillside Avenue
Somerset, MA 02726
1. (b) **Manufacturer Address:** Inolase 2002, Ltd.
13 Giborei Israel St., PO Box 8497
Netanya (New Industrial Area)
Israel, 42504
- Mfg. Phone:** 972-9-865-6750
- Contact Person:** Raphi Shavit, CEO
- Date:** Revised- November 12, 2006
2. **Device & Classification Name:** Laser Handpiece Accessory System (Class 2), Product Code GEX, 21 CFR 878.4810 – Tradename of device: Serenity PSF™ (Pneumatic Skin Flattening) System
3. **Predicate Device:** Aesthera AIP™ Intense Pulsed Light System (K041554)
4. **Description:** The Serenity PSF™ (Pneumatic Skin Flattening) System is a device which has the ability to produce a vacuum at a level of 200-700 mmHg. Suction can be delivered through its handpiece, which has a sapphire window at its distal end, to be positioned over the skin for laser or IPL treatment. When the PFS system is activated, the negative pressure results in the flattening of the skin of the treatment site. During the time duration of suction, the treatment beam of the laser or IPL is administered through the sapphire window of the handpiece, which is transparent and thermally conductive. The resultant tight mechanical contact produced by the PSF expels blood from the treatment site which enhances the light penetration of the laser or IPL in tissue as well as the removal of its heat from the skin. This can reduce the possibility of post treatment erythema.
5. **Intended Use:** The Serenity PSF™ (Pneumatic Skin Flattening) System is an accessory for a compatible legally market Laser or Intense Pulse Light System for use in hair removal. Its handpiece produces a negative pressure over the skin surface just prior to the administration of the treatment beam through it.
6. **Comparison of Technological Characteristics:** With respect to technology, the Serenity PSF™ (Pneumatic Skin Flattening) System is substantially equivalent to one of its predicate device in that it produces a pneumatic connection of the treatment site for the delivery of photonic treatment. When the PSF™ System is used in combination with a legally marketed device such as an IPL system, it also has the same basic intended use as this predicate device.

NOV 30 2006



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Inolase 2002, Ltd.
% MedicSense, USA
Mr. George J. Hattub, RAC & CQE
Senior Staff Consultant
291 Hillside Avenue
Somerset, Massachusetts 02726

NOV 30 2006

Re: K062589

Trade/Device Name: Serenity PSF™ (Pneumatic Skin Flattening) System
Regulation Number: 21 CFR 878.4810
Regulation Name: Laser surgical instrument for use in general and plastic surgery
and in dermatology
Regulatory Class: II
Product Code: GEX
Dated: September 21, 2006
Received: September 25, 2006

Dear Mr. Hattub:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set


Page 2 – Mr. George J. Hattub, RAC & CQE

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K062589

Device Name: Serenity PSF™ (Pneumatic Skin Flattening) System

Indications For Use: The Serenity PSF™ (Pneumatic Skin Flattening) System is an accessory for a compatible legally marketed Laser or Intense Pulse Light System for use in hair removal. Its handpiece produces a negative pressure over the skin surface, just prior to the administration of the treatment beam through it.

(revised 11/12/06)


Prescription Use
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

Page 1 of 1

**Division of General, Restorative,
and Neurological Devices**

510(k) Number K062589