

VII. 510(k) Summary

JAN 24 2007

In accordance with Title 21 of the Code of Federal Regulations, Part 807, and in particular §807.92, the following summary of information is provided:

A. Submitted by:

Laetitia Cousin
Director of Regulatory Affairs and Quality Assurance
NuVasive, Incorporated
4545 Towne Centre Court
San Diego, California 92121
Telephone: (858) 909-1868
Fax: (858) 909-2068

B. Device Name

Trade or Proprietary Name: *NuVasive NeuroVision JJB System*
Common or Usual Name: Electromyography (EMG) monitor/stimulator
Classification Name: Surgical nerve stimulator/locator
Device Class: Class II
Classification: §874.1820, §882.1870
Product Code: 77ETN, 84GWF

C. Predicate Devices

The subject *NeuroVision JJB System* is substantially equivalent to the *NeuroVision JJB System* currently manufactured and distributed commercially in the U.S. by NuVasive.

D. Device Description

The *NVJJB System* utilizes conventional neurophysiologic monitoring to reduce the incidence of injury to nerve roots during instrumented spine surgery. In the procedure, stimulus evoked electromyography is used to determine changes in nerves. Corresponding muscle groups are monitored using surface electrodes, while stimulation is used to detect nerve responses.

The NeuroVision JJB System consists of a reusable Patient Module, a Control Unit comprised of an embedded computer with touch screen controls and an interface card, and an assortment of disposable and reusable conductive probes, electrodes, and electrode leads.

NuVasive®, Inc.

E. Intended Use

The NeuroVision JJB System is used for intraoperative monitoring and neurological status assessment by the administration of brief electrical stimulus pulses to neural tissues and the EMG monitoring of the associated muscle groups. The System is used in conjunction with other NuVasive devices to assist in gaining controlled access to, and visualization of, the spine.

F. Comparison to Predicate Devices

The subject device has indications for use identical to those of its predicate, and employs the same principles of operation. Due to this equivalency, the device raises no new safety or effectiveness issues.



JAN 24 2007

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

NuVasive, Incorporated
c/o Ms. Laetitia Cousin
Director of Regulatory Affairs and Quality Assurance
4545 Towne Centre Court
San Diego, California, 92121

Re: K062765
Trade Name: NuVasive NeuroVision® JJB System
Regulation Number: 21 CFR 882.1870
Regulation Name: Evoked response electrical stimulator
Regulatory Class: II
Product Code: GWF
Dated: December 26, 2006
Received: December 28, 2006

Dear Ms. Cousin:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 - Ms. Laetitia Cousin

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read 'Mark N. Melkerson', with a large, sweeping flourish extending to the right.

Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

D. Indications for Use Statement

510(k) Number (if known): K062765

Device Name: NeuroVision JJB System

Indications for Use:

The NeuroVision JJB System is used for intraoperative monitoring and neurological status assessment by the administration of brief electrical stimulus pulses to neural tissues and the EMG monitoring of the associated muscle groups. The System is used in conjunction with other NuVasive devices to assist in gaining controlled access to, and visualization of, the spine.

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use _____
(Per 21 CFR 801.109)

OR

Over-The-Counter Use _____


(Division Sign-Off)
**Division of General, Restorative,
and Neurological Devices**

510(k) Number K062765