

K062807

TSRH® Spinal System  
510(k) Summary  
September 2006

NOV - 6 2006

I. **Company:** Medtronic Sofamor Danek USA, Inc.  
1800 Pyramid Place  
Memphis, TN 38132  
(901) 396-3133

**Contact:** Edward S. Chin, D. Ph, MBA  
Group Director, Regulatory and Clinical Affairs

II. **Proposed Proprietary Trade Name:** TSRH® Spinal System

III. **Classification Name(s)/Product Code(s):**

Regulation Names: Spinal Interlaminar Fixation Orthosis, Spinal Intervertebral Body Fixation Orthosis, Spondylolisthesis Spinal Fixation Device System, and Pedicle Screw Spinal System.

Regulation Numbers: 888.3050, 888.3060 and 888.3070

Product Codes: NKB, KWP, MNH and MNI

IV. **Description**

The purpose of this 510(k) submission is to add modified connectors to the TSRH® Spinal System. The TSRH® Spinal System is intended to help provide immobilization and stabilization of spinal segments as an adjunct to fusion of the thoracic, lumbar, and/or sacral spine.

The TSRH® Spinal System consists of a variety of shapes and sizes of rods, hooks, screws, cross connectors, and connecting components. In addition, GDLH® rods, DYNALOK® PLUS bolts, CD HORIZON® Low Profile MULTI-SPAN® CROSSLINK® Plates, GDLH® rod/bolt connectors, GDLH® Variable Angle T-Bolts, and GDLH® and CD HORIZON® set screws and locking screws may be used with the TSRH® Spinal System.

The TSRH® Spinal System implant components can be rigidly locked into a variety of configurations, with each construct being tailor-made for the individual case. The hooks are intended for posterior use only and the staples are for anterior use only. The TSRH-3D® connectors and TSRH-3D® screws are intended for posterior use only.

The TSRH® Spinal System components are fabricated from stainless steel. Alternatively, they may be fabricated from medical grade titanium alloy or medical grade titanium. The TSRH® Spinal System may be sold sterile or non-sterile.

V. **Indications for Use:**

When used as a pedicle screw fixation system in the non-cervical posterior spine in skeletally mature patients, the TSRH® Spinal System is indicated for one or more of the following: (1) degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) degenerative spondylolisthesis with objective evidence of neurologic impairment, (3) fracture, (4) dislocation, (5) scoliosis, (6) kyphosis, (7) spinal tumor, and/or (8) failed previous fusion (pseudarthrosis).

In addition, when used as a pedicle screw fixation system, the TSRH® Spinal System is indicated for skeletally mature patients: (1) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) joint; (2) who are receiving fusions using autogenous bone graft only; (3) who are

Page 1 of 2

having the device fixed or attached to the lumbar and sacral spine (L3 and below); and (4) who are having the device removed after the development of a solid fusion mass.

When used as a posterior, non-cervical, non-pedicle screw fixation system, the TSRH<sup>®</sup> Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spondylolisthesis, (3) fracture, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) spinal stenosis, (6) pseudarthrosis, (7) tumor resection, and/or (8) unsuccessful previous attempts at spinal fusion.

For anterior use only the TSRH<sup>®</sup> Spinal System has the additional indications of: (1) spinal stenosis and/or, (2) spondylolysis.

**VI. Substantial Equivalence**

This submission reflects merely a wording change and therefore no additional testing is required. The modifications to the package insert and to the wording of the indications of use do not alter the safety and effectiveness of the device, nor do they impose any additional risks to the patient. Wording for the modified contraindications is equivalent to that found in package inserts for similar spinal device system including the Moss Miami Spinal System (DePuySpine - K030383, SE 02/26/03) and the Pangea<sup>™</sup> Degenerative Spine System (Synthes - K052151, SE 12/07/05).



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Medtronic Sofamor Danek, Inc.  
% Edward Chin, D.Ph, MBA  
Group Director, Regulatory and Clinical Affairs  
1800 Pyramid Place  
Memphis, Tennessee 38132

NOV - 6 2006

Re: K062807  
Trade/Device Name: TSRH® Spinal System  
Regulation Number: 21 CFR 888.3070  
Regulation Name: Pedicle screw spinal system  
Regulatory Class: Class III  
Product Code: NKB, MNH, MNI, KWP  
Dated: September 18, 2006  
Received: September 19, 2006

Dear Dr. Chin:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 – Edward Chin, D.Ph, MBA

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at 240-276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in cursive script, appearing to read "Mark N. Melkerson".

Mark N. Melkerson  
Director  
Division of General, Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

510(k) Number (if known): K062807

Device Name: TSRH® Spinal System

**Indications for Use:**

When used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, the TSRH® Spinal System is indicated for one or more of the following: (1) degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) degenerative spondylolisthesis with objective evidence of neurologic impairment, (3) fracture, (4) dislocation, (5) scoliosis, (6) kyphosis, (7) spinal tumor, and/or (8) failed previous fusion (pseudarthrosis).

In addition, when used as a pedicle screw fixation system, the TSRH® Spinal System is indicated for skeletally mature patients: (1) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (2) who are receiving fusions using autogenous bone graft only; (3) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); and (4) who are having the device removed after the development of a solid fusion mass.

When used as a posterior, non-cervical, non-pedicle screw fixation system, the TSRH® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spondylolisthesis, (3) fracture, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) spinal stenosis, (6) pseudarthrosis, (7) tumor resection, and/or (8) unsuccessful previous attempts at spinal fusion.

For anterior use only the TSRH® Spinal System has the additional indications of: (1) spinal stenosis and/or, (2) spondylolysis

Prescription Use   X   AND/OR Over-The-Counter Use \_\_\_\_\_  
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

**(Division Sign-Off)**

**Division of General, Restorative,  
and Neurological Devices**

510(k) Number K062807

*page 1 of 1*