

K063477

Non-Confidential Summary of Safety and Effectiveness

Page 1 of 2

15-Nov-06

NovaMed LLC
4 Nursery Lane
Rye, NY 10580

Tel – 914-967-3500
Fax – 914-967-3209

FEB 9 2007

Official Contact: Peter Derrico - President
Proprietary or Trade Name: NovaMed MRI Conditional Laryngoscope set
Common/Usual Name: Laryngoscope handle and blades – MR conditional
Classification Name: Rigid Laryngoscope
Device: NovaMed MRI Conditional
Predicate Devices: Minrad MR tested laryngoscope – K041852

Device Description:

The NovaMed MRI Conditional laryngoscope handle, battery, and blades are standard blades and handles that have been specially treated to be useable in a magnetic resonance (MR) environment. Testing has been performed according to ASTM F2052-02 in a 3.0 Tesla environment. There are many handle and blade configurations offered.

Indications for Use:

Indicated Use -- The NovaMed MRI Conditional laryngoscope set is used to facilitate and aid in tracheal intubation in a Magnetic Resonance (MR) environment, not to exceed a 3.0 Tesla static magnetic field.

Patient Population -- No limitations

Environment of Use -- Magnetic resonance (MR) environment up to 3.0 Tesla

Non-Confidential Summary of Safety and Effectiveness

Page 2 of 2

15-Nov-06

Device Attributes:

Design	
Configurations offered	Handles, many blades, i.e. Miller, Mac, etc.
Cleaning	Cold solution, autoclave
Testing	
ASTM F2052-02	In 3.0 Tesla < 45 degree string deflection

Differences Between Other Legally Marketed Predicate Devices

The NovaMed MRI Conditional laryngoscope set viewed as substantially equivalent to the following predicate device – Minrad – K041852.

There are no significant differences that affect the safety or effectiveness of the intended device as compared to the predicate devices.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

NovaMed, LLC
C/O Mr. Paul Dryden
ProMedic, Incorporated
3460 Pointe Creek Court, #102
Bonita Springs, Florida 34134-2015

FEB 9 2007

Re: K063477

Trade/Device Name: NovaMed MRI Conditional Laryngoscope Set
Regulation Number: 21 CFR 868.5540
Regulation Name: Rigid Laryngoscope
Regulatory Class: I
Product Code: CCW
Dated: November 15, 2006
Received: November 16, 2006

Dear Mr. Dryden:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,


Chiu Lin, Ph.D.

Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

4 **Indications for Use Statement**

Page 1 of 1

510(k) Number: K 063477 (To be assigned)

Device Name: NovaMed MRI Conditional

Indications for Use: The NovaMed MRI Conditional laryngoscope set is used to facilitate and aid in tracheal intubation in a Magnetic Resonance (MR) environment, not to exceed a 3.0 Tesla static magnetic field.

Prescription Use XX
(Part 21 CFR 801 Subpart D)

or

Over-the-counter use ___
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Susan Burns

Medical Director, General Hospital
Medical Director, Dental Division

K063477