

K070189

510(k) Summary

FEB 21 2007

Introduction According to the requirements of 21 CFR 807.92, the following information provides sufficient detail to understand the basis for a determination of substantial equivalence.

1) Submitter name, address, contact **Submitted by:**
 Disetronic Medical Systems AG
 Kirchbergstrasse 190, Postfach
 CH-3401 Burgdorf, Switzerland

United States Contact Person:
 Mike Flis
 Roche Diagnostics
 9115 Hague Road
 Indianapolis, Indiana 46250
 317-521-2110
 Mike.flis@roche.com

Date Prepared: January 15, 2007

2) Device name Proprietary name: ACCU-CHEK Ultraflex Infusion Set
 Common name: subcutaneous infusion set
 Classification name: intravascular administration set
 Product Code: FPA

3) Predicate device We claim substantial equivalence to the current legally marketed version of the same device.

4) Device Description The ACCU-CHEK Ultraflex is a disconnectable infusion set with soft cannula perpendicular to the adhesive, for transfusion of insulin into the subcutaneous tissue. The unit is designed to interface with commercially available insulin infusion pumps with suitable connections. The insulin infusion pump systems are designed to control the delivery of insulin as prescribed by a health care professional. The system (infusion set, insulin infusion pump, and insulin) is indicated for patients with insulin dependent diabetes mellitus.

5) Intended use ACCU-CHEK Ultraflex is an infusion set for the subcutaneous infusion of insulin administered with microdosage insulin pumps.

Continued on next page

510(k) Summary, Continued

6) Data demonstrating substantial equivalence

Testing of the modified ACCU-CHEK Ultraflex Infusion Set demonstrated that the device meets the requirements for its intended use. The data also demonstrates that the ACCU-CHEK Ultraflex Infusion Set is substantially equivalent to the predicate device.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

FEB 21 2007

Disetronic Medical Systems AG
C/O Mr. Mike Flis
Roche Diagnostics
9115 Hague Road
Indianapolis, Indiana 46250

Re: K070189

Trade/Device Name: ACCU-CHEK Ultraflex Infusion Set
Regulation Number: 21 CFR 880.5440
Regulation Name: Intravascular Administration Set
Regulatory Class: II
Product Code: FPA
Dated: January 15, 2007
Received: January 22, 2007

Dear Mr. Flis:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

INDICATIONS FOR USE STATEMENT

510(k) Number (if known): K070189

Device Name: ACCU-CHEK Ultraflex Infusion Set

The ACCU-CHEK Ultraflex is an infusion set for the subcutaneous infusion of insulin administered with microdosage insulin pumps.

Prescription Use X
(Per 21 CFR 801.109)

OR

Over the Counter Use _____

(PLEASE DO NOT WRITE BELOW THIS LINE -CONTINUE ON ANOTHER PAGE IF NEEDED)

Aut [Signature]

K070189