

K071143

510(k) SUMMARY

1.0 Submitted By

BD Biosciences
2350 Qume Drive
San Jose, CA 95131-1807

JUN - 6 2007

Contact:

Nobuko Nakajima
Senior Regulatory Affairs Specialist
Phone (408) 954-4109
Fax (408) 954-2495
Nobuko_Nakajima@bd.com

Submission date

April 20th, 2007

2.0 Device Name and Classification

- a) BD™ Tritest CD3/CD4/CD45
- b) 864.5220 Automated differential cell counter, GKZ class II

3.0 Intended Use

The BD TriTEST™ CD3FITC/CD4PE/CD45 PerCP reagent is a three-color, direct immunofluorescence reagent for identifying and enumerating percentages of T lymphocytes (CD3+) and T-helper/inducer (CD3+CD4+) cells in erythrocyte-lysed whole blood (LWB).

4.0 Basic description of the device

The BD TriTEST™ CD3FITC/CD4PE/CD45 PerCP reagent is a three-color, direct immunofluorescence reagent for identifying and enumerating percentages of T lymphocytes (CD3+) and T-helper/inducer (CD3+CD4+) cells in erythrocyte-lysed whole blood (LWB). If used with Becton Dickinson flow cytometers, the product can be used with MultiSET™ software for analysis as an accessory, or customers may perform analysis using CELLQuest™, CELLQuest Pro™ or LYSYS™ II software.

The reagent vials and counting bead vials are packaged separately. Each vial of this reagent yields 50 tests. Each package of counting bead tubes yields 50 tests.

5.0 Predicate Device

The BD™Tritest CD3/CD4/CD45 claims substantial equivalence to the BD™Tritest CD3/CD4/CD45 currently in distribution (FDA 510(k) number K946055/S2)

6.0 Comparison to the Predicate(s)

The modifications to the legally marketed device (BD™Tritest CD3/CD4/CD45) intends to extend the sample stability claim for EDTA from 48 to 72 hours

The Intended use and the indications of the modified device, as described in its labeling are the same as the intended use and indications for the original predicate device.

7.0 Summary of Performance Data

Performance data from validation testing supports equivalency.

This Summary of safety and effectiveness is being submitted in accordance with the requirements of compliance with SMDA 1990 and 21 CFR 807.92.



JUN - 6 2007

Nobuko Nakajima
BD Biosciences
2350 Qume Drive
San Jose, California 95131-1807

Re: k071143

Trade/Device Name: BD Tritest CD3/CD4/CD5
Regulation Number: 21 CFR 864.5220
Regulation Name: Automated differential cell counter
Regulatory Class: Class II
Product Code: GKZ
Dated: April 20, 2007
Received: April 24, 2007

Dear Mr. Nakajima:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket

Page 2 – Nobuko Nakajima

notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (240) 276-0450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150, or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>.

Sincerely yours,



Robert L. Becker, Jr., MD, PhD
Director
Division of Immunology and Hematology
Office of *In Vitro* Diagnostic Device Evaluation
and Safety
Center for Devices and Radiological Health

Enclosure

Attachment 1:

Indications for Use Statement

510(k) Number: K071143

Device Name: BD Tritest CD3/CD4/CD45

Indications For Use:

BD Tritest CD3 FITC/CD4 PE/CD45 PerCP is a three-color direct immunofluorescence reagent for identifying and enumerating percentage of mature human T lymphocytes (CD3+) and T-helper/inducer (CD3+CD4+) cells in erythrocyte-lysed whole blood (LWB).


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER
PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (OIVD)


Division Sign-Off

**Office of In Vitro Diagnostic Device
Evaluation and Safety**

510(k) K071143