

K072088

Vision-Sciences, Inc.

Special 510(k) Premarket Notification: Device Modification

July 27, 2007

Modified Trans-Nasal Esophagoscope with Digital Video Processor and Disposable EndoSheath® Systems

510(k) Summary

AUG 29 2007

Owner's Name: Vision-Sciences, Inc.
Address: 40 Ramland Road South
Orangeburg, NY 10962
Telephone Number: (845) 365-0600
Fax Number: (845) 365-0620
Contact Person: Lillian Quintero; Director QA/RA

Subject Device Name: Modified Flexible Trans-Nasal Esophagoscope with Digital Video Processor and Disposable EndoSheath® Systems
Common/Usual Name: Flexible video endoscope with video processor and disposable sheaths
Product Codes: EOX
FDA Regulations: 21 CFR 874.4710
Device Classification: Class II

Predicate Device Name: Flexible Fiberoptic Trans-Nasal Esophagoscope with EndoSheath® Systems
Common/Usual Name: Flexible fiberoptic endoscope with sheaths and accessories
Product Codes: EOX
FDA Regulations: 21 CFR 874.4710
Device Classification: Class II
Premarket Notification: K031786

Device Description

The VSI flexible endoscope is a flexible endoscope with connections to a video processor and display monitor. The EndoSheath® Systems are sterile, single-use protective sheath systems, with or without a working channel, that are intended to cover the entire insertion tube of the videoscope. The digital video processors are used with the flexible videoscope for image visualization and capture.

Intended Use

The flexible trans-nasal video esophagoscope with digital video processor and disposable sheath system is intended for use in endoscopic access and examination of the larynx, esophagus and gastro-esophageal junction. The system can also be used to aid in intubation.

Performance Testing

This submission is a Special 510(k): Device Modification as described in FDA's guidance document entitled "The New 510(k) Paradigm - Alternate Approaches to Demonstrating Substantial Equivalence in Premarket Notifications." In support of this 510(k), Vision-Sciences has provided certification of compliance to 21 CFR 820.30 Design Control requirements, and a description of the internal Risk Analysis procedure. V & V activities, including scope/sheath/processor system functional and performance testing, and software validation was addressed through comprehensive Design Validation and Verification planning.

Conclusion

Based on the indications for use, technological characteristics, performance testing and comparison to predicate devices, the VSI flexible video TNE endoscope with digital video processors and disposable EndoSheath® Systems have been shown to be safe and effective for their intended use.

010189



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Vision-Sciences, Inc.
c/o Ms. Pamela Papineau, RAC
Delphi Medical Device Consulting, Inc.
5 Whitcomb Ave.
Ayer, MA 01432

AUG 29 2007

Re: K072088

Trade/Device Name: Flexible Trans-Nasal Esophagoscope with Digital Video Processor and Disposable EndoShealth® Systems

Regulation Number: 21 CFR 874.4710

Regulation Name: Esophagoscope (flexible or rigid) and accessories

Regulatory Class: Class II

Product Code: EOX

Dated: July 27, 2007

Received: July 30, 2007

Dear Ms. Papineau:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at 240-276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at 240-276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Malvina B. Eydelman, M.D.
Director
Division of Ophthalmic and Ear, Nose
and Throat Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

K072088

Vision-Sciences, Inc.
July 27, 2007

Special 510(k) Premarket Notification: Device Modification
Modified Trans-Nasal Esophagoscope with Digital Video Processor and Disposable EndoSheath® Systems

510(k) Number (if known): _____

Device Name: Flexible Trans-Nasal Video Esophagoscope with Digital Video Processor and Disposable EndoSheath® Systems

Indications for Use:

The flexible trans-nasal video esophagoscope with EndoSheath® System is intended for use in endoscopic access and examination of the larynx, esophagus and gastro-esophageal junction. The system may also be used to aid in intubation.

The digital video processor is intended for use with the VSI flexible video scope.


Prescription Use X
(Per 21 CFR 801 Subpart D)

OR

Over-the -Counter Use _____
(Per 21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Ophthalmic Ear,
Nose and Throat Devices
510(k) Number K072088

Page 1 of 1

000019