



May 31, 2016

AngioScore, Inc.  
c/o Ms. Melinda Swanson  
5055 Brandlin Court  
Fremont, CA 94538

Re: K072225

Trade/Device Name: AngioSculpt® PTA Scoring Balloon Catheter  
Regulation Number: 21 CFR 870.1250  
Regulation Name: Percutaneous catheter  
Regulatory Class: Class II (two)  
Product Code: PNO  
Dated: September 11, 2007  
Received: September 12, 2007

Dear Ms. Swanson:

This letter corrects our substantially equivalent letter of October 12, 2007.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Misti L. Malone -S**

for Bram D. Zuckerman, M.D.  
Director  
Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure



OCT 12 2007

K072225  
p. 1 of 1

**510(k) Summary for the Modified AngioSculpt Scoring Balloon Catheter**

**1. Submitter's Name / Contact Person**

Submitter: AngioScore, Inc.  
5055 Brandin Court  
Fremont, CA 94538

Contact Person: Melinda Swanson  
Director, Clinical and Regulatory Affairs  
Phone: 510-933-7910  
Fax: 510-933-7994

Summary Preparation Date: September 11, 2007

**2. General Information**

Trade Name: AngioSculpt® PTA Scoring Balloon Catheter  
Common / Usual Name: Angioplasty catheter  
Classification Name: Catheter, angioplasty, peripheral, transluminal (DQY and I.IT)  
Predicate Devices: AngioSculpt® Scoring Balloon Catheter (K050629)  
Cordis Savvy® PTA Dilatation Catheter (K971010)

**3. Intended Use / Indications**

The modified AngioSculpt PTA Scoring Balloon Catheter is intended for balloon dilatation of lesions in infrapopliteal arteries. Not for use in the coronary or neuro-vasculature.

**4. Device Description**

The modified AngioSculpt catheter is a standard two-lumen catheter with a scoring balloon near the distal tip. The distal end of the catheter has a conventional nylon-blend balloon with a nitinol scoring element that wraps around the balloon. The scoring element creates focal concentrations of dilating force which minimizes balloon slippage and assists with luminal expansion of stenotic arteries. The balloon has radiopaque markers to aid in positioning the balloon in the stenosis, and is designed to provide an expandable segment of known diameter and length at a specific pressure.

**5. Substantial Equivalence Comparison**

There is no difference between the modified AngioSculpt catheter intended use and that of the predicate devices. The modified AngioSculpt catheter shares the same fundamental design, scientific technology (operating principle), functional performance, and similar materials as the original AngioSculpt catheter. The modified AngioSculpt catheter has similar dimensions and size configurations as the Savvy catheter. Design verification and validation testing demonstrated adequate device performance and confirmed that no new questions of safety or effectiveness for peripheral balloon angioplasty devices were raised.

The modified AngioSculpt PTA Scoring Balloon Catheter is therefore, substantially equivalent to the predicate devices.