

**SMITHS MEDICAL MD, INC.  
510(k) Premarket Notification**

**510(k) Summary**

**I. Applicant (Sponsor) Name and Address**

Smiths Medical MD, Inc.  
1265 Grey Fox Road  
St. Paul, MN 55112  
Establishment Reg. No.: 2183502

**II. Contact Name and Phone**

Brian L. Haugstad  
Senior Regulatory Affairs Associate  
Company Phone: (651) 628-7513  
Company Fax: (651) 628-7457

DEC 14 2007

**III. Device Trade/Proprietary Name**

Catheter/Port

POWER P.A.C. Implantable Venous Access Systems

Needle

GRIPPER PLUS® POWER P.A.C. Safety Huber Needle

**IV. Device Classification/Common Name/Panel**

Catheter/Port

21 CFR Reference: §880.5965  
21 CFR Common Name: Implanted Infusion Port  
Classification: Class II  
Product Code: LJT  
Panel: General Hospital

Needle

21 CFR Reference: §880.5440  
21 CFR Common Name: Huber Needle Intravascular Administration Set  
Classification: Class II  
Product Code: FPA  
Review Panel: General Hospital and Personal Use

1072657  
283

**SMITHS MEDICAL MD, INC.  
510(k) Premarket Notification**

**510(k) Summary**

**V. Identification of Predicate Device**

Smiths Medical MD, Inc. believes the POWER P.A.C. Implantable Venous Access Systems and GRIPPER PLUS® POWER P.A.C. Safety Huber Needles are substantially equivalent to the following devices and thus may market the devices under these equivalencies.

Catheter/Port

Device Name:

1. Smiths Medical MD, Inc.: PORT-A-CATH® and PORT-A-CATH® II POWER P.A.C. Implantable Venous Access Systems and Gripper Plus® POWER P.A.C. Needle
2. Smiths Medical MD, Inc.: P.A.S. Port® T2 Implantable Venous Access System
3. Smiths Medical MD, Inc.: PORT-A-CATH® II 7 French, Dual Lumen Low Profile™ Implantable Venous Access System
4. Medcomp®: Power Injectable Implantable Infusion Port

Needle

Device Name:

1. Smiths Medical MD, Inc. PORT-A-CATH® and PORT-A-CATH® II POWER P.A.C. Implantable Venous Access Systems and Gripper Plus® POWER P.A.C. Needle

**VI. Device Description**

POWER P.A.C. Implantable Venous Access Systems

The POWER P.A.C. implantable venous access systems are similar to the current PORT-A-CATH® and PORT-A-CATH® II implantable venous access systems and Power Injectable Implantable Infusion Port. They both are totally implantable venous access systems designed to permit repeated access to the venous system for the parenteral delivery of medications, fluids, nutritional solutions, and for the sampling of venous blood. However, when used with a power indicated huber needle, the POWER P.A.C. implantable venous access systems are indicated for power injection of contrast media. The POWER P.A.C. implantable venous access systems can be placed in the chest or arm.

The POWER P.A.C. implantable venous access systems are supplied sterile and non-pyrogenic. All POWER P.A.C. implantable venous access systems are designed and intended for single patient use only. A system consists of a portal with one or two self-sealing septa and a single or dual lumen catheter and is accessible by percutaneous puncture with a non-coring needle.

GRIPPER PLUS® POWER P.A.C. Needle

The GRIPPER PLUS® POWER P.A.C. Needle is similar to the current GRIPPER PLUS® needle, with the addition of an added indication. Both are supplied sterile and non-pyrogenic and are intended for the administration into or withdrawal of fluids from implanted ports. In addition, they both have a passive needle stick protection feature that is designed to help protect against exposure to bloodborne pathogens caused by accidental needlestick injuries. It does not protect against other routes of bloodborne pathogen transmission.

The GRIPPER PLUS® POWER P.A.C. Needle is indicated for power injection of contrast media.

**SMITHS MEDICAL MD, INC.  
510(k) Premarket Notification**

**510(k) Summary**

**VII. Intended Use of the Device**

The intended use of each device has not changed from that of the predicate.

POWER P.A.C. Implantable Venous Access Systems

The POWER P.A.C. implantable venous access systems are implantable vascular access devices designed to provide long term repeated access to the vascular system.

GRIPPER PLUS® POWER P.A.C. Needle

The GRIPPER PLUS® POWER P.A.C. Safety Huber Needle is designed for the administration into or withdrawal of fluids from implanted ports.

**VIII. Indications for Use**

POWER P.A.C. Implantable Venous Access Systems

Smiths Medical MD, Inc. POWER P.A.C. Implantable Venous Access Systems are indicated when patient therapy requires repeated vascular access for injection or infusion therapy and/or venous blood sampling. When used with a GRIPPER PLUS® POWER P.A.C. Safety Huber Needle or other power injectable huber needle, the Smiths Medical MD, Inc. POWER P.A.C. implantable venous access systems are indicated for power injection of contrast media. For power injection of contrast media, the maximum recommended infusion rate is 5 ml/sec.

GRIPPER PLUS® POWER P.A.C. Needle

The GRIPPER PLUS® POWER P.A.C. Safety Huber Needle is indicated for the administration into or withdrawal of fluids from implanted ports. It is designed to help protect against exposure to bloodborne pathogens caused by accidental needlestick injuries. When used with power injectable implantable venous access ports, the GRIPPER PLUS® POWER P.A.C. Safety Huber Needle is indicated for power injection of contrast media.

**IX. Summary of Studies**

**A. Functional Testing**

The POWER P.A.C. Implantable Venous Access Systems and GRIPPER PLUS® POWER P.A.C. Needle met all established acceptance criteria for performance testing and design verification testing. Packaging and Sterilization systems are unchanged from predicate devices.

**B. Clinical Studies**

Clinical studies for the POWER P.A.C. Implantable Venous Access Systems and GRIPPER PLUS® POWER P.A.C. Safety Huber Needle were deemed not necessary due to their similarity in materials, design and function to current Smiths Medical MD, INC. devices and other commercially available systems.

**C. Conclusions Drawn from the Studies**

Based upon the information provided; the POWER P.A.C. Implantable Venous Access Systems and GRIPPER PLUS® POWER P.A.C. Safety Huber Needle met all acceptance criteria for performance testing and design verification testing. Therefore, these products are considered acceptable for human use.



DEC 14 2007

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Mr. Brian L. Haugstad  
Senior Regulatory Affairs Associate  
Smiths Medical MD, Incorporated  
1265 Grey Fox Road  
St. Paul, Minnesota 55112

Re: K072657

Trade/Device Name: GRIPPER PLUS™ POWER P.A.C. Safety Huber Needle and  
POWER P.A.C. Implantable Venous Access Systems

Regulation Number: 21 CFR 880.5965

Regulation Name: Subcutaneous, Implanted, Intravascular Infusion Port and Catheter

Regulatory Class: II

Product Code: LJT, FPA

Dated: September 18, 2007

Received: September 20, 2007

Dear Mr. Haugstad:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

  
Chiu Lin, Ph.D.

Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

SMITHS MEDICAL MD, INC.  
510(k) Premarket Notification

152

Indications for Use

510(k) Number: K072657

Device Name: POWER P.A.C. Implantable Venous Access Systems

Smiths Medical MD, Inc. POWER P.A.C. Implantable Venous Access Systems are indicated when patient therapy requires repeated vascular access for injection or infusion therapy and/or venous blood sampling.

When used with a GRIPPER PLUS® POWER P.A.C. Safety Huber Needle or other power injectable huber needle, the Smiths Medical MD, Inc. POWER P.A.C. implantable venous access systems are indicated for power injection of contrast media. For power injection of contrast media, the maximum recommended infusion rate is 5 ml/sec.

Prescription Use  OR Over-The Counter Use  Per 21 CFR 801.109)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

*Antonia...*  
(Division Sign-Off)  
Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

K072657  
~~510(k) Number~~ of CDRH, Office of Device Evaluation (ODE)  
Concurrence of

SMITHS MEDICAL MD, INC.  
510(k) Premarket Notification

Indications for Use

2.8

510(k) Number: K070657

Device Name: GRIPPER PLUS™ POWER P.A.C. Safety Huber Needle

The GRIPPER PLUS® POWER P.A.C. Safety Huber Needle is indicated for the administration into or withdrawal of fluids from implanted ports. It is designed to help protect against exposure to bloodborne pathogens caused by accidental needlestick injuries.

When used with power injectable implantable venous access ports, the GRIPPER PLUS® POWER P.A.C. Safety Huber Needle is indicated for power injection of contrast media.

Prescription Use  X  OR Over-The Counter Use \_\_\_\_\_ Per 21 CFR 801.109)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

\_\_\_\_\_  
(Division Sign-Off)  
Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

510(k) Number: Concurrence of CDRH, Office of Device Evaluation (ODE)