

**510(k) Summary of Safety and Effectiveness**

**Carematix™ Wellness System**

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR 807.92.

**Submitter**

Carematix™, Inc.  
120 S. Riverside Plaza, Suite 2100  
Chicago, IL 60606

**Contact Person**

Sukhwant Khanuja, Ph. D.  
Chief Executive Officer  
Carematix™, Inc.  
120 S. Riverside Plaza, Suite 2100  
Chicago, IL 60606

JAN 11 2008

Phone: (312) 627-9300  
e-mail: skhanuja@carematix.com

**Date Prepared:**

October 23, 2007

**Name of Device**

Carematix™ Wellness System

**Classification Names**

Radiofrequency Physiological Signal Transmitters and Receivers

**Device Classification**

Regulatory Class: Class II  
Product Code: DRG  
Classification Panel: Cardiovascular Device Panel  
Regulation Number: 21 CFR 870.2910

**Predicate Devices**

K040966      Carematix™ Wellness System      Carematix™, Inc.

**Description of Device**

The Carematix™ Wellness System is intended to gather and transmit patient data from the home site or remote location to caregivers at a clinical facility where it provides supplemental information for the care of the patient.

The **Carematix™ Wellness System** consists of wireless radiofrequency transmitter adapters, a communication hub or receiving station, and an internet server:

- The radiofrequency transmitter adapter is connected, either wired directly or to a serial port, of monitoring devices currently in distribution having capability to monitor patient parameters for blood pressure, pulse rate, blood sugar, blood oxygen saturation, PT/INR, and FEVR/PEF.
- The communications hub, or receiving station, collects and stores data transmitted from each of the radiofrequency transmitter adapters and transmits patient data to the internet server at specified intervals.

- The internet server receives the patient data from the home setting or remote location where it is made available to the caregiver to track, graph, trend, note variances, set alert criteria, and receive alerts when parameters are outside the criteria set.

Page 1 of 2

### **Indications For Use**

The Carematix™ Wellness System is a physiological monitoring system. The system collects, accumulates and transmits patient vital signs and other physiological data from a patient who may be remote from the healthcare practitioner to the practitioner. It is intended for patient home use for the following and can record physiological information such as:

Non-invasive blood pressure measurement;  
Pulse rate measurement;  
Blood glucose level using a Glucometer;  
Blood hemoglobin oxygen saturation (%SpO<sub>2</sub>) using a digital Pulse Oximeter;  
Prothrombin Time (PT) and International Normalized Ratio (INR) measurement using an in-home coagulation measurement system;  
Peak Expiratory Flow Rate (PEFR) and Forced Expiratory Volume (FEV) measurements using an electronic peak flow meter;  
Patient weight using a stand-on electronic scale

The results of these measurements are transmitted to a computer monitoring station in a clinical setting via common telephone lines from the patient's home setting.

### **Nonclinical Performance**

The Carematix™ Wellness System was tested and passed all required electrical and mechanical testing.

### **Clinical Performance**

The Carematix™ Wellness System performance was tested with clinical data and the results met the acceptable criteria.

### **Conclusion**

The Carematix™ Wellness System is substantially equivalent to the following 510(k) cleared devices:

Carematix™ Wellness System cleared under K040966 on June 2, 2004



JAN 11 2008

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Carematix Inc.  
c/o Sukhwant Khanuja, Ph.D.  
Chief Executive Officer  
120 S. Riverside Plaza, Suite 2100  
Chicago, IL 60606

Re: K073038  
Trade/Device Name: Carematix™ Wellness System Model CWS-5000.1-B  
Regulation Number: 21 CFR 870.2910  
Regulation Name: Radiofrequency Physiological Signal Transmitter and Receiver  
Regulatory Class: Class II (two)  
Product Code: DRG  
Dated: December 11, 2007  
Received: December 13, 2007

Dear Dr. Khanuja:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at 240-276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at 240-276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Bram D. Zuckerman, M.D.  
Director  
Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

