FF7 - 5 : 4-1

K073282

## 510(k) Summary 21 CFR 807.92

Submitter's Name & Address

Manufacturer:

BioHorizons Implant Systems, Inc.

2300 Riverchase Center Birmingham, AL 35244 Phone (205) 967-7880 Fax (205) 870-0304

Official contact: Date prepared:

Winston Greer, Vice-President, QA & RA

November 20, 2007

Name of the Device

Trade Name:

BioHorizons Single-stage Implant System

Common or Usual Name: Screw-type dental implant Classification Name: Classification Number:

Endosseous dental implant Class II (21 CFR 872.3640)

**Predicate Devices** 

1. BioHorizons Single-stage Implant System, documented under 510(k) number K053152, concurrence date of December 2, 2005.

2. BioHorizons Tapered Internal Implant System, documented under 510(k) number K071638, concurrence date of October 10, 2007.

Device Description

Bio-forizons Single-stage Implants are machined titanium, screw-form endosseous denial implants supplied in 3.5mm, 4.0mm, 5.0mm, 6.0mm diameters across lengths of 7mm, 9mm, 10.5mm, 12mm and 15mm. Implant material is titanium alloy as specified in ASTM F136 Standard Specification for Wrought Titanium-6Aluminum-4Vanadium ELI (Extra Low Interstitial) Alloy (UNS R56401) for Surgical Implant Applications.

The devices are further processed by roughening the surface with tricalcium phosphate plast media, or by applying hydroxylapatite coating conforming to ASTM F1185 indard Specification for Composition of Ceramic Hydroxylapatite for Surgical mulants, to promote implant fixation. The product is packaged using materials known fille industry to be appropriate for medical device packaging and is provided with a minimum sterility assurance level of 10<sup>-6</sup>, validated in compliance to ANSI / AAMI / ISO 1137-1 Sterilization of healthcare products -- Radiation -- Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices.

BioHorizons Single-stage Implant System includes a series of implant catalog item mbers with Laser-Lok® technology applied to the implant collar to provide additional leatment options for the dental implant clinician.

Intended Use

BioHorizons Single-stage Implants are intended for use in the mandible or maxilla as an artificial root structure for single tooth replacement or for fixed bridgework and dental retention.

BioHorizons Single-stage Implants may be restored immediately

1) with a temporary prosthesis that is not in functional occlusion, or

2) when splinted together for multiple tooth replacement, or when stabilized with an overdenture supported by multiple implants.

Technological Characteristics

The fundamental scientific technology of the BioHorizons Single-stage Implant System of endosseous dental implant devices subject to this 510(k) is substantially equivalent to the referenced predicate devices. The addition to the indications for use for conditional immediate restoration is substantially equivalent to the predicate BioHorizons Tapered Internal Implant System (K071638), and immediate restoration is an accepted and prevalent clinical practice of demonstrated safety and efficacy. Laser-Lok is a surface feature in which patterns of micro-machined grooves are applied to the collar surface of a dental implant to (1) inhibit epithelial cell downgrowth around the implant, and (2) attach and retain crestal bone adjacent to the implant. All materials, suppliers, processing, packaging and sterilization methods remain the same as for the predicate BioHorizons Single-stage (K053152) and Tapered Internal endosseous implants, and the Laser-Lok feature is substantially equivalent as that cleared for the Tapered Internal Implant System. The BioHorizons Single-stage Implant System which is the subject of this 510(k) is substantially equivalent to all features of the predicate BioHorizons Single-stage and Tapered Internal implant devices which could affect safety or effectiveness because of the similarities in design, materials and intended use.

## DEPARTMENT OF HEALTH & HUMAN SERVICES



FEB 1 5 2008

Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

Mr. Winston Greer Vice-President, Quality Assurance & Regulatory Affairs BioHorizons Implant Systems, Incorporated 2300 Riverchase Center Birmingham, Alabama 35244

Re: K073282

Trade/Device Name: BioHorizons Single-stage Implant System

Regulation Number: 872.3640

Regulation Name: Endosseous Dental Implant

Regulatory Class: II Product Code: DZE

Dated: November 20, 2007 Received: November 21, 2007

## Dear Mr. Greer:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal</u> Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <a href="http://www.fda.gov/cdrh/industry/support/index.html">http://www.fda.gov/cdrh/industry/support/index.html</a>.

Sincerely yours,

Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

sutte y. Michae Omo.

Enclosure

Page 1 of 1
510(k) Number:
Device Name: <u>BioHorizons Single-stage Implant System</u>
Indications for Use:
BioHorizons Single-stage Implants are intended for use in the mandible or maxilla as an artificial root structure for single tooth replacement or for fixed bridgework and dental retention.
BigHorizons Single-stage Implants may be restored immediately 1) with a temporary prosthesis that is not in functional occlusion, or 2) when splinted together for multiple tooth replacement, or when stabilized with an overdenture supported by multiple implants.
Concurrence of CDRH, Office of Device Evaluation (ODE)
Division Sign-Off)
ivision of Anesthesiology, General Hospital
fection Control, Dental Devices
10(k) Number: KO 78282
Prescription Use X OR Over-the-Counter Use (per 21 CFR 801.109)