

K080479

510(k) Summary

Trade Name: Dental Ceramic **MAY 21 2008**

Sponsor: DMG USA, Inc.
23 Frank Mossberg Drive
Attleboro, MA 02703
Registration # not yet assigned
Owner/Operator No. 9005969

Device Generic Name: Dental Ceramic

Classification: According to Section 513 of the Federal Food, Drug, and Cosmetic Act, the device classification is Class II.

Product Description:

The present Dental Ceramic blocks are grindable leucite reinforced glass ceramic blocks for use with CEREC[®] and inLab[®] devices marketed by Sirona, Bensheim, Germany. The blocks are compatible with all CEREC models up to CEREC 3 and inLab but are not compatible with the CEREC model MC XL and Inlab model MC XL.

Dental Ceramic is indicated for

- preparation of inlays
- preparation of onlays
- preparation of veneers
- preparation of crowns

Predicate Devices:

The Dental Ceramic material is substantially equivalent to the following currently marketed dental restorative material:

Dental Ceramic:

| Product Name | Predicates |
|--------------|-----------------------------|
| PROCAD | K980986 (Ivoclar USA, inc.) |

Safety and Performance:

This submission is an Abbreviated 510(k) as described in FDA's guidance document entitled "The New 510(k) Paradigm - Alternate Approaches to Demonstrating Substantial Equivalence in Premarket Notifications." Data provided in this submission to establish substantial equivalence includes chemical composition and certification of compliance with relevant consensus standards.

Conclusion:

Based on the indications for use, technological characteristics, and comparison to the predicate devices, the Dental Ceramic has been shown to be safe and effective for its intended use.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

MAY 21 2008

DMG USA, Incorporated
C/O Ms. Pamela Papineau
Consultant
Delphi Medical Device Consulting, Incorporated
5 Whitcomb Avenue
Ayer, Massachusetts 01432

Re: K080479
Trade/Device Name: Dental Ceramic
Regulation Number: Porcelain Powder for Clinical use
Regulation Name: 872.6660
Regulatory Class: II
Product Code: EIH
Dated: April 29, 2008
Received: April 2, 2008

Dear Ms. Papineau:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Chiu Lin, Ph.D.
Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K080479

Device Name: Dental Ceramic

Product Indications for Use:

The Dental Ceramic is indicated for use with the:


- preparation of inlays
- preparation of onlays
- preparation of veneers
- preparation of crowns

The Dental Ceramic is indicated for use with CEREC devices marketed by Sirona, Bensheim, Germany. The blocks are compatible with all CEREC devices up to CEREC 3 and inLab but not compatible with the CEREC model MC XL and Inlab model MC XL.

Prescription Use X OR Over-the -Counter Use _____
(Per 21 CFR 801 Subpart D) (Per 21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K080479