

7. 510(k) Summary according to 807.92(c)

Contact:	Kevin Gemas President 866-822-7800	Titan Spine, LLC Mequon Technology Center 10520 Baehr Rd., Suite A Mequon, WI 53092
Trade Name:	Endoskeleton TA® Interbody Fusion Device	
Product Class:	Class II	
Classification:	21 CFR §888.3080 Orthosis, intervertebral fusion	
Product Codes:	MAX	
Panel Code:	87	

Indications for Use:

The Endoskeleton TA® Interbody Fusion Device is indicated for use in skeletally mature patients with Degenerative Disc Disease (DDD) at one or two contiguous levels from L2-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by patient history and radiographic studies. Patients should have received 6 months of non-operative treatment prior to treatment with the devices. The device may be used with supplemental fixation. These DDD patients may also have up to Grade I spondylolisthesis or retrolisthesis at the involved level(s). It is indicated to be use with autograft bone.

Device Description:

The Endoskeleton TA® Interbody Fusion Device is comprised of a variety of implant sizes to accommodate various patient anatomy and pathology, and associated instrumentation. All implantable components are manufactured from medical grade titanium alloy (Ti6Al4V-ELI).

Predicate Device(s):

The Endoskeleton TA® Interbody Fusion Device was shown to be substantially equivalent to previously cleared devices and has the same indications for use, design, function, and materials used. The four predicate devices include the BAK Interbody Fusion Device (Spine-Tech, P950002), Inter Fix Threaded Fusion Device (Sofamor Danek, P970015) and the Ray Threaded Fusion Cage (Surgical Dynamics, P950019) and the Titan Spine Endoskeleton TA® VBR (K032812).

Performance Testing:

The pre-clinical testing performed indicated that the Endoskeleton TA® Interbody Fusion Device is adequate for the intended use.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Titan Spine, LLC
% Silver Pine Consulting
Mr. Richard Jansen
President
13540 Guild Avenue
Apple Valley, Minnesota 55124

JUN 17 2008

Re: K080615
Trade/Device Name: Endoskeleton TA[®] Interbody Fusion Device
Regulation Number: 21 CFR §888.3080
Regulation Name: Intervertebral body fusion device.
Regulatory Class: Class II
Product Code: MAX
Dated: February 26, 2008
Received: April 22, 2008

Dear Mr. Jansen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 - Mr. Richard Jansen

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at (240) 276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at (240) 276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at toll-free number (800) 638-2041 or (240) 276-3150 or the Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

4. Statement of Indications for Use

510(k) Number (if known): K080615

Indications for Use:

The Endoskeleton TA® Interbody Fusion Device is indicated for use in skeletally mature patients with Degenerative Disc Disease (DDD) at one or two contiguous levels from L2-S1. DDD is defined as discogenic back pain with degeneration of the disc confirmed by patient history and radiographic studies. Patients should have received 6 months of non-operative treatment prior to treatment with the devices. The device may be used with supplemental fixation. These DDD patients may also have up to Grade I spondylolisthesis or retrolisthesis at the involved level(s). It is indicated to be use with autograft bone.

Prescription Use AND/OR Over-The-Counter Use
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Neil R. Ogle for man
(Division Sign-Off)

**Division of General, Restorative,
and Neurological Devices**

510(k) Number K080615

Concurrence of CDRH, Office of Device Evaluation (ODE)