

**Summary of Safety and Effectiveness
Omega™ 3 Trochanteric Stabilizing Plate**

MAY 29 2008

Proprietary Name: Omega™ 3 Trochanteric Stabilizing Plate

Common Name: Compression Screw System

Classification Name and Reference: Single/multiple component metallic bone fixation appliances and accessories, 21 CFR §888.3030

Device Product Code: 87 KTT

For Information Contact: Andrea Dwyer, Regulatory Affairs Associate
Howmedica Osteonics Corp.
325 Corporate Drive
Mahwah, NJ 07430
Phone: (201) 831-6250
Fax: (201) 831-6038

Date Summary Prepared: May 5, 2008

Description:

The Omega™ 3 Trochanteric Stabilizing Plate is a compression screw system designed to treat various types of fractures of the proximal femur.

Intended Use:

The modifications do not alter the intended use of the predicate system as cleared in the referenced premarket notifications. The subject and predicate devices are available sterile and non-sterile intended for use in the temporary stabilization of fractures of the proximal and distal femur. The indications for use for the Omega™ 3 System are provided below.

Indications for Use:

The Omega™ Plus, Omega™ 2, and Omega™ 3 Systems are intended for use in the temporary stabilization of types of fractures of the proximal and distal femur. The subject devices are indicated for fixation of proximal and distal femoral fractures including:

- Intracapsular and basal neck fractures including transcervical and subcapital fractures
- Intertrochanteric fractures

- Subtrochanteric fractures
- Supracondylar fractures
- Intracondylar fractures
- Osteotomies for patients with diseases or deformities of the hip
- Hip arthrodesis

Substantial Equivalence:

The subject Omega™ 3 Trochanteric Stabilizing Plate shares the same intended use, and basic design concepts as that of the currently available Omega™ 3 System. Mechanical testing demonstrated comparable mechanical properties to the predicate components and is substantially equivalent to these devices.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

MAY 29 2008

Howmedica Osteonics Corp.
% Ms. Andrea Dwyer
Regulatory Affairs Associate
325 Corporate Drive
Mahwah, New Jersey 07430

Re: K081278
Trade/Device Name: Omega™ 3 System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/multiple component metallic bone fixation appliances and accessories
Regulatory Class: II
Product Code: KTT
Dated: May 5, 2008
Received: May 6, 2008

Dear Ms. Dwyer:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 - Ms. Andrea Dwyer

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at (240) 276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at (240) 276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at toll-free number (800) 638-2041 or (240) 276-3150 or the Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K081278

Device Name: Omega™ 3 System

Indications for Use:

The Omega™ Plus, Omega™ 2, and Omega™ 3 Systems are intended for use in the temporary stabilization of types of fractures of the proximal and distal femur. The subject devices are indicated for fixation of proximal and distal femoral fractures including:

- Intracapsular and basal neck fractures including transcervical and subcapital fractures
- Intertrochanteric fractures
- Subtrochanteric fractures
- Supracondylar fractures
- Intracondylar fractures
- Osteotomies for patients with diseases or deformities of the hip
- Hip arthrodesis

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Neil K. Ogden for MKA
(Division Sign-Off)

**Division of General, Restorative,
and Neurological Devices**

510(k) Number K081278