

SEP 25 2008

## 510(k) Summary

<b>SUBMITTER:</b>	HAMILTON MEDICAL AG Via Crusch 8 7402 Bonaduz, Switzerland
<b>CONTACT PERSON:</b>	Joerg Schneider Phone: +41 81 660 6479 Fax: +41 81 660 60 20 e-mail: jschneider@hamilton-medical.ch
<b>PREPARATION DATE:</b>	May 21, 2008
<b>TRADE NAME:</b>	HAMILTON-G5
<b>COMMON NAME:</b>	Continuous Ventilator
<b>CLASSIFICATION NAME:</b>	Ventilator, Continuous, Facility Use (21 CFR 868.5895, Product Code: CBK)
<b>LEGALLY MARKETED DEVICES TO WHICH EQUIVALENCE IS BEING CLAIMED:</b>	HAMILTON-G5 510(k) Number: K070513
	MAQUET Servo-i 510(k) Number: K041223
	Viasys AVEA ventilator 510(k) Number: K062093
	Datex-Ohmeda Engström Carestation 510(k) Number: K062710
	Dräger EvitaXL with NeoFlow Option 510(k) Numbers: K051263, K983219



## **DEVICE DESCRIPTION**

The HAMILTON-G5 is an electronically controlled pneumatic intensive care ventilator ventilation system.

It uses oxygen and air or (optionally) heliox to ventilate adults, pediatrics, and optionally infants and neonates. It is powered by ac with battery backup to protect against power failure or unstable power and to facilitate intrahospital transport.

The HAMILTON-G5's pneumatics deliver gas, and its electrical systems control pneumatics, monitor alarms, and distribute power.

The user interface consists of a LCD-display with touch screen, keys, and a press-and-turn knob.

The CO<sub>2</sub> option allows continuous mainstream monitoring of carbon dioxide.

## **INTENDED USE**

The HAMILTON-G5 ventilator is designed for intensive care ventilation of adult and pediatric patients, and optionally infant and neonatal patients.

The device is intended for use in the hospital and institutional environment where healthcare professionals provide patient care.

The HAMILTON-G5 ventilator is intended for use by properly trained personnel under the direct supervision of a licensed physician.

The HAMILTON-G5 ventilator may be used for transport within a hospital or hospital-type facility provided compressed gas is supplied. The device is not to be used in the presence of flammable anesthetic agents or other ignition sources. The ventilator is not to be used in an environment with magnetic resonance imaging (MRI) equipment. The device is not intended for transportation outside the hospital or for use in the home environment.

In the USA, federal law restricts this device to sale by or on the order of a physician.

## **SUMMARY OF THE TECHNOLOGY AND PERFORMANCE SPECIFICATIONS COMPARISON WITH THE PREDICATED DEVICES**

The indication statements for the HAMILTON-G5 ventilator are comparable to those for the predicate devices.

Technological characteristics and performance specifications of the HAMILTON-G5 ventilator are substantially equivalent to those of the predicate devices.

The Carbon dioxide gas analyzer is considered to be substantial equivalent to the CO<sub>2</sub> gas analyzer of the currently marketed predicate device Servo-i.

The heliox option is comparable with the substantial equivalent option of the predicate device AVEA ventilator.

The neonate ventilation characteristics are substantially equivalent to those of the predicate devices Servo-i, Engström Carestation, and Evita XL.

HAMILTON MEDICAL has demonstrated the HAMILTON-G5 ventilator including the new options to be safe and effective.

The HAMILTON-G5 is considered to be substantial equivalent to currently marketed predicate devices which have been previously cleared by FDA.



### **NON-CLINICAL PERFORMANCE TESTS**

Safety testing of the HAMILTON-G5 with the new options was conducted according to IEC60601-1, IEC60601-1-2, IEC 60601-2-12 and other applicable standards. The test results show that the device is safe and effective for its intended use.

The ventilator was further subject to wave-form performance testing as described in the standard ASTM F1100-90. The data provided from these tests, were shown to be substantially equivalent to a legally marketed device.

The results of the software verification and validation testing demonstrate that all specified requirements have been implemented correctly and completely.

### **CONCLUSION**

The results of verification, validation, and testing activities demonstrate that the HAMILTON-G5 ventilator including the new options is as safe, as effective, and performs as well as or better than the legally marketed devices identified above.



SEP 25 2008

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Mr. Joerg Schneider  
Quality Engineer, R&D  
Hamilton Medical AG  
Via Crusch 8  
CH-7402 Bonaduz  
SWITZERLAND

Re: K081521  
Trade/Device Name: HAMILTON-G5  
Regulation Number: 21 CFR 868.5895  
Regulation Name: Continuous Ventilator  
Regulatory Class: II  
Product Code: CBK  
Dated: September 12, 2008  
Received: September 17, 2008

Dear Mr. Schneider:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at 240-276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at 240-276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Chiu Lin, Ph.D.  
Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

# INDICATIONS FOR USE STATEMENT

510(k) Number: \_\_\_\_\_

Device Name: HAMILTON-G5

Indication for Use: The HAMILTON-G5 ventilator is intended to provide positive pressure ventilatory support to adult and pediatric patients, and optionally to infant and neonatal patients.  
The device is intended for use in the hospital and institutional environment where healthcare professionals provide patient care, including use as a patient bedside for intra-facility transport, provided compressed gas is supplied.  
The device is not intended for transportation outside the hospital or for use in the home environment.


Prescription Use   X    
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)  
Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

510(k) Number:   K081521