

OCT 07 2008

K02608

Radius Spinal System Line Extension – Vitallium Rod

Special 510(k) Premarket Notification

**Special 510(k) Summary of Safety and Effectiveness:
Line Extension to the Radius[®] Spinal System**

Proprietary Name: Radius[®] Spinal System – Vitallium[®] Rod

Common Name: Spinal Fixation Appliances

Proposed Regulatory Class: Class III

Orthosis, Spinal Pedicle Fixation, For Degenerative
Disc Disease
21 CFR 888.3070

Device Product Code: NKB, KWP, KWQ, MNH, MNI

For Information contact: Curtis Truesdale
Regulatory Affairs Project Manager
2 Pearl Court
Allendale, NJ 07401
Telephone: (201) 760-8296
Fax: (201) 760-8496
Email: Curtis.Truesdale@Stryker.com

Date Summary Prepared: September 26, 2008

Predicate Devices

- Stryker Spine Radius[®] Spinal System, K070631, K062270
- Stryker Spine Xia[®] Spinal System, K060979.
- Medtronic Sofamor Danek CD Horizon Spinal System, K020709

Description of Device Modification This 510(k) is intended to introduce an extension to the existing Radius[®] Spinal System. The proposed line extension includes the addition of a Vitallium rod.

Intended Use The Radius[®] Spinal System is intended for anterior/anterolateral and posterior, noncervical pedicle and non-pedicle fixation for the following indications: degenerative disc disease (DDD) (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis, and/or lordosis); tumor; pseudoarthrosis; and failed previous fusion.

The Radius[®] Spinal System can also be linked to the Xia[®] Titanium Spinal System via the Ø5.5 mm to Ø6.0 mm Radius[®] rod-to-rod connector.

Summary of the Technological Characteristics Documentation is provided which demonstrates the new components of the Stryker Spine Radius[®] Spinal System to be substantially equivalent to the predicate devices in terms of material, design, and indications for use. Engineering analysis and testing to demonstrate compliance with FDA's Guidance for Spinal System 510(k)'s May 3, 2004 was completed for the Stryker Spine Radius[®] Spinal System, including the subject components.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

OCT 07 2008

Stryker Spine
% Mr. Curtis Truesdale
Regulatory Affairs Project Manager
2 Pearl Court
Allendale, New Jersey 07401

Re: K082608

Trade/Device Name: Radius[®] Spinal System Line Extension- Vitallium[®] Rod
Regulation Number: 888.3070
Regulation Name: Pedicle Screw Spinal System
Regulatory Class: III
Product Code: NKB, MNI, MNH, KWP, KWQ
Dated: August 11, 2008
Received: September 8, 2008

Dear Mr. Truesdale:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2- Mr. Curtis Truesdale

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at (240) 276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at (240) 276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkerson
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

