

K082619

NOV - 7 2008

Page 1 of 2

510(K) SUMMARY:

510(k) Summary of Safety and Effectiveness:

SUBMITTER: Surgical Devices, a global business unit of Tyco  
Healthcare Group LP (d/b/a Covidien)  
60 Middletown Avenue  
North Haven, CT 06473

CONTACT PERSON: Robert Zott  
Program Director, Regulatory Affairs  
Phone: (203) 492-6013  
Fax: (203) 492-5029

DATE PREPARED: September 8, 2008

TRADE/PROPRIETARY NAME: SILS™ Port

COMMON/USUAL NAME: Surgical Trocar

CLASSIFICATION NAME: Endoscope and/or accessories

PREDICATE DEVICE(S): K981941: Dexide™ Multiport Cannula Reducer  
and Accessories  
K012539: Versa Step™ Reposable System  
K954108: Modified Versaport™ Trocar  
K072814: Convenience Kit for "Single-Incision  
Laparoscopic Surgery and other advanced  
laparoscopic procedures."  
K945457: Auto Suture™ Modified Grip  
K073170: R-Port II Laparoscopic Access Device

DEVICE DESCRIPTION: Three laparoscopic trocars and an insufflation  
tube bound by a flexible port.

INTENDED USE:

For multiple instrument or camera access to the abdominal cavity through a single incision for performing minimally invasive laparoscopic procedures

TECHNOLOGICAL CHARACTERISTICS:

The SILS™ Port provides the ability to use three conventional laparoscopic trocars and an insufflation port through a single incision while providing the ability to maintain pneumoperitoneum.

MATERIALS:

All components of the SILS™ Port are comprised of materials that have been evaluated in accordance with ISO 10993-1: 2003, Biological Evaluation of medical devices -- Part 1: Evaluation and Testing.

PERFORMANCE DATA:

In-vitro and in-vivo testing has been performed in support of the intended use of this device.



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Covidien  
% Mr. Robert Zott  
Regulatory Affairs Program Director  
60 Middletown Avenue  
North Haven, Connecticut 06473

NOV - 7 2008

Re: K082619  
Trade/Device Name: SILS™ Port  
Regulation Number: 21 CFR 876.1500  
Regulation Name: Endoscope and accessories  
Regulatory Class: II  
Product Code: GCJ  
Dated: October 27, 2008  
Received: October 28, 2008

Dear Mr. Zott:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at (240) 276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at (240) 276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkerson  
Director  
Division of General, Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

K082619

## Indications for Use

510(k) Number (if known): K082619

Device Name: SILS™ Port

Indications For Use: For multiple instrument or camera access to the abdominal cavity through a single incision for performing minimally invasive laparoscopic procedures

Prescription Use    
 (Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_   
 (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

Neil R. Ghan for MDR

(Division Sign-Off)  
Division of General, Restorative,  
and Neurological Devices

510(k) Number K082619

Page 1 of 1