

DEC 19 2008

**stryker**<sup>®</sup>

**Endoscopy**

**510(k) SUMMARY OF SAFETY AND EFFECTIVENESS**

**Device Name**

Proprietary Name: Stryker LED Light Source (L9000) K082813  
Common and Usual Names: Light Source, illuminator  
Classification Name: Light Source, Fiberoptic, Routine, CFR 21 § 876.1500

**Product Description:**

**Intended Use:** The Stryker LED Light Source is used to illuminate the site of surgery during minimally invasive surgical procedures in Arthroscopy (orthopedic surgery), Laparoscopy (general and gynecological surgery) and in Endoscopy (general, gastroenterological and ENT surgery). The light is transmitted from the source through an optical cable and a scope.

**Voluntary Safety and Performance Standards:** The Stryker LED Light Source will conform to the voluntary standards listed in section 5.1.

**Predicate Devices:** The Stryker LED Light Source is substantially equivalent in terms of safety and effectiveness to currently marketed devices including the Stryker Quantum 5000 Light Source (K961971).

**Substantial Equivalence:** The technological differences between the LED Light Source (L9000) and the Stryker Quantum 5000 Light Source do not raise new questions of safety or effectiveness. Therefore, the Stryker LED Light Source (L9000) is substantially equivalent to the predicate marketed device. Refer to Section 7.0 for a detailed comparison.

Contact:



Desiree Crisolo  
Stryker Endoscopy  
5900 Optical Court  
San Jose, CA 95138  
Phone: 408-754-2784  
Fax: 408-754-2521  
Email: [desiree.crisolo@stryker.com](mailto:desiree.crisolo@stryker.com)

23 September 2008

Date:



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

**MAR 24 2009**

Stryker Corporation  
% Ms. Desiree Crisolo  
Sr. Regulatory Representative  
5900 Optical Court  
San Jose, California 95138

Re: K082813  
Trade/Device Name: Stryker LED Lightsource  
Regulation Number: 21 CFR 876.1500  
Regulation Name: Endoscope and accessories  
Regulatory Class: II  
Product Code: FCW  
Dated: December 19, 2008  
Received: November 14, 2008

Dear Ms. Crisolo:

This letter corrects our substantially equivalent letter of December 19, 2008.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to continue marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Mark N. Melkerson", with a long horizontal flourish extending to the right.

Mark N. Melkerson  
Director  
Division of General, Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

INTENDED USE

Device Name: Stryker LED Light Source (L9000)

510(k) Number if known: K082813

The Stryker LED Light Source is used to illuminate the site of surgery during minimally invasive surgical procedures in arthroscopy (orthopedic surgery), laparoscopy (general and gynecological surgery) and in Endoscopy (general, gastroenterological and ENT surgery). The light is transmitted from the source through an optical cable and a scope.

The light source is an integral component of a visualization system which consists of a video camera, video monitor, video recorder, video printer, light cable and scope.

Prescription Use X  
(Part 21 CFR 801 Subpart D)

AND/OR Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

*Michael John Forman*  
(Division Sign-Off)  
Division of General, Restorative,  
and Neurological Devices

510(k) Number K02813