



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room –WO66-G609
Silver Spring, MD 20993-0002

MAR 31 2011

Ms. Lisa Stahl
Senior Regulatory Affairs Specialist
Medtronic, Inc.
8200 Coral Sea St. NE
Mounds View, MN 55112

Re: P080006/S004
Attain Ability Straight Model 4396 Left Ventricular Lead
Filed: October 19, 2009
Amended: October 20, 2009; September 15, 2010; March 1, 2011
Procode: OJX

Dear Ms. Stahl:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its evaluation of your premarket approval application (PMA) supplement, which requested approval for the Attain Ability Straight Model 4396 Left Ventricular Lead. The Model 4396 lead is a 4 French, transvenous, steroid eluting, dual electrode, polyurethane insulated, single coil, straight tip, tine fixation lead intended for use in left heart coronary veins. Based upon the information submitted, the PMA supplement is approved. You may begin commercial distribution of the device as modified by your PMA supplement in accordance with the conditions described below.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the act). The device is further restricted under section 515(d)(1)(B)(ii) of the act insofar as the labeling must specify the specific training or experience practitioners need in order to use the device. FDA has determined that these restrictions on sale and distribution are necessary to provide reasonable assurance of the safety and effectiveness of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the act, in addition to the many other FDA requirements governing the manufacture, distribution, and marketing of devices.

Expiration dating for this device has been established and approved at 1 year.

Continued approval of this PMA is contingent upon the submission of periodic reports, required under 21 CFR 814.84, at intervals of one year (unless otherwise specified) from the date of approval of the original PMA. Two copies of this report, identified as "Annual Report" (please use this title even if the specified interval is more frequent than one year) and bearing the applicable PMA reference number, should be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and should include the information required by 21 CFR 814.84.

In addition to the above, and in order to provide continued reasonable assurance of the safety and effectiveness of the device, the Annual Report must include, separately for each model number (if applicable), the number of devices sold and distributed during the reporting period, including those distributed to distributors. The distribution data will serve as a denominator and provide necessary context for FDA to ascertain the frequency and prevalence of adverse events, as FDA evaluates the continued safety and effectiveness of the device.

As a condition of approval, you have agreed to collect and provide the following data as part of a future PMA supplement:

Medtronic commits to evaluate the tightening of the in vitro drug elution acceptance criteria based on the collection of drug elution profile data from release and stability batches. At a minimum, Medtronic will collect drug elution profile data on the first full lead lot manufactured per month for one year (total of 12 lots) for further elution testing. In addition 3 lots stored at controlled room temperature for 2 years will be elution tested at the following time points: 0 months, 12 months and 24 months. For the collection of the elution data, the acceptance criteria will be set to +/-12% and data from elution testing at Level L1 (total of 6 units), and if necessary L2 (total of 12 units), or L3 (total of 24 units) will be provided. Medtronic understands that before any failure on meeting the +/- 12% acceptance criteria is reported, the elution testing must be extended to L3 for a total of 24 tested units.

In addition, because your device is a pacemaker, implantable cardioverter-defibrillator (ICD), or system lead, FDA has determined that the following additional information is necessary to provide continued reasonable assurance of the safety and effectiveness of the device. In the Annual Report, provide the following information known by or reported to the applicant:

1. The number of leads domestically implanted and the number of reported explants and deaths.
2. A breakdown of the reported deaths into lead-related and non-lead-related.
3. A breakdown of the reported explants into the number reported that were associated with mechanical failure, associated with clinical complications, and as applicable, the numbers that experienced other safety and effectiveness complications as ascertained by the user, applicant, or otherwise.

4. The number of leads returned to the applicant for cause from domestic sources, with a breakdown into the number currently in analysis, the number operating properly, the number failed (with failure mechanisms described); broken down into groupings for full leads and partial leads.
5. A cumulative survival table for the leads.

In addition to the Annual Report requirements, you must provide the following data in post-approval study reports (PAS). Two copies, identified as "PMA Post-Approval Study Report" and bearing the applicable PMA reference number, should be submitted to the address below.

A nonrandomized, multi-site, world-wide study of implanted commercially available Model 4396 leads. The study will be conducted per the final protocol version 2 dated February 28, 2011 (P080006/S004/A003). The primary objective of the study is to evaluate the chronic performance of the Model 4396 lead. The hypothesis is to demonstrate that the Model 4396 lead-related complication-free rate is greater than 92.5% at five years post-implant. The complication free rate will be estimated based on clinical adverse events including: failure to capture, failure to sense/under-sensing, threshold rise, over-sensing, abnormal pacing impedance, lead insulation breach, lead conductor fracture, extra-cardiac stimulation, cardiac perforation, lead dislodgement, and structural lead failure. The secondary objectives will provide descriptive information on device performance. A total of 1,016 subjects will be enrolled. A minimum of 600 are required for the primary analysis. Subjects will be followed from their enrollment date, and at six month intervals through five years post-implant.

FDA would like to remind you that you are required to submit PAS Progress Reports every six months. The PAS Progress Reports should be submitted separately from the Annual Reports. Please refer to the guidance document on how to handle post-approval studies imposed by approval orders, located at the following website:

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070974.htm>

Failure to comply with all post-approval requirements constitutes a ground for withdrawal of approval of a PMA. Commercial distribution of a device that is not in compliance with these conditions is a violation of the act.

Be advised that the failure to conduct any such study in compliance with the good clinical laboratory practices in 21 CFR part 58 (if a non-clinical study subject to part 58) or the institutional review board regulations in 21 CFR part 56 and the informed consent regulations in 21 CFR part 50 (if a clinical study involving human subjects) may be grounds for FDA withdrawal of approval of the PMA.

Before making any change affecting the safety or effectiveness of the device, you must submit a PMA supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.39. All PMA supplements and alternate submissions (30-day notice) must comply with the applicable requirements in 21 CFR 814.39. For more information, please refer to the FDA guidance document entitled, "Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process" (www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089274.htm).

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52, you are required to report adverse events for this device. Manufacturers of medical devices, including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or
2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm.

In accordance with the recall requirements specified in 21 CFR 806.10, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at www.fda.gov/Safety/Recalls/IndustryGuidance/default.htm.

CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this PMA submission with copies of all approved labeling in final printed form. Final printed labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final printed labeling is identical to the labeling approved in draft form. If the final printed labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted in triplicate, unless otherwise specified, to the address below and should reference the above PMA number to facilitate processing. One of those three copies may be an electronic copy (eCopy), in an electronic format that FDA can process, review and archive (general information:


<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PremarketSubmissions/ucm134508.htm>; clinical and statistical data:

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PremarketSubmissions/ucm136377.htm>).

U.S. Food and Drug Administration
Center for Devices and Radiological Health
PMA Document Mail Center – WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

If you have any questions regarding this letter, you may contact (b) (6) at (b) (6) or email at (b) (6)

Sincerely yours,

for 

Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health