



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Virulite LLC
% Ms. Beky Pine
Noblitt & Rueland
5405 Alton Parkway 5A, #530
Irvine, CA 92604

September 24, 2018

Re: DEN090012 (K083767)
VirusLite Cold Sore Machine
Evaluation of Automatic Class III Designation – De Novo Request
Regulation Number: 21 CFR 878.4860
Regulation Name: Light based energy source device for topical application
Regulatory Classification: Class II
Product Code: OKJ
Dated: June 25, 2009
Received: June 30, 2009

Dear Ms. Pine:

This letter corrects our classification orders dated October 18, 2012 and February 5, 2018.

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your De Novo request for classification of the VirusLite Cold Sore Machine, an over-the-counter device under 21 CFR Part 801 Subpart C that is indicated for the following:

The VirusLite Cold Sore Machine is indicated for shortening the time to healing of herpes simplex labialis lesions on or around the lips with time to healing defined as the time to patient described re-epithelialization.

FDA concludes that this device should be classified into class II. This order, therefore, classifies the VirusLite Cold Sore Machine, and substantially equivalent devices of this generic type, into class II under the generic name, light based energy source device for topical application.

FDA identifies this generic type of device as:

Light based energy source device for topical application. The device emits light energy at near infrared spectrum and is applied externally to the surface of herpes simplex labialis lesions on or around the lips.

In accordance with section 513(f)(1) of the Federal Food, Drug, and Cosmetic Act (21 U.S.C. 360c(f)(1)) (the FD&C Act), devices that were not in commercial distribution prior to May 28, 1976 (the date of enactment of the Medical Device Amendments of 1976 (the amendments)),

generally referred to as postamendments devices, are classified automatically by statute into class III without any FDA rulemaking process. These devices remain in class III and require premarket approval, unless and until the device is classified or reclassified into class I or II or FDA issues an order finding the device to be substantially equivalent, in accordance with section 513(i) of the FD&C Act (21 U.S.C. 360c(i)), to a predicate device that does not require premarket approval. The agency determines whether new devices are substantially equivalent to previously marketed devices by means of premarket notification procedures in section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and Part 807 of the FDA regulations (21 CFR 807).

Section 513(f)(2) of the Food, Drug and Cosmetic Act (the FD&C Act) was amended by section 607 of the Food and Drug Administration Safety and Innovation Act (FDASIA) on July 9, 2012. This new law provides two options for De Novo classification. First, any person who receives a "not substantially equivalent" (NSE) determination in response to a 510(k) for a device that has not been previously classified under the Act may, within 30 days of receiving notice of the NSE determination, request FDA to make a risk-based classification of the device under section 513(a)(1) of the Act. Alternatively, any person who determines that there is no legally marketed device upon which to base a determination of substantial equivalence may request FDA to make a risk-based classification of the device under section 513(a)(1) of the Act without first submitting a 510(k). FDA shall, within 120 days of receiving such a request, classify the device. This classification shall be the initial classification of the device. Within 30 days after the issuance of an order classifying the device, FDA must publish a notice in the **Federal Register** classifying the device type.

In accordance with section 513(f)(1) of the FD&C Act, FDA issued an order on June 10, 2009 automatically classifying the ViruLite Cold Sore Machine in class III, because it was not within a type of device which was introduced or delivered for introduction into interstate commerce for commercial distribution before May 28, 1976, nor which was subsequently reclassified into class I or class II. On June 30, 2009, FDA received your De Novo requesting classification of the ViruLite Cold Sore Machine into class II. The request was submitted under section 513(f)(2) of the FD&C Act. In order to classify the ViruLite Cold Sore Machine into class I or II, it is necessary that the proposed class have sufficient regulatory controls to provide reasonable assurance of the safety and effectiveness of the device for its intended use.

After review of the information submitted in the De Novo request, FDA has determined that the ViruLite Cold Sore Machine, indicated for shortening the time to healing of herpes simplex labialis lesions on or around the lips with time to healing defined as the time to patient described re-epithelialization, can be classified in class II with the establishment of special controls. FDA believes that class II (special controls) provide reasonable assurance of the safety and effectiveness of the device type. The identified risks and mitigation measures associated with the device type are summarized in Table 1.

Table 1 – Identified Risks to Health and Mitigation Measures

Identified Risk	Mitigation Measures
Redness and discomfort	Clinical performance testing Usability testing Labeling

Identified Risk	Mitigation Measures
Burns and blisters	Clinical performance testing Usability testing Labeling
Adverse tissue reaction	Biocompatibility evaluation
Infection/transmissibility	Labeling Cleaning and disinfection validation Usability testing
Electrical shock	Electrical safety testing Labeling
Electromagnetic incompatibility	Electromagnetic compatibility testing Labeling
User error	Usability testing Labeling
Ocular injury	Labeling Non-clinical performance testing for ocular safety

In combination with the general controls of the FD&C Act, the light based energy source device for topical application is subject to the following special controls:

1. The technical parameters of the device, including wavelength, treatment time, treatment area, energy density, spot size, and power, must be characterized.
2. The cleaning and disinfection instructions for the device must be validated.
3. The device must be demonstrated to be biocompatible.
4. Performance testing must validate electromagnetic compatibility (EMC), ocular safety, and electrical safety of the device.
5. Labeling must direct end-users to contact the device manufacturer and MedWatch if they experience any adverse events when using this device.
6. Labeling must include specific information pertinent to use of the device by the intended patient population and the treatment regimen.
7. Simulated use testing must include information from a usability, label comprehension and self-selection study to demonstrate that the device can be used by the intended patient population without any assistance.
8. Clinical data must show adequate reduction in time to healing and assess risks of redness, discomfort, burns, and blisters.

Section 510(m) of the FD&C Act provides that FDA may exempt a class II device from the premarket notification requirements under section 510(k) of the FD&C Act, if FDA determines that premarket notification is not necessary to provide reasonable assurance of the safety and effectiveness of the device type. FDA has determined premarket notification is necessary to

provide reasonable assurance of the safety and effectiveness of the device type, and, therefore, the device is not exempt from the premarket notification requirements of the FD&C Act. Thus, persons who intend to market this device type must submit a premarket notification containing information on the light based energy source device for topical application they intend to market prior to marketing the device and receive clearance to market from FDA.

Please be advised that FDA's decision to grant this De Novo request does not mean that FDA has made a determination that your device complies with other requirements of the FD&C Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the FD&C Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the FD&C Act); 21 CFR 1000-1050.

A notice announcing this classification order will be published in the **Federal Register**. A copy of this order and supporting documentation are on file in the Dockets Management Branch (HFA-305), Food and Drug Administration, 5630 Fishers Lane, Room 1061, Rockville, MD 20852 and are available for inspection between 9 a.m. and 4 p.m., Monday through Friday.

As a result of this order, you may immediately market your device as described in the De Novo request, subject to the general control provisions of the FD&C Act and the special controls identified in this order.

If you have any questions concerning this classification order, please contact Neil Ogden at 301-796-6397.

Sincerely,

Angela C. Krueger
Deputy Director
for Engineering and Science Review
Office of Device Evaluation
Center for Devices and Radiological Health