

K090054

JUL 29 2009

510(K) Summary

Submitter
Cowellmedi Co., Ltd.
Kim Young Tae
155-4, Gamjeon – 2 Dong,
Sasang-Gu, 617-801,
Busan, South Korea
Phone: 82-51-314-2028
Fax: 82-51-314-2026

Official Correspondent
Kodent Inc.
Jung Bae Bang
13340 E. Firestone Blvd. Suite J
Santa Fe Springs, CA 90670
Email: kodentinc@kodent.co.kr
Phone: 562-404-8466
Fax: 562-404-2757

Device Information

Product Name: Atlas Implant System – Wide Plus

Common Name: Endosseous Dental Implant

Classification Name: Implant, Endosseous, Root-Form

Product Code: DZE

Regulation Number: 872.3640

Device Class: Class II

Device Description

The Atlas Implant System – Wide Plus is a dental implant made of titanium metal intended to be surgically placed in the bone of the upper and / or lower jaw arches. This system is similar to other commercially available products based on the intended use, the technology used, the claims, the material composition employed and performance characteristics.

Indication for Use

The Atlas Implant System – Wide Plus is indicated for use in partially or fully edentulous mandibles and maxillae, in support of single or multiple-unit restorations including; cemented retained, screw retained, or overdenture restorations, and terminal or intermediate abutment support for fixed bridgework. Also, there are no angled abutments intended to be used with this system. Smaller (ø 4.5, ø 5.5) implants are dedicated for immediate loading when good primary stability is achieved and with appropriate occlusal loading. Larger (ø 6.0 to ø 8.0) implants are dedicated for the molar region and not indicated for immediate loading.

Materials

This device are manufactured from TI6Al-4V ELI alloy following ASTM and ISO standards.

Predicate Devices

The subject device is substantially equivalent to the following predicate devices:

- Atlas Implant System (K071148) manufactured by Cowellmedi Co., Ltd.
- Rescue External Implant System (K081302) manufactured by Megagen Implant Co., Ltd.
- US Ultra Wide System (K073465) manufactured by Osstem Co., Ltd.

Comparison to Predicate Devices

Testing and other comparisons have established that the subject of Atlas Implant System – Wide Plus is substantially equivalent in design, materials, indications and intended use, packaging, labeling, and performance to other predicate devices of the type currently marketed in the U.S.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Cowellmedi Company, Limited
C/O Mr. Jung Bae Bang
Regulatory Affairs
Kodent Incorporated
13340 East Firestone Boulevard, Suite J
Santa Fe Springs, California 90670

JUL 29 2009

Re: K090054
Trade/Device Name: Atlas Implant System – Wide Plus
Regulation Number: 21 CFR 872.3640
Regulation Name: Endosseous Dental Implant
Regulatory Class: II
Product Code: DZE
Dated: June 9, 2009
Received: June 9, 2009

Dear Mr. Bang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/cdrh/mdr/> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Susan Runner, B.D.S., M.A.

Acting Division Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

K090054

Indication for Use

510(K) Number (if known): K090054

Device Name: Atlas Implant System – Wide Plus

Indication for Use:

The Atlas Implant System – Wide Plus is indicated for use in partially or fully edentulous mandibles and maxillae, in support of single or multiple-unit restorations including; cemented retained, screw retained, or overdenture restorations, and terminal or intermediate abutment support for fixed bridgework. Also, there are no angled abutments intended to be used with this system. Smaller (ø 4.5, ø 5.5) implants are dedicated for immediate loading when good primary stability is achieved and with appropriate occlusal loading. Larger (ø 6.0 to ø 8.0) implants are dedicated for the molar region and not indicated for immediate loading.

Prescription Use X

AND/OR

Over-The-Counter

(Part 21 CFR 801 Subpart D)

(Per 21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page 1 of 1

Kevin M. Kelly for MSR
(Division Sign-Off)

Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K090054