

K090410

## Premarket Notification 510(k) Summary CoolTouch Thermal Sensing Handpiece Accessory

This 510(K) Summary of safety and effectiveness for the CoolTouch Thermal Sensing Handpiece Accessory is submitted in accordance with the requirements of 21CFR 807.92.

Applicant: New Star Lasers, Inc. dba CoolTouch, Inc.

Address: 9085 Foothills Boulevard  
Roseville, CA 95747

Contact Person: Natalie R. Vollrath

Telephone: (916) 677-1900

Fax: (916) 677-1901

Preparation Date: February 17, 2009

Device Trade Name: CoolTouch Thermal Sensing Handpiece Accessory

Common Name: Handpiece Accessory

Classification Name: Instrument, Surgical Powered, Laser 79-GEX

Legally Marketed Predicate Device: New Star Temperature Diagnostic Accessory

Description of the CoolTouch Thermal Sensing Handpiece Accessory: The Thermal Sensing Handpiece Accessory is a temperature detector which will provide the laser operator with a readout of the temperature of the treatment area

Intended use of the CoolTouch Thermal Sensing Handpiece Accessory: For use as a sensing device to measure and display the temperature of the treatment area during procedures with the NS130 laser

Performance Data: None

Conclusion: Based on the evaluation of the risks and hazards and including various testing of the modifications, the CoolTouch Thermal Sensing Handpiece Accessory is substantially equivalent to the predicate device, the New Star Temperature Diagnostic Accessory

MAR 27 2009



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

New Star Lasers, Inc.  
% Ms. Natalie Vollrath  
QA/RA Manager  
9085 Foothills Boulevard  
Roseville, California 95747

MAR 27 2009

Re: K090410

Trade/Device Name: CoolTouch Thermal Sensing Handpiece Accessory  
Regulation Number: 21 CFR 878.4810  
Regulation Name: Laser surgical instrument for use in general and plastic surgery and in dermatology  
Regulatory Class: II  
Product Code: GEX  
Dated: March 19, 2009  
Received: March 23, 2009

Dear Ms. Vollrath:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set

Page 2 - Ms. Natalie Vollrath

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at (240) 276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at (240) 276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkerson

Director

Division of General, Restorative  
and Neurological Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

**510(k) Number**

**Device Name** CoolTouch Thermal Sensing Handpiece Accessory

**Indications for Use** The CoolTouch Thermal Sensing Handpiece Accessory is intended for use as a sensing device to measure and display the temperature of the treatment area during procedures with the NS-130 Laser.

Prescription Use  X   
(Part 21 CFR 801 Subpart D)

AND/OR

Over-the-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

David Krone for AXM 3/27/2009  
(Division Sign-Off)

Division of General, Restorative,  
and Neurological Devices

Page  1  of  1

510(k) Number  K090410